

Data Path : Z:\HPCHEM1\BNA E\DATA\BE022315\
 Data File : BE089656.D
 Acq On : 24 Feb 2015 00:47
 Operator : TP/JJ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Quant Time: Feb 24 04:03:36 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE022315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 24 03:46:57 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	120	0.00
2	1,4-Dioxane	1.000	1.011	-1.1	97	0.00
3	n-Nitrosodimethylamine	1.000	1.044	-4.4	106	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	120	0.00
5 S	Nitrobenzene-d5	1.000	0.957	4.3	97	0.00
6	Naphthalene	1.000	1.003	-0.3	104	0.00
7	2-Methylnaphthalene	1.000	1.029	-2.9	104	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	119	0.00
9 S	2-Fluorobiphenyl	1.000	0.975	2.5	100	0.00
10	Acenaphthylene	1.000	1.038	-3.8	105	-0.01
11	Acenaphthene	1.000	1.012	-1.2	105	0.00
12	Fluorene	1.000	1.032	-3.2	104	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	102	0.00
14	Hexachlorobenzene	1.000	1.037	-3.7	94	0.00
15	Pentachlorophenol	1.000	0.943	5.7	95	0.00
16	Phenanthrene	1.000	1.015	-1.5	90	0.00
17	Anthracene	1.000	1.032	-3.2	90	0.00
18	Fluoranthene	1.000	1.051	-5.1	102	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	117	0.00
20	Pvrene	1.000	1.037	-3.7	103	0.00
21 S	Terphenyl-d14	1.000	1.085	-8.5	78	0.00
22	Benzo(a)anthracene	1.000	0.992	0.8	101	0.00
23	Chrysene	1.000	1.017	-1.7	104	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	0.876	12.4	95	0.00
25 I	Perylene-d12	5.000	5.000	0.0	116	0.00
26	Benzo(b)fluoranthene	1.000	0.853	14.7	102	0.00
27	Benzo(k)fluoranthene	1.000	1.019	-1.9	100	0.00
28 C	Benzo(a)pyrene	1.000	0.883	11.7	98	0.00
29	Dibenzo(a,h)anthracene	1.000	0.880	12.0	96	0.00
30	Benzo(g,h,i)perylene	1.000	0.920	8.0	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0