

Data Path : Z:\HPCHEM1\BNA E\DATA\BE022315\
 Data File : BE089664.D
 Acq On : 24 Feb 2015 6:36
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Quant Time: Feb 24 07:45:25 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE022315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 24 03:46:57 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.642	0.761	-18.5	104	0.00
3	n-Nitrosodimethylamine	0.745	1.010	-35.6#	109	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
5 S	Nitrobenzene-d5	0.291	0.310	-6.5	105	0.00
6	Naphthalene	1.021	1.183	-15.9	110	0.00
7	2-Methylnaphthalene	0.575	0.694	-20.7	112	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	111	0.00
9 S	2-Fluorobiphenyl	1.355	1.459	-7.7	112	0.00
10	Acenaphthylene	7.371	8.903	-20.8	114	0.00
11	Acenaphthene	1.194	1.388	-16.2	112	0.00
12	Fluorene	1.311	1.555	-18.6	112	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
14	Hexachlorobenzene	0.136	0.162	-19.1	99	0.00
15	Pentachlorophenol	0.020	0.016	20.0	90	0.00
16	Phenanthrene	1.103	1.285	-16.5	95	0.00
17	Anthracene	0.931	1.126	-20.9	97	0.00
18	Fluoranthene	0.917	1.103	-20.3	107	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
20	Pvrene	1.256	1.530	-21.8	108	0.00
21 S	Terphenyl-d14	0.714	1.170	-63.9#	105	0.00
22	Benzo(a)anthracene	3.489	3.839	-10.0	99	0.00
23	Chrysene	4.726	5.526	-16.9	106	0.00
24	Indeno(1,2,3-cd)pyrene	1.955	2.016	-3.1	93	0.00
25 I	Perylene-d12	1.000	1.000	0.0	102	0.00
26	Benzo(b)fluoranthene	0.525	0.472	10.1	93	0.00
27	Benzo(k)fluoranthene	1.190	1.581	-32.9#	100	0.00
28 C	Benzo(a)pyrene	0.565	0.579	-2.5	93	0.00
29	Dibenzo(a,h)anthracene	0.466	0.483	-3.6	93	0.00
30	Benzo(g,h,i)perylene	2.566	2.981	-16.2	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0