

Data Path : Z:\HPCHEM1\BNA E\DATA\BE022315\
 Data File : BE089664.D
 Acq On : 24 Feb 2015 6:36
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Quant Time: Feb 24 07:45:25 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE022315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 24 03:46:57 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	109	0.00
2	1,4-Dioxane	1.000	1.187	-18.7	104	0.00
3	n-Nitrosodimethylamine	1.000	1.173	-17.3	109	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	109	0.00
5 S	Nitrobenzene-d5	1.000	1.110	-11.0	105	0.00
6	Naphthalene	1.000	1.159	-15.9	110	0.00
7	2-Methylnaphthalene	1.000	1.208	-20.8	112	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	111	0.00
9 S	2-Fluorobiphenyl	1.000	1.197	-19.7	112	0.00
10	Acenaphthylene	1.000	1.208	-20.8	114	0.00
11	Acenaphthene	1.000	1.162	-16.2	112	0.00
12	Fluorene	1.000	1.186	-18.6	112	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	94	0.00
14	Hexachlorobenzene	1.000	1.194	-19.4	99	0.00
15	Pentachlorophenol	1.000	0.960	4.0	90	0.00
16	Phenanthrene	1.000	1.165	-16.5	95	0.00
17	Anthracene	1.000	1.210	-21.0	97	0.00
18	Fluoranthene	1.000	1.204	-20.4	107	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	104	0.00
20	Pvrene	1.000	1.219	-21.9	108	0.00
21 S	Terphenyl-d14	1.000	1.638	-63.8#	105	0.00
22	Benzo(a)anthracene	1.000	1.100	-10.0	99	0.00
23	Chrysene	1.000	1.169	-16.9	106	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	0.932	6.8	93	0.00
25 I	Perylene-d12	5.000	5.000	0.0	102	0.00
26	Benzo(b)fluoranthene	1.000	0.870	13.0	93	0.00
27	Benzo(k)fluoranthene	1.000	1.162	-16.2	100	0.00
28 C	Benzo(a)pyrene	1.000	0.933	6.7	93	0.00
29	Dibenzo(a,h)anthracene	1.000	0.930	7.0	93	0.00
30	Benzo(g,h,i)perylene	1.000	1.028	-2.8	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0