

Data Path : Z:\HPCHEM1\BNA E\DATA\BE022318\
 Data File : BE095678.D
 Acq On : 23 Feb 2018 12:08
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE022318

Quant Time: Feb 23 15:55:09 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE022318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 23 12:12:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.630	0.669	-6.2	111	0.00
3	n-Nitrosodimethylamine	0.781	0.786	-0.6	106	-0.01
4 S	2-Fluorophenol	1.056	1.040	1.5	104	0.00
5 S	Phenol-d6	1.462	1.448	1.0	101	0.00
6	bis(2-Chloroethyl)ether	1.500	1.513	-0.9	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	104	-0.02
8 S	Nitrobenzene-d5	0.412	0.381	7.5	99	-0.02
9	Naphthalene	4.611	4.652	-0.9	103	0.00
10	Hexachlorobutadiene	0.269	0.300	-11.5	110	0.00
11 SURR	2-Methylnaphthalene-d10	0.669	0.675	-0.9	104	0.00
12	2-Methylnaphthalene	0.796	0.797	-0.1	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	106	0.00
14 S	2,4,6-Tribromophenol	0.143	0.135	5.6	112	0.00
15 S	2-Fluorobiphenyl	1.787	1.817	-1.7	106	0.00
16	Acenaphthylene	2.247	2.169	3.5	103	0.00
17	Acenaphthene	1.411	1.366	3.2	102	0.00
18	Fluorene	1.750	1.722	1.6	104	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
20	4-Bromophenyl-phenylether	0.251	0.245	2.4	101	0.00
21	Hexachlorobenzene	0.260	0.259	0.4	104	0.00
22	Pentachlorophenol	0.060	0.051	15.0	110	-0.01
23	Phenanthrene	1.250	1.239	0.9	101	0.00
24	Anthracene	1.148	1.108	3.5	103	0.00
25 SURR	Fluoranthene-d10	5.225	5.073	2.9	102	0.00
26	Fluoranthene	1.538	1.497	2.7	101	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
28	Pvrene	1.636	1.520	7.1	97	0.00
29 S	Terphenyl-d14	0.797	0.788	1.1	102	0.00
30	Benzo(a)anthracene	1.305	1.287	1.4	103	0.00
31	Chrysene	1.298	1.289	0.7	102	0.00
32	Bis(2-ethylhexyl)phthalate	2.308	2.049	11.2	101	0.00
33	Indeno(1,2,3-cd)pyrene	1.257	1.219	3.0	101	0.00
34 I	Perylene-d12	1.000	1.000	0.0	103	0.00
35	Benzo(b)fluoranthene	1.425	1.406	1.3	99	0.00
36	Benzo(k)fluoranthene	1.417	1.407	0.7	101	0.00
37 C	Benzo(a)pyrene	1.284	1.261	1.8	100	0.00
38	Dibenzo(a,h)anthracene	1.143	1.106	3.2	100	0.00
39	Benzo(a,h,i)perylene	1.224	1.193	2.5	99	0.00

Data Path : Z:\HPCHEM1\BNA E\DATA\BE022318\
Data File : BE095678.D
Acq On : 23 Feb 2018 12:08
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE022318

Quant Time: Feb 23 15:55:09 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE022318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 23 12:12:42 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			