

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE022415\
 Data File : BE089666.D
 Acq On : 24 Feb 2015 11:46
 Operator : TP/JJ
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.1

Manual Integrations
 APPROVED

apatel
 2/25/2015 3:34:14 PM

Quant Time: Feb 24 15:09:12 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE022415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 24 14:51:37 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.34	152	45260	5.00	ng	0.00
4) Naphthalene-d8	12.24	136	158728	5.00	ng	0.00
8) Acenaphthene-d10	16.41	164	69802	5.00	ng	0.00
13) Phenanthrene-d10	19.77	188	126209	5.00	ng	0.00
19) Chrysene-d12	23.65	240	90466	5.00	ng	0.00
25) Perylene-d12	25.34	264	68326	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	10.60	82	1063	0.12	ng	0.00
9) 2-Fluorobiphenyl	14.88	172	2013	0.11	ng	0.00
21) Terphenyl-d14	22.30	244	1024	0.09	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.73	88	547	0.10	ng	93
3) n-Nitrosodimethylamine	3.52	42	490m	0.07	ng	
6) Naphthalene	12.29	128	3246	0.10	ng	96
7) 2-Methylnaphthalene	13.95	142	1697	0.10	ng	97
10) Acenaphthylene	16.06	152	9410	0.09	ng	99
11) Acenaphthene	16.49	154	1644	0.10	ng	87
12) Fluorene	17.74	166	1741	0.10	ng	98
14) Hexachlorobenzene	18.99	284	342	0.10	ng	# 81
16) Phenanthrene	19.81	178	2863	0.11	ng	98
17) Anthracene	19.91	178	2046	0.09	ng	100
18) Fluoranthene	21.72	202	1921	0.08	ng	99
20) Pyrene	22.04	202	2124	0.10	ng	99
22) Benzo(a)anthracene	23.65	228	5739	0.10	ng	94
23) Chrysene	23.68	228	9133	0.11	ng	96
24) Indeno(1,2,3-cd)pyrene	26.74	276	1059	0.03	ng	# 38
26) Benzo(b)fluoranthene	24.91	252	320	0.05	ng	# 71
27) Benzo(k)fluoranthene	24.94	252	802m	0.05	ng	
28) Benzo(a)pyrene	25.27	252	353	0.05	ng	# 64
29) Dibenzo(a,h)anthracene	26.77	278	166	0.03	ng	# 56
30) Benzo(g,h,i)perylene	27.15	276	1347	0.04	ng	# 61

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE022415\
Data File : BE089666.D
Acq On : 24 Feb 2015 11:46
Operator : TP/JJ
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC0.1

Manual Integrations
APPROVED

apatel
2/25/2015 3:34:14 PM

Quant Time: Feb 24 15:09:12 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE022415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 24 14:51:37 2015
Response via : Initial Calibration

