

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE022415\
 Data File : BE089669.D
 Acq On : 24 Feb 2015 13:43
 Operator : TP/JJ
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.8

Quant Time: Feb 24 15:06:32 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE022415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 24 14:51:37 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.34	152	43402	5.00	ng	0.00
4) Naphthalene-d8	12.24	136	154064	5.00	ng	0.00
8) Acenaphthene-d10	16.42	164	67645	5.00	ng	0.01
13) Phenanthrene-d10	19.76	188	123979	5.00	ng	0.00
19) Chrysene-d12	23.66	240	97656	5.00	ng	0.00
25) Perylene-d12	25.34	264	93716	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	10.60	82	6327	0.74	ng	0.00
9) 2-Fluorobiphenyl	14.87	172	13789	0.78	ng	0.00
21) Terphenyl-d14	22.30	244	9949	0.81	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.70	88	4132	0.76	ng	95
3) n-Nitrosodimethylamine	3.38	42	5307	0.84	ng	# 97
6) Naphthalene	12.29	128	24142	0.80	ng	99
7) 2-Methylnaphthalene	13.94	142	13702	0.80	ng	97
10) Acenaphthylene	16.06	152	78321	0.81	ng	100
11) Acenaphthene	16.49	154	12484	0.80	ng	99
12) Fluorene	17.74	166	13840	0.81	ng	100
14) Hexachlorobenzene	18.99	284	2600	0.80	ng	99
15) Pentachlorophenol	19.43	266	236	0.49	ng	93
16) Phenanthrene	19.81	178	21281	0.81	ng	100
17) Anthracene	19.91	178	18329	0.82	ng	99
18) Fluoranthene	21.72	202	18025	0.81	ng	100
20) Pyrene	22.04	202	19649	0.85	ng	100
22) Benzo(a)anthracene	23.64	228	48706	0.75	ng	99
23) Chrysene	23.68	228	73512	0.82	ng	99
24) Indeno(1,2,3-cd)pyrene	26.73	276	25932	0.74	ng	# 46
26) Benzo(b)fluoranthene	24.91	252	4860	0.52	ng	97
27) Benzo(k)fluoranthene	24.94	252	17527	0.83	ng	99
28) Benzo(a)pyrene	25.27	252	5799	0.58	ng	98
29) Dibenzo(a,h)anthracene	26.77	278	4460	0.55	ng	93
30) Benzo(g,h,i)perylene	27.15	276	31493	0.69	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE022415\
Data File : BE089669.D
Acq On : 24 Feb 2015 13:43
Operator : TP/JJ
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC0.8

Quant Time: Feb 24 15:06:32 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE022415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 24 14:51:37 2015
Response via : Initial Calibration

