

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE022415\
 Data File : BE089674.D
 Acq On : 24 Feb 2015 16:59
 Operator : TP/JJ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE022415

Quant Time: Feb 25 14:26:23 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE022415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 25 14:06:35 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	128	0.00
2	1,4-Dioxane	0.627	0.541	13.7	102	0.00
3	n-Nitrosodimethylamine	0.766	0.686	10.4	101	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	129	0.00
5 S	Nitrobenzene-d5	0.282	0.220	22.0	136	0.00
6	Naphthalene	0.968	0.880	9.1	127	0.00
7	2-Methylnaphthalene	0.550	0.504	8.4	126	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	128	0.00
9 S	2-Fluorobiphenyl	1.295	1.035	20.1	128	0.00
10	Acenaphthylene	7.176	6.404	10.8	124	0.00
11	Acenaphthene	1.137	1.022	10.1	125	0.00
12	Fluorene	1.261	1.128	10.5	124	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	126	0.00
14	Hexachlorobenzene	0.129	0.116	10.1	124	0.00
15	Pentachlorophenol	0.014	0.012	14.3	133	0.00
16	Phenanthrene	1.067	0.953	10.7	124	0.00
17	Anthracene	0.908	0.800	11.9	121	0.00
18	Fluoranthene	0.890	0.778	12.6	119	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
20	Pyrene	1.208	1.124	7.0	119	0.00
21 S	Terphenyl-d14	0.587	0.541	7.8	120	0.00
22	Benzo(a)anthracene	3.288	2.905	11.6	104	0.00
23	Chrysene	4.532	3.974	12.3	97	0.00
24	Indeno(1,2,3-cd)pyrene	1.340	1.193	11.0	116	0.00
25 I	Perylene-d12	1.000	1.000	0.0	119	0.00
26	Benzo(b)fluoranthene	0.289	0.293	-1.4	123	0.00
27	Benzo(k)fluoranthene	0.999	1.093	-9.4	115	0.00
28 C	Benzo(a)pyrene	0.416	0.370	11.1	117	0.00
29	Dibenzo(a,h)anthracene	0.243	0.290	-19.3	118	0.00
30	Benzo(g,h,i)perylene	1.814	1.929	-6.3	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0