

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE022415\
 Data File : BE089674.D
 Acq On : 24 Feb 2015 16:59
 Operator : TP/JJ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE022415

Quant Time: Feb 25 14:26:23 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE022415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 25 14:06:35 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	128	0.00
2	1,4-Dioxane	1.000	0.864	13.6	102	0.00
3	n-Nitrosodimethylamine	1.000	0.886	11.4	101	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	129	0.00
5 S	Nitrobenzene-d5	1.000	0.914	8.6	136	0.00
6	Naphthalene	1.000	0.910	9.0	127	0.00
7	2-Methylnaphthalene	1.000	0.918	8.2	126	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	128	0.00
9 S	2-Fluorobiphenyl	1.000	0.911	8.9	128	0.00
10	Acenaphthylene	1.000	0.892	10.8	124	0.00
11	Acenaphthene	1.000	0.899	10.1	125	0.00
12	Fluorene	1.000	0.894	10.6	124	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	126	0.00
14	Hexachlorobenzene	1.000	0.900	10.0	124	0.00
15	Pentachlorophenol	1.000	1.061	-6.1	133	0.00
16	Phenanthrene	1.000	0.894	10.6	124	0.00
17	Anthracene	1.000	0.881	11.9	121	0.00
18	Fluoranthene	1.000	0.874	12.6	119	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	101	0.00
20	Pyrene	1.000	0.930	7.0	119	0.00
21 S	Terphenyl-d14	1.000	0.922	7.8	120	0.00
22	Benzo(a)anthracene	1.000	0.883	11.7	104	0.00
23	Chrysene	1.000	0.877	12.3	97	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	0.921	7.9	116	0.00
25 I	Perylene-d12	5.000	5.000	0.0	119	0.00
26	Benzo(b)fluoranthene	1.000	0.954	4.6	123	0.00
27	Benzo(k)fluoranthene	1.000	0.935	6.5	115	0.00
28 C	Benzo(a)pyrene	1.000	0.961	3.9	117	0.00
29	Dibenzo(a,h)anthracene	1.000	0.983	1.7	118	0.00
30	Benzo(g,h,i)perylene	1.000	0.892	10.8	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0