

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE022415\
 Data File : BE089682.D
 Acq On : 24 Feb 2015 22:10
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 25 14:27:12 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE022415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 25 14:06:35 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.627	0.706	-12.6	108	-0.03
3	n-Nitrosodimethylamine	0.766	0.845	-10.3	101	-0.02
4 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
5 S	Nitrobenzene-d5	0.282	0.184	34.8#	90	0.00
6	Naphthalene	0.968	1.023	-5.7	116	0.00
7	2-Methylnaphthalene	0.550	0.567	-3.1	112	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
9 S	2-Fluorobiphenyl	1.295	1.304	-0.7	119	0.00
10	Acenaphthylene	7.176	7.250	-1.0	103	0.00
11	Acenaphthene	1.137	1.182	-4.0	107	0.00
12	Fluorene	1.261	1.272	-0.9	103	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
14	Hexachlorobenzene	0.129	0.139	-7.8	103	0.00
15	Pentachlorophenol	0.014	0.012	14.3	88	0.00
16	Phenanthrene	1.067	1.112	-4.2	100	0.00
17	Anthracene	0.908	0.907	0.1	94	0.00
18	Fluoranthene	0.890	0.852	4.3	90	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	60	0.00
20	Pyrene	1.208	1.399	-15.8	89	0.00
21 S	Terphenyl-d14	0.587	0.858	-46.2#	115	0.00
22	Benzo(a)anthracene	3.288	2.738	16.7	59	0.00
23	Chrysene	4.532	5.042	-11.3	74	0.00
24	Indeno(1,2,3-cd)pyrene	1.340	1.199	10.5	70	0.00
25 I	Perylene-d12	1.000	1.000	0.0	70	0.00
26	Benzo(b)fluoranthene	0.289	0.253	12.5	62	0.00
27	Benzo(k)fluoranthene	0.999	1.163	-16.4	71	0.00
28 C	Benzo(a)pyrene	0.416	0.377	9.4	69	0.00
29	Dibenzo(a,h)anthracene	0.243	0.282	-16.0	67	0.00
30	Benzo(g,h,i)perylene	1.814	2.306	-27.1#	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0