

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE022415\
 Data File : BE089682.D
 Acq On : 24 Feb 2015 22:10
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 25 14:27:12 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE022415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 25 14:06:35 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	104	0.00
2	1,4-Dioxane	1.000	1.126	-12.6	108	-0.03
3	n-Nitrosodimethylamine	1.000	1.075	-7.5	101	-0.02
4 I	Naphthalene-d8	5.000	5.000	0.0	102	0.00
5 S	Nitrobenzene-d5	1.000	0.795	20.5	90	0.00
6	Naphthalene	1.000	1.057	-5.7	116	0.00
7	2-Methylnaphthalene	1.000	1.032	-3.2	112	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	94	0.00
9 S	2-Fluorobiphenyl	1.000	1.154	-15.4	119	0.00
10	Acenaphthylene	1.000	1.010	-1.0	103	0.00
11	Acenaphthene	1.000	1.039	-3.9	107	0.00
12	Fluorene	1.000	1.009	-0.9	103	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	87	0.00
14	Hexachlorobenzene	1.000	1.077	-7.7	103	0.00
15	Pentachlorophenol	1.000	1.027	-2.7	88	0.00
16	Phenanthrene	1.000	1.042	-4.2	100	0.00
17	Anthracene	1.000	0.999	0.1	94	0.00
18	Fluoranthene	1.000	0.956	4.4	90	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	60	0.00
20	Pyrene	1.000	1.157	-15.7	89	0.00
21 S	Terphenyl-d14	1.000	1.463	-46.3#	115	0.00
22	Benzo(a)anthracene	1.000	0.833	16.7	59	0.00
23	Chrysene	1.000	1.113	-11.3	74	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	0.925	7.5	70	0.00
25 I	Perylene-d12	5.000	5.000	0.0	70	0.00
26	Benzo(b)fluoranthene	1.000	0.847	15.3	62	0.00
27	Benzo(k)fluoranthene	1.000	0.989	1.1	71	0.00
28 C	Benzo(a)pyrene	1.000	0.976	2.4	69	0.00
29	Dibenzo(a,h)anthracene	1.000	0.962	3.8	67	0.00
30	Benzo(g,h,i)perylene	1.000	1.048	-4.8	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0