

Data Path : Z:\HPCHEM1\BNA E\DATA\BE022417\
 Data File : BE092755.D
 Acq On : 24 Feb 2017 11:53
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 24 15:34:58 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE021517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 15 17:05:20 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	122	-0.02
2	1,4-Dioxane	0.710	0.706	0.6	122	-0.01
3	n-Nitrosodimethylamine	0.740	0.920	-24.3	173#	-0.05
4 S	2-Fluorophenol	0.973	1.313	-34.9#	171#	-0.03
5 S	Phenol-d6	1.194	1.724	-44.4#	182#	-0.05
6	bis(2-Chloroethyl)ether	1.321	1.558	-17.9	156#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	114	-0.02
8 S	Nitrobenzene-d5	0.279	0.321	-15.1	136	-0.05
9	Naphthalene	1.082	1.283	-18.6	134	-0.02
10	Hexachlorobutadiene	0.159	0.156	1.9	107	-0.02
11	2-Methylnaphthalene	0.680	0.736	-8.2	127	-0.03
12 I	Acenaphthene-d10	1.000	1.000	0.0	129	-0.02
13 S	2,4,6-Tribromophenol	0.106	0.183	-72.6#	228#	-0.03
14 S	2-Fluorobiphenyl	1.219	1.186	2.7	124	-0.02
15	Acenaphthylene	7.544	7.741	-2.6	134	-0.02
16	Acenaphthene	1.096	1.112	-1.5	130	-0.02
17	Fluorene	1.407	1.592	-13.1	148	-0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	140	0.00
19	Hexachlorobenzene	0.193	0.190	1.6	125	-0.02
20	Pentachlorophenol	0.042	0.117	-178.6#	456#	-0.12
21	Phenanthrene	1.188	1.226	-3.2	137	-0.02
22	Anthracene	1.094	1.165	-6.5	152#	-0.02
23	Fluoranthene	5.096	6.351	-24.6	174#	-0.02
24 I	Chrysene-d12	1.000	1.000	0.0	165#	-0.02
25	Pyrene	4.761	4.907	-3.1	175#	-0.02
26 S	Terphenyl-d14	0.638	0.651	-2.0	174#	-0.02
27	Benzo(a)anthracene	4.849	5.502	-13.5	191#	-0.02
28	Chrysene	5.752	4.831	16.0	139	-0.02
29	Bis(2-ethylhexyl)phthalate	0.674	0.692	-2.7	196#	0.00
30	Indeno(1,2,3-cd)pyrene	6.486	5.024	22.5	127	-0.03
31 I	Perylene-d12	1.000	1.000	0.0	145	-0.02
32	Benzo(b)fluoranthene	1.414	1.369	3.2	145	-0.02
33	Benzo(k)fluoranthene	1.495	1.362	8.9	145	-0.02
34 C	Benzo(a)pyrene	1.285	1.261	1.9	151#	-0.02
35	Dibenzo(a,h)anthracene	1.355	1.186	12.5	129	-0.03
36	Benzo(g,h,i)perylene	5.728	4.952	13.5	126	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0