

Data Path : Z:\HPCHEM1\BNA_E\Data\BE022515\
 Data File : BE089684.D
 Acq On : 25 Feb 2015 16:57
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 26 18:08:11 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE022415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 25 14:06:35 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	61	0.00
2	1,4-Dioxane	0.627	0.577	8.0	52	-0.08
3	n-Nitrosodimethylamine	0.766	0.702	8.4	49#	-0.04
4 I	Naphthalene-d8	1.000	1.000	0.0	71	0.00
5 S	Nitrobenzene-d5	0.282	0.180	36.2#	61	0.00
6	Naphthalene	0.968	0.988	-2.1	79	0.00
7	2-Methylnaphthalene	0.550	0.644	-17.1	89	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	82	0.00
9 S	2-Fluorobiphenyl	1.295	1.215	6.2	97	0.00
10	Acenaphthylene	7.176	7.419	-3.4	92	0.00
11	Acenaphthene	1.137	1.232	-8.4	97	0.00
12	Fluorene	1.261	1.361	-7.9	96	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	78	0.00
14	Hexachlorobenzene	0.129	0.149	-15.5	98	0.00
15	Pentachlorophenol	0.014	0.024	-71.4#	161#	0.00
16	Phenanthrene	1.067	1.162	-8.9	93	0.00
17	Anthracene	0.908	0.924	-1.8	86	0.00
18	Fluoranthene	0.890	0.835	6.2	79	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	50#	0.00
20	Pyrene	1.208	1.528	-26.5#	80	0.00
21 S	Terphenyl-d14	0.587	0.734	-25.0#	80	0.00
22	Benzo(a)anthracene	3.288	3.042	7.5	54	0.00
23	Chrysene	4.532	5.007	-10.5	60	0.00
24	Indeno(1,2,3-cd)pyrene	1.340	1.267	5.4	61	0.00
25 I	Perylene-d12	1.000	1.000	0.0	50	0.00
26	Benzo(b)fluoranthene	0.289	0.392	-35.6#	69	0.00
27	Benzo(k)fluoranthene	0.999	1.320	-32.1#	58	0.00
28 C	Benzo(a)pyrene	0.416	0.444	-6.7	59	0.00
29	Dibenzo(a,h)anthracene	0.243	0.381	-56.8#	65	0.00
30	Benzo(g,h,i)perylene	1.814	2.410	-32.9#	60	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0