

Data Path : Z:\HPCHEM1\BNA_E\Data\BE022515\
 Data File : BE089684.D
 Acq On : 25 Feb 2015 16:57
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 26 18:08:11 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE022415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 25 14:06:35 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	61	0.00
2	1,4-Dioxane	1.000	0.920	8.0	52	-0.08
3	n-Nitrosodimethylamine	1.000	0.906	9.4	49	-0.04
4 I	Naphthalene-d8	5.000	5.000	0.0	71	0.00
5 S	Nitrobenzene-d5	1.000	0.781	21.9	61	0.00
6	Naphthalene	1.000	1.021	-2.1	79	0.00
7	2-Methylnaphthalene	1.000	1.171	-17.1	89	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	82	0.00
9 S	2-Fluorobiphenyl	1.000	1.073	-7.3	97	0.00
10	Acenaphthylene	1.000	1.034	-3.4	92	0.00
11	Acenaphthene	1.000	1.084	-8.4	97	0.00
12	Fluorene	1.000	1.080	-8.0	96	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	78	0.00
14	Hexachlorobenzene	1.000	1.152	-15.2	98	0.00
15	Pentachlorophenol	1.000	1.725	-72.5#	161	0.00
16	Phenanthrene	1.000	1.089	-8.9	93	0.00
17	Anthracene	1.000	1.018	-1.8	86	0.00
18	Fluoranthene	1.000	0.938	6.2	79	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	50	0.00
20	Pyrene	1.000	1.265	-26.5#	80	0.00
21 S	Terphenyl-d14	1.000	1.252	-25.2#	80	0.00
22	Benzo(a)anthracene	1.000	0.925	7.5	54	0.00
23	Chrysene	1.000	1.105	-10.5	60	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	0.966	3.4	61	0.00
25 I	Perylene-d12	5.000	5.000	0.0	50	0.00
26	Benzo(b)fluoranthene	1.000	1.198	-19.8	69	0.00
27	Benzo(k)fluoranthene	1.000	1.114	-11.4	58	0.00
28 C	Benzo(a)pyrene	1.000	1.110	-11.0	59	0.00
29	Dibenzo(a,h)anthracene	1.000	1.200	-20.0	65	0.00
30	Benzo(g,h,i)perylene	1.000	1.091	-9.1	60	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0