

Data Path : Z:\HPCHEM1\BNA_E\Data\BE022515\
 Data File : BE089690.D
 Acq On : 25 Feb 2015 21:29
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 26 00:27:00 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE022415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 26 00:03:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	124	0.00
2	1,4-Dioxane	0.627	0.615	1.9	112	-0.05
3	n-Nitrosodimethylamine	0.766	0.728	5.0	104	-0.04
4 I	Naphthalene-d8	1.000	1.000	0.0	130	0.00
5 S	Nitrobenzene-d5	0.282	0.259	8.2	161#	0.00
6	Naphthalene	0.968	0.850	12.2	124	0.00
7	2-Methylnaphthalene	0.550	0.490	10.9	124	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
9 S	2-Fluorobiphenyl	1.295	1.557	-20.2	179#	0.00
10	Acenaphthylene	7.176	6.278	12.5	114	0.00
11	Acenaphthene	1.137	0.993	12.7	114	0.00
12	Fluorene	1.261	1.019	19.2	105	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
14	Hexachlorobenzene	0.129	0.117	9.3	94	0.00
15	Pentachlorophenol	0.014	0.017	-21.4	135	0.00
16	Phenanthrene	1.067	0.940	11.9	93	0.00
17	Anthracene	0.908	0.782	13.9	89	0.00
18	Fluoranthene	0.890	0.694	22.0	80	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	64	0.00
20	Pyrene	1.208	1.156	4.3	78	0.00
21 S	Terphenyl-d14	0.587	0.886	-50.9#	125	0.00
22	Benzo(a)anthracene	3.288	2.827	14.0	64	0.00
23	Chrysene	4.532	4.002	11.7	62	0.00
24	Indeno(1,2,3-cd)pyrene	1.340	2.009	-49.9#	124	0.00
25 I	Perylene-d12	1.000	1.000	0.0	86	0.00
26	Benzo(b)fluoranthene	0.289	0.375	-29.8#	114	0.00
27	Benzo(k)fluoranthene	0.999	1.074	-7.5	82	0.00
28 C	Benzo(a)pyrene	0.416	0.449	-7.9	103	0.00
29	Dibenzo(a,h)anthracene	0.243	0.403	-65.8#	119	0.00
30	Benzo(g,h,i)perylene	1.814	2.563	-41.3#	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0