

Data Path : Z:\HPCHEM1\BNA_E\Data\BE022515\
 Data File : BE089690.D
 Acq On : 25 Feb 2015 21:29
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 26 00:27:00 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE022415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 26 00:03:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	124	0.00
2	1,4-Dioxane	1.000	0.981	1.9	112	-0.05
3	n-Nitrosodimethylamine	1.000	0.936	6.4	104	-0.04
4 I	Naphthalene-d8	5.000	5.000	0.0	130	0.00
5 S	Nitrobenzene-d5	1.000	1.041	-4.1	161	0.00
6	Naphthalene	1.000	0.879	12.1	124	0.00
7	2-Methylnaphthalene	1.000	0.891	10.9	124	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	119	0.00
9 S	2-Fluorobiphenyl	1.000	1.382	-38.2#	179	0.00
10	Acenaphthylene	1.000	0.875	12.5	114	0.00
11	Acenaphthene	1.000	0.873	12.7	114	0.00
12	Fluorene	1.000	0.808	19.2	105	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	95	0.00
14	Hexachlorobenzene	1.000	0.903	9.7	94	0.00
15	Pentachlorophenol	1.000	1.321	-32.1#	135	0.00
16	Phenanthrene	1.000	0.881	11.9	93	0.00
17	Anthracene	1.000	0.861	13.9	89	0.00
18	Fluoranthene	1.000	0.779	22.1	80	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	64	0.00
20	Pyrene	1.000	0.956	4.4	78	0.00
21 S	Terphenyl-d14	1.000	1.510	-51.0#	125	0.00
22	Benzo(a)anthracene	1.000	0.860	14.0	64	0.00
23	Chrysene	1.000	0.883	11.7	62	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	1.381	-38.1#	124	0.00
25 I	Perylene-d12	5.000	5.000	0.0	86	0.00
26	Benzo(b)fluoranthene	1.000	1.158	-15.8	114	0.00
27	Benzo(k)fluoranthene	1.000	0.919	8.1	82	0.00
28 C	Benzo(a)pyrene	1.000	1.120	-12.0	103	0.00
29	Dibenzo(a,h)anthracene	1.000	1.249	-24.9	119	0.00
30	Benzo(g,h,i)perylene	1.000	1.155	-15.5	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0