

Method Path : Z:\HPCHEM1\BNA_E\METHODS\
 Method File : SIM-BE022615.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Feb 27 05:03:26 2015
 Response Via : Initial Calibration

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Calibration Files

0.1 =BE089700.D 0.2 =BE089701.D 0.5 =BE089702.D
 0.8 =BE089703.D 1 =BE089704.D 2 =BE089705.D

	Compound	0.1	0.2	0.5	0.8	1	2	Avg	%RSD

1) I	1,4-Dichlorobenzene-d	-----ISTD-----							
2)	n-Nitrosodimethyl	0.491	0.630	0.680	0.714	0.709	0.673	0.699	16.66
3) S	2-Fluorophenol	1.126	1.186	1.233	1.215	1.131	1.120	1.210	7.29
4) S	Phenol-d6	1.559	1.650	1.701	1.639	1.534	1.515	1.643	6.29
5) C	Phenol	1.727	1.806	1.848	1.808	1.681	1.651	1.794	5.63
6)	bis(2-Chloroethyl	1.382	1.469	1.481	1.435	1.308	1.282	1.407	5.39
7) I	Naphthalene-d8	-----ISTD-----							
8) S	Nitrobenzene-d5	0.289	0.459	0.236	0.282	0.234	0.244	0.295	24.82
9)	Nitrobenzene	0.234	0.261	0.276	0.298	0.279	0.293	0.295	14.86
10)	2,4-Dimethylpheno	0.207	0.212	0.225	0.226	0.213	0.218	0.229	10.15
11) C	2,4-Dichloropheno	0.139	0.154	0.160	0.164	0.154	0.162	0.168	14.16
12)	Hexachlorobutadie	0.099	0.102	0.099	0.095	0.087	0.085	0.095	6.17
13) I	Acenaphthene-d10	-----ISTD-----							
14) S	2,4,6-Tribromophe	0.036	0.038	0.039	0.042	0.040	0.049	0.051	38.54
15) S	2-Fluorobiphenyl	1.377	2.190	1.152	1.258	1.037	0.986	1.277	30.46
16) P	2,4-Dinitrophenol			0.008	0.011	0.012	0.014	0.019#	65.75
17) I	Phenanthrene-d10	-----ISTD-----							
18)	Hexachlorobenzene	0.131	0.138	0.137	0.130	0.119	0.117	0.130	6.11
19) C	Pentachlorophenol	0.017	0.012	0.014	0.014	0.015	0.020	0.022	62.82#
20) I	Chrysene-d12	-----ISTD-----							
21)	Benzidine	0.382	0.153	0.151	0.166	0.165	0.229	0.267	51.43
22) S	Terphenyl-d14	0.580	0.621	0.622	0.644	0.557	0.546	0.606	6.52
23)	Benzo(a)anthracen	3.476	3.314	3.520	3.414	3.182	3.275	3.553	10.40
24)	3,3'-Dichlorobenz	0.111	0.108	0.128	0.144	0.143	0.183	0.179	46.66
25)	Chrysene	4.760	4.742	4.656	4.454	3.968	3.811	4.372	8.04
26)	Indeno(1,2,3-cd)p	0.213	0.235	0.440	0.457	0.437	0.528	0.495	46.97
27) I	Perylene-d12	-----ISTD-----							
28)	Benzo(b)fluoranth	0.295	0.298	0.404	0.426	0.396	0.502	0.517	48.67
29)	Benzo(k)fluoranth	0.712	1.009	1.221	1.260	1.177	1.205	1.166	18.58
30) C	Benzo(a)pyrene	0.306	0.342	0.449	0.496	0.473	0.566	0.555	41.95#
31)	Dibenzo(a,h)anthr	0.183	0.220	0.369	0.415	0.403	0.510	0.467	51.98

(#)= Out of Range ### Number of calibration levels exceeded format ###									