

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE022615\
 Data File : BE089698.D
 Acq On : 26 Feb 2015 19:30
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001EC

Quant Time: Feb 27 01:17:28 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE022415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 25 14:06:35 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.34	152	59527	5.00	ng	0.00
4) Naphthalene-d8	12.24	136	254474	5.00	ng	0.00
8) Acenaphthene-d10	16.41	164	118367	5.00	ng	0.00
13) Phenanthrene-d10	19.76	188	218752	5.00	ng	0.00
19) Chrysene-d12	23.65	240	184858	5.00	ng	0.00
25) Perylene-d12	25.33	264	149783	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	10.60	82	11483	0.93	ng	0.00
9) 2-Fluorobiphenyl	14.87	172	24416	0.91	ng	0.00
21) Terphenyl-d14	22.30	244	22273	1.03	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.71	88	6227	0.83	ng	97
3) n-Nitrosodimethylamine	3.36	42	8205	0.89	ng	# 98
6) Naphthalene	12.28	128	45032	0.91	ng	99
7) 2-Methylnaphthalene	13.94	142	26494	0.95	ng	100
10) Acenaphthylene	16.06	152	155840	0.92	ng	100
11) Acenaphthene	16.49	154	24085	0.89	ng	99
12) Fluorene	17.74	166	27019	0.91	ng	100
14) Hexachlorobenzene	18.99	284	5037	0.89	ng	97
15) Pentachlorophenol	19.43	266	710	1.29	ng	98
16) Phenanthrene	19.81	178	42026	0.90	ng	100
17) Anthracene	19.90	178	36462	0.92	ng	100
18) Fluoranthene	21.72	202	39102	1.00	ng	98
20) Pyrene	22.03	202	41379	0.93	ng	100
22) Benzo(a)anthracene	23.64	228	130839	1.08	ng	96
23) Chrysene	23.68	228	145335	0.87	ng	98
24) Indeno(1,2,3-cd)pyrene	26.74	276	82696	1.50	ng	# 47
26) Benzo(b)fluoranthene	24.91	252	16900	1.57	ng	97
27) Benzo(k)fluoranthene	24.94	252	35195	1.00	ng	98
28) Benzo(a)pyrene	25.27	252	17333	1.36	ng	96
29) Dibenzo(a,h)anthracene	26.77	278	15765	1.51	ng	# 87
30) Benzo(g,h,i)perylene	27.15	276	77881	1.17	ng	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE022615\
Data File : BE089698.D
Acq On : 26 Feb 2015 19:30
Operator : TP/JJ
Sample : SSTDCCC001
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDCCC001EC

Quant Time: Feb 27 01:17:28 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE022415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 25 14:06:35 2015
Response via : Initial Calibration

