

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE022615\
 Data File : BE089702.D
 Acq On : 26 Feb 2015 22:47
 Operator : TP/JJ
 Sample : SSTDICC0.5
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.5

Manual Integrations
 APPROVED

apatel
 2/27/2015 4:12:59 PM

Quant Time: Feb 27 05:03:49 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE022615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 27 04:47:17 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.33	152	57953	5.00	ng	0.00
7) Naphthalene-d8	12.23	136	246963	5.00	ng	0.00
13) Acenaphthene-d10	16.40	164	113883	5.00	ng	0.00
17) Phenanthrene-d10	19.76	188	210524	5.00	ng	0.00
20) Chrysene-d12	23.65	240	170496	5.00	ng	0.00
27) Perylene-d12	25.34	264	136911	5.00	ng	0.00
System Monitoring Compounds						
3) 2-Fluorophenol	6.49	112	7147	0.40	ng	0.00
4) Phenol-d6	8.61	99	9858	0.45	ng	0.00
8) Nitrobenzene-d5	10.60	82	5839	0.35	ng	0.00
14) 2,4,6-Tribromophenol	18.30	330	447	0.19	ng	0.00
15) 2-Fluorobiphenyl	14.87	172	13117	0.39	ng	0.00
22) Terphenyl-d14	22.30	244	10605	0.56	ng	0.00
Target Compounds						
2) n-Nitrosodimethylamine	3.38	42	3939m	0.42	ng	Qvalue
5) Phenol	8.63	94	10709	0.46	ng	# 6
6) bis(2-Chloroethyl)ether	8.82	93	8585	0.49	ng	# 98
9) Nitrobenzene	10.64	77	6813	0.36	ng	99
10) 2,4-Dimethylphenol	11.57	122	5555	0.38	ng	99
11) 2,4-Dichlorophenol	11.95	162	3958	0.36	ng	95
12) Hexachlorobutadiene	12.62	225	2446	0.31	ng	99
16) 2,4-Dinitrophenol	16.66	184	96	0.72	ng	# 77
18) Hexachlorobenzene	18.98	284	2882	0.31	ng	98
19) Pentachlorophenol	19.43	266	287	0.40	ng	94
21) Benzidine	21.96	184	2581	0.20	ng	# 95
23) Benzo(a)anthracene	23.64	228	60017	0.39	ng	98
24) 3,3'-Dichlorobenzidine	23.63	252	2182	0.47	ng	99
25) Chrysene	23.68	228	79382	0.42	ng	100
26) Indeno(1,2,3-cd)pyrene	26.73	276	7504	0.37	ng	# 58
28) Benzo(b)fluoranthene	24.91	252	5528	0.46	ng	99
29) Benzo(k)fluoranthene	24.94	252	16715	0.44	ng	98
30) Benzo(a)pyrene	25.27	252	6150	0.53	ng	98
31) Dibenzo(a,h)anthracene	26.77	278	5048	0.52	ng	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE022615\
Data File : BE089702.D
Acq On : 26 Feb 2015 22:47
Operator : TP/JJ
Sample : SSTDICC0.5
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.5

Manual Integrations
APPROVED

apatel
2/27/2015 4:12:59 PM

Quant Time: Feb 27 05:03:49 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE022615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 27 04:47:17 2015
Response via : Initial Calibration

