

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE022615\
 Data File : BE089703.D
 Acq On : 26 Feb 2015 23:26
 Operator : TP/JJ
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.8

Manual Integrations
 APPROVED

apatel
 2/27/2015 4:13:03 PM

Quant Time: Feb 27 05:04:47 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE022615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 27 04:47:17 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.33	152	57340	5.00	ng	0.00
7) Naphthalene-d8	12.23	136	243752	5.00	ng	0.00
13) Acenaphthene-d10	16.41	164	110284	5.00	ng	0.00
17) Phenanthrene-d10	19.76	188	200863	5.00	ng	0.00
20) Chrysene-d12	23.65	240	161364	5.00	ng	0.00
27) Perylene-d12	25.33	264	126012	5.00	ng	0.00

System Monitoring Compounds

3) 2-Fluorophenol	6.49	112	11144	0.64	ng	0.00
4) Phenol-d6	8.61	99	15035	0.69	ng	0.00
8) Nitrobenzene-d5	10.60	82	11006	0.66	ng	0.00
14) 2,4,6-Tribromophenol	18.30	330	734	0.32	ng	0.00
15) 2-Fluorobiphenyl	14.87	172	22201	0.68	ng	0.00
22) Terphenyl-d14	22.30	244	16630	0.94	ng	0.00

Target Compounds

						Qvalue
2) n-Nitrosodimethylamine	3.37	42	6550m	0.71	ng	
5) Phenol	8.64	94	16584	0.72	ng	# 57
6) bis(2-Chloroethyl)ether	8.82	93	13169	0.76	ng	# 97
9) Nitrobenzene	10.64	77	11627	0.62	ng	100
10) 2,4-Dimethylphenol	11.57	122	8833	0.61	ng	98
11) 2,4-Dichlorophenol	11.95	162	6379	0.58	ng	99
12) Hexachlorobutadiene	12.62	225	3722	0.48	ng	100
16) 2,4-Dinitrophenol	16.65	184	195	0.80	ng	93
18) Hexachlorobenzene	18.98	284	4176	0.48	ng	92
19) Pentachlorophenol	19.43	266	456	0.49	ng	95
21) Benzidine	21.95	184	4277	0.34	ng	98
23) Benzo(a)anthracene	23.64	228	88150	0.61	ng	99
24) 3,3'-Dichlorobenzidine	23.63	252	3714	0.56	ng	98
25) Chrysene	23.68	228	114991	0.65	ng	99
26) Indeno(1,2,3-cd)pyrene	26.74	276	11808	0.48	ng	# 48
28) Benzo(b)fluoranthene	24.91	252	8588	0.61	ng	99
29) Benzo(k)fluoranthene	24.94	252	25394	0.72	ng	99
30) Benzo(a)pyrene	25.27	252	9994	0.67	ng	99
31) Dibenzo(a,h)anthracene	26.77	278	8365	0.63	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE022615\
Data File : BE089703.D
Acq On : 26 Feb 2015 23:26
Operator : TP/JJ
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.8

Manual Integrations
APPROVED

apatel
2/27/2015 4:13:03 PM

Quant Time: Feb 27 05:04:47 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE022615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 27 04:47:17 2015
Response via : Initial Calibration

