

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE022615\
 Data File : BE089708.D
 Acq On : 27 Feb 2015 2:43
 Operator : TP/JJ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE022615

Quant Time: Feb 27 13:54:10 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE022615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 27 13:38:44 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.33	152	48545	5.00	ng	0.00
7) Naphthalene-d8	12.23	136	206581	5.00	ng	0.00
13) Acenaphthene-d10	16.40	164	94957	5.00	ng	0.00
17) Phenanthrene-d10	19.76	188	176653	5.00	ng	0.00
20) Chrysene-d12	23.65	240	149230	5.00	ng	0.00
27) Perylene-d12	25.34	264	118051	5.00	ng	0.00

System Monitoring Compounds						
3) 2-Fluorophenol	6.49	112	12634	1.08	ng	0.00
4) Phenol-d6	8.61	99	16981	1.06	ng	0.00
8) Nitrobenzene-d5	10.60	82	11153	1.00	ng	0.00
14) 2,4,6-Tribromophenol	18.31	330	886	0.93	ng	0.00
15) 2-Fluorobiphenyl	14.87	172	22786	1.07	ng	0.00
22) Terphenyl-d14	22.30	244	23178	1.28	ng	0.00

Target Compounds						Qvalue
2) n-Nitrosodimethylamine	3.35	42	7062	1.00	ng	95
5) Phenol	8.63	94	18671	1.07	ng	# 20
6) bis(2-Chloroethyl)ether	8.82	93	14642	1.07	ng	# 98
9) Nitrobenzene	10.64	77	13134	1.08	ng	99
10) 2,4-Dimethylphenol	11.57	122	10125	1.07	ng	99
11) 2,4-Dichlorophenol	11.95	162	7384	1.07	ng	99
12) Hexachlorobutadiene	12.62	225	4157	1.06	ng	99
16) 2,4-Dinitrophenol	16.65	184	224	1.02	ng	85
18) Hexachlorobenzene	18.98	284	4816	1.05	ng	97
19) Pentachlorophenol	19.43	266	602	1.11	ng	97
21) Benzidine	21.96	184	5905	0.92	ng	99
23) Benzo(a)anthracene	23.64	228	110638	1.04	ng	# 92
24) 3,3'-Dichlorobenzidine	23.63	252	5247	0.95	ng	100
25) Chrysene	23.68	228	140165	1.07	ng	99
26) Indeno(1,2,3-cd)pyrene	26.74	276	15533	0.96	ng	# 48
28) Benzo(b)fluoranthene	24.91	252	11667	0.92	ng	99
29) Benzo(k)fluoranthene	24.93	252	32425	1.09	ng	99
30) Benzo(a)pyrene	25.27	252	13251	0.95	ng	99
31) Dibenzo(a,h)anthracene	26.77	278	11576	0.95	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE022615\
Data File : BE089708.D
Acq On : 27 Feb 2015 2:43
Operator : TP/JJ
Sample : SSTDICV001
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE022615

Quant Time: Feb 27 13:54:10 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE022615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 27 13:38:44 2015
Response via : Initial Calibration

