

Data Path : Z:\HPCHEM1\BNA_E\Data\BE022615\
 Data File : BE089692.D
 Acq On : 26 Feb 2015 12:50
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 27 01:16:05 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE022415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 25 14:06:35 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	125	0.00
2	1,4-Dioxane	1.000	0.944	5.6	109	-0.07
3	n-Nitrosodimethylamine	1.000	0.972	2.8	109	-0.05
4 I	Naphthalene-d8	5.000	5.000	0.0	140	0.00
5 S	Nitrobenzene-d5	1.000	0.752	24.8	115	0.00
6	Naphthalene	1.000	0.925	7.5	140	0.00
7	2-Methylnaphthalene	1.000	1.007	-0.7	150	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	158	0.00
9 S	2-Fluorobiphenyl	1.000	0.907	9.3	158	0.00
10	Acenaphthylene	1.000	0.906	9.4	156	0.00
11	Acenaphthene	1.000	0.918	8.2	159	0.00
12	Fluorene	1.000	0.913	8.7	157	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	142	0.00
14	Hexachlorobenzene	1.000	0.974	2.6	152	0.00
15	Pentachlorophenol	1.000	3.686	-268.6#	962	0.00
16	Phenanthrene	1.000	0.898	10.2	141	0.00
17	Anthracene	1.000	0.898	10.2	139	0.00
18	Fluoranthene	1.000	0.796	20.4	122	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	82	0.00
20	Pyrene	1.000	1.169	-16.9	122	0.00
21 S	Terphenyl-d14	1.000	1.179	-17.9	125	0.00
22	Benzo(a)anthracene	1.000	0.868	13.2	83	0.00
23	Chrysene	1.000	0.920	8.0	83	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	1.076	-7.6	115	0.00
25 I	Perylene-d12	5.000	5.000	0.0	96	0.00
26	Benzo(b)fluoranthene	1.000	1.094	-9.4	117	0.00
27	Benzo(k)fluoranthene	1.000	0.896	10.4	89	0.00
28 C	Benzo(a)pyrene	1.000	1.037	-3.7	103	0.00
29	Dibenzo(a,h)anthracene	1.000	1.176	-17.6	122	0.00
30	Benzo(g,h,i)perylene	1.000	1.051	-5.1	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0