

Data Path : Z:\HPCHEM1\BNA_E\Data\BE022615\
 Data File : BE089692.D
 Acq On : 26 Feb 2015 12:50
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 27 01:16:05 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE022415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 25 14:06:35 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	125	0.00
2	1,4-Dioxane	0.627	0.592	5.6	109	-0.07
3	n-Nitrosodimethylamine	0.766	0.759	0.9	109	-0.05
4 I	Naphthalene-d8	1.000	1.000	0.0	140	0.00
5 S	Nitrobenzene-d5	0.282	0.171	39.4#	115	0.00
6	Naphthalene	0.968	0.895	7.5	140	0.00
7	2-Methylnaphthalene	0.550	0.554	-0.7	150#	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	158#	0.00
9 S	2-Fluorobiphenyl	1.295	1.030	20.5	158#	0.00
10	Acenaphthylene	7.176	6.500	9.4	156#	0.00
11	Acenaphthene	1.137	1.044	8.2	159#	0.00
12	Fluorene	1.261	1.151	8.7	157#	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	142	0.00
14	Hexachlorobenzene	0.129	0.126	2.3	152#	0.00
15	Pentachlorophenol	0.014	0.080	-471.4#	962#	0.00
16	Phenanthrene	1.067	0.958	10.2	141	0.00
17	Anthracene	0.908	0.815	10.2	139	0.00
18	Fluoranthene	0.890	0.709	20.3	122	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	82	0.00
20	Pyrene	1.208	1.413	-17.0	122	0.00
21 S	Terphenyl-d14	0.587	0.692	-17.9	125	0.00
22	Benzo(a)anthracene	3.288	2.854	13.2	83	0.00
23	Chrysene	4.532	4.170	8.0	83	0.00
24	Indeno(1,2,3-cd)pyrene	1.340	1.454	-8.5	115	0.00
25 I	Perylene-d12	1.000	1.000	0.0	96	0.00
26	Benzo(b)fluoranthene	0.289	0.349	-20.8	117	0.00
27	Benzo(k)fluoranthene	0.999	1.045	-4.6	89	0.00
28 C	Benzo(a)pyrene	0.416	0.407	2.2	103	0.00
29	Dibenzo(a,h)anthracene	0.243	0.371	-52.7#	122	0.00
30	Benzo(g,h,i)perylene	1.814	2.313	-27.5#	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0