

Data Path : Z:\HPCHEM1\BNA_E\Data\BE022615\
 Data File : BE089698.D
 Acq On : 26 Feb 2015 19:30
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 27 01:17:28 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE022415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 25 14:06:35 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	158	0.00
2	1,4-Dioxane	1.000	0.835	16.5	122	0.00
3	n-Nitrosodimethylamine	1.000	0.890	11.0	126	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	158	0.00
5 S	Nitrobenzene-d5	1.000	0.932	6.8	170	0.00
6	Naphthalene	1.000	0.914	8.6	156	0.00
7	2-Methylnaphthalene	1.000	0.947	5.3	160	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	162	0.00
9 S	2-Fluorobiphenyl	1.000	0.908	9.2	161	0.00
10	Acenaphthylene	1.000	0.917	8.3	162	0.00
11	Acenaphthene	1.000	0.895	10.5	158	0.00
12	Fluorene	1.000	0.905	9.5	159	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	164	0.00
14	Hexachlorobenzene	1.000	0.890	11.0	160	0.00
15	Pentachlorophenol	1.000	1.294	-29.4#	225	0.00
16	Phenanthrene	1.000	0.900	10.0	163	0.00
17	Anthracene	1.000	0.918	8.2	163	0.00
18	Fluoranthene	1.000	1.004	-0.4	177	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	146	0.00
20	Pyrene	1.000	0.926	7.4	172	0.00
21 S	Terphenyl-d14	1.000	1.027	-2.7	194	0.00
22	Benzo(a)anthracene	1.000	1.076	-7.6	184	0.00
23	Chrysene	1.000	0.867	13.3	139	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	1.498	-49.8#	316	0.00
25 I	Perylene-d12	5.000	5.000	0.0	191	0.00
26	Benzo(b)fluoranthene	1.000	1.573	-57.3#	378	0.00
27	Benzo(k)fluoranthene	1.000	0.999	0.1	198	0.00
28 C	Benzo(a)pyrene	1.000	1.361	-36.1#	292	0.00
29	Dibenzo(a,h)anthracene	1.000	1.509	-50.9#	343	0.00
30	Benzo(g,h,i)perylene	1.000	1.170	-17.0	246	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1