

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE022615\
 Data File : BE089705.D
 Acq On : 27 Feb 2015 00:43
 Operator : TP/JJ
 Sample : SSTDICC002
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC002

Quant Time: Feb 27 04:57:47 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE022615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 27 04:47:17 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.33	152	61621	5.00	ng	0.00
7) Naphthalene-d8	12.23	136	266014	5.00	ng	0.00
13) Acenaphthene-d10	16.41	164	124860	5.00	ng	0.00
17) Phenanthrene-d10	19.76	188	222869	5.00	ng	0.00
20) Chrysene-d12	23.65	240	182075	5.00	ng	0.00
27) Perylene-d12	25.33	264	142973	5.00	ng	0.00

System Monitoring Compounds						
3) 2-Fluorophenol	6.49	112	27609	1.46	ng	0.00
4) Phenol-d6	8.61	99	37336	1.60	ng	0.00
8) Nitrobenzene-d5	10.60	82	25951	1.43	ng	0.00
14) 2,4,6-Tribromophenol	18.30	330	2432	0.94	ng	0.00
15) 2-Fluorobiphenyl	14.87	172	49255	1.34	ng	0.00
22) Terphenyl-d14	22.30	244	39788	1.98	ng	0.00


Target Compounds				Qvalue		
2) n-Nitrosodimethylamine	3.34	42	16579	1.66	ng	# 25
5) Phenol	8.63	94	40689	1.64	ng	# 25
6) bis(2-Chloroethyl)ether	8.82	93	31593	1.70	ng	# 99
9) Nitrobenzene	10.64	77	31125	1.52	ng	99
10) 2,4-Dimethylphenol	11.57	122	23216	1.48	ng	99
11) 2,4-Dichlorophenol	11.95	162	17234	1.44	ng	98
12) Hexachlorobutadiene	12.62	225	9080	1.08	ng	100
16) 2,4-Dinitrophenol	16.65	184	700	1.12	ng	95
18) Hexachlorobenzene	18.98	284	10415	1.07	ng	97
19) Pentachlorophenol	19.43	266	1771	1.00	ng	94
21) Benzidine	21.96	184	16681	1.18	ng	98
23) Benzo(a)anthracene	23.63	228	238537	1.46	ng	# 91
24) 3,3'-Dichlorobenzidine	23.63	252	13327	1.01	ng	97
25) Chrysene	23.68	228	277562	1.39	ng	99
26) Indeno(1,2,3-cd)pyrene	26.74	276	38434	0.98	ng	98
28) Benzo(b)fluoranthene	24.91	252	28699	1.34	ng	99
29) Benzo(k)fluoranthene	24.94	252	68902	1.73	ng	99
30) Benzo(a)pyrene	25.27	252	32352	1.26	ng	99
31) Dibenzo(a,h)anthracene	26.77	278	29169	1.13	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE022615\
Data File : BE089705.D
Acq On : 27 Feb 2015 00:43
Operator : TP/JJ
Sample : SSTDICC002
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICC002

Quant Time: Feb 27 04:57:47 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE022615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 27 04:47:17 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 9.33 min Scan# 1104

Delta R.T. 0.00 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43

Instrument :

BNA_E

ClientSampleId :

SSTDICC002


Tgt Ion: 152 Resp: 61621

Ion Ratio Lower Upper

152 100

150 154.7 127.3 190.9




115 62.9 52.0 78.0



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

n-Nitrosodimethylamine

Concen: 1.66 ng

RT: 3.34 min Scan# 286

Delta R.T. -0.02 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43


Tgt Ion: 42 Resp: 16579

Ion Ratio Lower Upper

42 100

74 16.3 83.8 125.8#



44 12.0 19.0 28.6#




Abundance Ion 42.00 (41.70 to 42.70): BE0

Ion 74.00 (73.70 to 74.70): BE0

Ion 44.00 (43.70 to 44.70): BE0





#3

2-Fluorophenol

Concen: 1.46 ng

RT: 6.49 min Scan# 751

Delta R.T. -0.00 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43

Instrument :

BNA_E

ClientSampleId :

SSTDICC002


Tgt Ion: 112 Resp: 27609

Ion Ratio Lower Upper

112 100

64 63.9 51.3 76.9




63 32.2 26.2 39.2



Abundance Ion 112.00 (111.70 to 112.70): BE0

Ion 64.00 (63.70 to 64.70): BE0

Ion 63.00 (62.70 to 63.70): BE0



#4

Phenol-d6

Concen: 1.60 ng

RT: 8.61 min Scan# 1022

Delta R.T. -0.01 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43


Tgt Ion: 99 Resp: 37336

Ion Ratio Lower Upper

99 100

42 22.8 12.3 18.5#



71 34.3 22.9 34.3




Abundance Ion 99.00 (98.70 to 99.70): BE0

Ion 42.00 (41.70 to 42.70): BE0

Ion 71.00 (70.70 to 71.70): BE0





#5

Phenol

Concen: 1.64 ng

RT: 8.63 min Scan# 1025

Delta R.T. -0.01 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43

Instrument :

BNA_E

ClientSampleId :

SSTDICC002


Tgt Ion: 94 Resp: 40689

Ion Ratio Lower Upper




94 100

65 135.4 31.7 71.7#

66 46.7 17.9 57.9



Abundance Ion 94.00 (93.70 to 94.70): BE0
Ion 65.00 (64.70 to 65.70): BE0
Ion 66.00 (65.70 to 66.70): BE0



#6

bis(2-Chloroethyl)ether

Concen: 1.70 ng

RT: 8.82 min Scan# 1046

Delta R.T. 0.00 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43


Tgt Ion: 93 Resp: 31593

Ion Ratio Lower Upper



93 100


63 103.3 83.8 125.8

95 0.0 0.0 0.0



Abundance Ion 93.00 (92.70 to 93.70): BE0
Ion 63.00 (62.70 to 63.70): BE0
Ion 95.00 (94.70 to 95.70): BE0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 12.23 min Scan# 1496

Delta R.T. 0.00 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43

Instrument :

BNA_E

ClientSampleId :

SSTDICC002


Tgt Ion: 136 Resp: 266014

Ion Ratio Lower Upper




136 100

137 10.9 8.6 13.0

54 10.3 8.1 12.1



Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BE0



#8

Nitrobenzene-d5

Concen: 1.43 ng

RT: 10.60 min Scan# 1284

Delta R.T. 0.00 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43


Tgt Ion: 82 Resp: 25951

Ion Ratio Lower Upper



82 100


128 40.3 31.4 47.2

54 51.8 43.8 65.8



Abundance Ion 82.00 (81.70 to 82.70): BE0
Ion 128.00 (127.70 to 128.70): E
Ion 54.00 (53.70 to 54.70): BE0





#9

Nitrobenzene

Concen: 1.52 ng

RT: 10.64 min Scan# 1293

Delta R.T. -0.00 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43

Instrument :

BNA_E

ClientSampled :

SSTDICC002


Tgt Ion: 77 Resp: 31125

Ion Ratio Lower Upper




77 100

123 42.4 33.2 49.8

65 13.7 10.6 15.8



Abundance Ion 77.00 (76.70 to 77.70): BE0
Ion 123.00 (122.70 to 123.70): BE0
Ion 65.00 (64.70 to 65.70): BE0



#10

2,4-Dimethylphenol

Concen: 1.48 ng

RT: 11.57 min Scan# 1428

Delta R.T. 0.00 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43


Tgt Ion: 122 Resp: 23216

Ion Ratio Lower Upper



122 100


107 108.2 87.6 131.4

121 53.7 43.4 65.2



Abundance Ion 122.00 (121.70 to 122.70): E
Ion 107.00 (106.70 to 107.70): E
Ion 121.00 (120.70 to 121.70): E









#11
 2,4-Dichlorophenol
 Concen: 1.44 ng
 RT: 11.95 min Scan# 1467
 Delta R.T. 0.00 min
 Lab File: BE089705.D
 Acq: 27 Feb 2015 00:43

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC002

Tgt Ion:162 Resp: 17234
 Ion Ratio Lower Upper
 162 100
 164 64.8 43.9 83.9
 98 33.4 15.7 55.7




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): E






#12
 Hexachlorobutadiene
 Concen: 1.08 ng
 RT: 12.62 min Scan# 1536
 Delta R.T. 0.00 min
 Lab File: BE089705.D
 Acq: 27 Feb 2015 00:43

Tgt Ion:225 Resp: 9080
 Ion Ratio Lower Upper
 225 100
 223 63.1 50.6 75.8
 227 63.1 50.8 76.2



Abundance Ion 225.00 (224.70 to 225.70): E
 Ion 223.00 (222.70 to 223.70): E
 Ion 227.00 (226.70 to 227.70): E





#13

Acenaphthene-d10

Concen: 5.00 ng

RT: 16.41 min Scan# 2802

Delta R.T. 0.00 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43

Instrument :

BNA_E

ClientSampleId :

SSTDICC002


Tgt Ion:164 Resp: 124860

Ion Ratio Lower Upper

164 100

162 93.2 75.0 112.6




160 42.3 34.3 51.5



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#14

2,4,6-Tribromophenol

Concen: 0.94 ng

RT: 18.30 min Scan# 3081

Delta R.T. 0.00 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43


Tgt Ion:330 Resp: 2432

Ion Ratio Lower Upper

330 100

332 94.6 0.0 0.0#



141 73.1 0.0 0.0#




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#15
2-Fluorobiphenyl
Concen: 1.34 ng
RT: 14.87 min Scan# 2371
Delta R.T. -0.00 min
Lab File: BE089705.D
Acq: 27 Feb 2015 00:43




Instrument :
BNA_E
ClientSampleId :
SSTDICC002


Tgt Ion:172 Resp: 49255
Ion Ratio Lower Upper
172 100
171 33.4 26.8 40.2
170 23.1 18.0 27.0



#16
2,4-Dinitrophenol
Concen: 1.12 ng
RT: 16.65 min Scan# 2837
Delta R.T. -0.01 min
Lab File: BE089705.D
Acq: 27 Feb 2015 00:43

Tgt Ion:184 Resp: 700
Ion Ratio Lower Upper
184 100
63 100.7 82.4 123.6
154 69.6 61.4 92.2









#17
Phenanthrene-d10
Concen: 5.00 ng
RT: 19.76 min Scan# 3293
Delta R.T. 0.00 min
Lab File: BE089705.D
Acq: 27 Feb 2015 00:43




Instrument :
BNA_E
ClientSampleId :
SSTDICC002


Tgt Ion:188 Resp: 222869
Ion Ratio Lower Upper
188 100
94 14.9 12.1 18.1
80 15.6 12.7 19.1



#18
Hexachlorobenzene
Concen: 1.07 ng
RT: 18.98 min Scan# 3178
Delta R.T. 0.00 min
Lab File: BE089705.D
Acq: 27 Feb 2015 00:43

Tgt Ion:284 Resp: 10415
Ion Ratio Lower Upper
284 100
142 68.4 52.2 78.4
249 34.3 26.4 39.6





#19

Pentachlorophenol

Concen: 1.00 ng

RT: 19.43 min Scan# 3243

Delta R.T. 0.00 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43

Instrument :

BNA_E

ClientSampleId :

SSTDICC002


Tgt Ion:266 Resp: 1771

Ion Ratio Lower Upper




266 100

268 61.4 51.4 77.2

264 63.4 55.8 83.8



Abundance Ion 265.70 (265.40 to 266.40): E
Ion 268.00 (267.70 to 268.70): E
Ion 264.00 (263.70 to 264.70): E



#20

Chrysene-d12

Concen: 5.00 ng

RT: 23.65 min Scan# 3931

Delta R.T. 0.00 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43


Tgt Ion:240 Resp: 182075

Ion Ratio Lower Upper



240 100


120 12.2 9.9 14.9

236 26.1 21.0 31.6



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E





#21

Benzidine

Concen: 1.18 ng

RT: 21.96 min Scan# 3660

Delta R.T. 0.00 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43

Instrument :

BNA_E

ClientSampleId :

SSTDICC002


Tgt Ion:184 Resp: 16681

Ion Ratio Lower Upper

184 100

185 14.9 12.8 19.2




183 10.7 8.8 13.2



Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



#22

Terphenyl-d14

Concen: 1.98 ng

RT: 22.30 min Scan# 3734

Delta R.T. 0.00 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43


Tgt Ion:244 Resp: 39788

Ion Ratio Lower Upper

244 100

212 8.3 6.7 10.1



122 18.8 15.7 23.5




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#23

Benzo(a)anthracene

Concen: 1.46 ng

RT: 23.63 min Scan# 3929

Delta R.T. -0.01 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43

Instrument :

BNA_E

ClientSampleId :

SSTDICC002


Tgt Ion:228 Resp: 238537

Ion Ratio Lower Upper

228 100

226 29.3 19.3 28.9#




229 17.6 16.6 25.0



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#24

3,3'-Dichlorobenzidine

Concen: 1.01 ng

RT: 23.63 min Scan# 3928

Delta R.T. 0.00 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43


Tgt Ion:252 Resp: 13327

Ion Ratio Lower Upper

252 100

254 60.7 49.4 74.2



126 19.0 17.4 26.2




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E





#25

Chrysene

Concen: 1.39 ng

RT: 23.68 min Scan# 3934

Delta R.T. 0.00 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43

Instrument :

BNA_E

ClientSampleId :

SSTDICC002


Tgt Ion:228 Resp: 277562

Ion Ratio Lower Upper

228 100

226 31.5 24.7 37.1




229 17.7 14.4 21.6



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#26

Indeno(1,2,3-cd)pyrene

Concen: 0.98 ng

RT: 26.74 min Scan# 4545

Delta R.T. 0.01 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43


Tgt Ion:276 Resp: 38434

Ion Ratio Lower Upper

276 100

138 39.3 29.8 44.6



277 24.9 19.8 29.6




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#27
Perylene-d12
 Concen: 5.00 ng
 RT: 25.33 min Scan# 4239
 Delta R.T. 0.00 min
 Lab File: BE089705.D
 Acq: 27 Feb 2015 00:43




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC002


Tgt Ion: 264 Resp: 142973
 Ion Ratio Lower Upper
 264 100
 260 24.9 20.2 30.2
 265 21.2 16.9 25.3



#28
Benzo(b)fluoranthene
 Concen: 1.34 ng
 RT: 24.91 min Scan# 4145
 Delta R.T. -0.00 min
 Lab File: BE089705.D
 Acq: 27 Feb 2015 00:43

Tgt Ion: 252 Resp: 28699
 Ion Ratio Lower Upper
 252 100
 253 21.2 17.8 26.6
 125 16.4 12.9 19.3





#29

Benzo(k)fluoranthene

Concen: 1.73 ng

RT: 24.94 min Scan# 4152

Delta R.T. 0.00 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43

Instrument :

BNA_E

ClientSampleId :

SSTDICC002


Tgt Ion:252 Resp: 68902

Ion Ratio Lower Upper

252 100

253 22.0 17.3 25.9




125 15.8 12.9 19.3



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#30

Benzo(a)pyrene

Concen: 1.26 ng

RT: 25.27 min Scan# 4225

Delta R.T. 0.00 min

Lab File: BE089705.D

Acq: 27 Feb 2015 00:43


Tgt Ion:252 Resp: 32352

Ion Ratio Lower Upper

252 100

253 21.8 17.9 26.9



125 17.3 14.0 21.0




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#31

Dibenzo(a,h)anthracene

Concen: 1.13 ng

RT: 26.77 min Scan# 4550

Delta R.T. -0.00 min

Lab File: BE089705.D


Acq: 27 Feb 2015 00:43

Instrument :

BNA_E

ClientSampleId :

SSTDIC002




Tgt Ion: 278 Resp: 29169

Ion Ratio Lower Upper

278 100

139 28.1 21.1 31.7

279 24.1 19.3 28.9



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E

