

Data Path : Z:\HPCHEM1\BNA_E\Data\BE022618\
 Data File : BE095695.D
 Acq On : 26 Feb 2018 18:54
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 27 00:46:37 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE022318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 23 12:12:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	-0.01
2	1,4-Dioxane	0.630	0.657	-4.3	123	0.00
3	n-Nitrosodimethylamine	0.781	0.788	-0.9	119	-0.01
4 S	2-Fluorophenol	1.056	1.079	-2.2	121	0.00
5 S	Phenol-d6	1.462	1.535	-5.0	121	0.00
6	bis(2-Chloroethyl)ether	1.500	1.568	-4.5	120	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	119	0.00
8 S	Nitrobenzene-d5	0.412	0.402	2.4	119	0.00
9	Naphthalene	4.611	4.649	-0.8	117	0.00
10	Hexachlorobutadiene	0.269	0.274	-1.9	114	0.00
11 SURR	2-Methylnaphthalene-d10	0.669	0.679	-1.5	119	0.00
12	2-Methylnaphthalene	0.796	0.803	-0.9	119	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	130	0.00
14 S	2,4,6-Tribromophenol	0.143	0.141	1.4	143	0.00
15 S	2-Fluorobiphenyl	1.787	1.679	6.0	120	0.00
16	Acenaphthylene	2.247	2.080	7.4	121	0.00
17	Acenaphthene	1.411	1.343	4.8	122	0.00
18	Fluorene	1.750	1.662	5.0	123	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	131	0.00
20	4-Bromophenyl-phenylether	0.251	0.240	4.4	123	0.00
21	Hexachlorobenzene	0.260	0.243	6.5	121	0.00
22	Pentachlorophenol	0.060	0.057	5.0	152#	-0.01
23	Phenanthrene	1.250	1.200	4.0	122	0.00
24	Anthracene	1.148	1.084	5.6	126	0.00
25 SURR	Fluoranthene-d10	5.225	4.952	5.2	124	0.00
26	Fluoranthene	1.538	1.444	6.1	122	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	128	0.00
28	Pyrene	1.636	1.588	2.9	125	0.00
29 S	Terphenyl-d14	0.797	0.776	2.6	124	0.00
30	Benzo(a)anthracene	1.305	1.268	2.8	125	0.00
31	Chrysene	1.298	1.240	4.5	122	0.00
32	Bis(2-ethylhexyl)phthalate	2.308	2.264	1.9	139	0.00
33	Indeno(1,2,3-cd)pyrene	1.257	1.286	-2.3	131	0.00
34 I	Perylene-d12	1.000	1.000	0.0	134	0.00
35	Benzo(b)fluoranthene	1.425	1.347	5.5	124	0.00
36	Benzo(k)fluoranthene	1.417	1.348	4.9	126	0.00
37 C	Benzo(a)pyrene	1.284	1.249	2.7	129	0.00
38	Dibenzo(a,h)anthracene	1.143	1.121	1.9	132	0.00
39	Benzo(g,h,i)perylene	1.224	1.155	5.6	125	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE022618\
Data File : BE095695.D
Acq On : 26 Feb 2018 18:54
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Feb 27 00:46:37 2018
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE022318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 23 12:12:42 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			