

Data Path : Z:\HPCHEM1\BNA E\Data\BE022718\
 Data File : BE095697.D
 Acq On : 27 Feb 2018 11:01
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 27 12:34:53 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE022318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 23 12:12:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2	1,4-Dioxane	0.630	0.546	13.3	104	0.00
3	n-Nitrosodimethylamine	0.781	0.725	7.2	112	0.00
4 S	2-Fluorophenol	1.056	0.973	7.9	111	0.00
5 S	Phenol-d6	1.462	1.359	7.0	109	0.00
6	bis(2-Chloroethyl)ether	1.500	1.395	7.0	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	121	-0.02
8 S	Nitrobenzene-d5	0.412	0.361	12.4	109	-0.02
9	Naphthalene	4.611	4.165	9.7	107	0.00
10	Hexachlorobutadiene	0.269	0.245	8.9	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.669	0.607	9.3	109	0.00
12	2-Methylnaphthalene	0.796	0.715	10.2	108	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	123	0.00
14 S	2,4,6-Tribromophenol	0.143	0.131	8.4	127	0.00
15 S	2-Fluorobiphenyl	1.787	1.587	11.2	108	0.00
16	Acenaphthylene	2.247	2.007	10.7	111	0.00
17	Acenaphthene	1.411	1.254	11.1	108	0.00
18	Fluorene	1.750	1.589	9.2	112	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	125	0.01
20	4-Bromophenyl-phenylether	0.251	0.223	11.2	109	0.00
21	Hexachlorobenzene	0.260	0.229	11.9	109	0.00
22	Pentachlorophenol	0.060	0.057	5.0	147	-0.02
23	Phenanthrene	1.250	1.116	10.7	108	0.00
24	Anthracene	1.148	0.996	13.2	111	0.00
25 SURR	Fluoranthene-d10	5.225	4.646	11.1	111	0.00
26	Fluoranthene	1.538	1.362	11.4	110	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	122	0.00
28	Pvrene	1.636	1.574	3.8	117	0.00
29 S	Terphenyl-d14	0.797	0.731	8.3	110	0.00
30	Benzo(a)anthracene	1.305	1.215	6.9	114	0.00
31	Chrysene	1.298	1.196	7.9	111	0.00
32	Bis(2-ethylhexyl)phthalate	2.308	2.177	5.7	126	0.00
33	Indeno(1,2,3-cd)pyrene	1.257	1.151	8.4	111	0.00
34 I	Perylene-d12	1.000	1.000	0.0	124	0.00
35	Benzo(b)fluoranthene	1.425	1.262	11.4	107	0.00
36	Benzo(k)fluoranthene	1.417	1.285	9.3	111	0.00
37 C	Benzo(a)pyrene	1.284	1.157	9.9	110	0.00
38	Dibenzo(a,h)anthracene	1.143	1.032	9.7	112	0.00
39	Benzo(a,h,i)perylene	1.224	1.099	10.2	109	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE022718\
Data File : BE095697.D
Acq On : 27 Feb 2018 11:01
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Feb 27 12:34:53 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE022318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 23 12:12:42 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			