

Data Path : Z:\HPCHEM1\BNA E\DATA\BE022817\  
 Data File : BE092762.D  
 Acq On : 28 Feb 2017 14:08  
 Operator : SJ/MA  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 SSTDCCC0.4

Quant Time: Mar 01 08:05:28 2017  
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE021517.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Feb 15 17:05:20 2017  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|---------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 8.10  | 152  | 8027     | 5.00 | ng    | -0.02    |
| 7) Naphthalene-d8         | 10.87 | 136  | 36298    | 5.00 | ng    | -0.02    |
| 12) Acenaphthene-d10      | 14.75 | 164  | 25677    | 5.00 | ng    | -0.02    |
| 18) Phenanthrene-d10      | 17.51 | 188  | 62780    | 5.00 | ng    | 0.00     |
| 24) Chrysene-d12          | 21.68 | 240  | 85174    | 5.00 | ng    | -0.02    |
| 31) Perylene-d12          | 24.01 | 264  | 74485    | 5.00 | ng    | -0.02    |


|                             |       |     |      |      |    |       |
|-----------------------------|-------|-----|------|------|----|-------|
| System Monitoring Compounds |       |     |      |      |    |       |
| 4) 2-Fluorophenol           | 5.62  | 112 | 625  | 0.40 | ng | -0.02 |
| 5) Phenol-d6                | 7.31  | 99  | 699  | 0.36 | ng | 0.00  |
| 8) Nitrobenzene-d5          | 9.16  | 82  | 105  | 0.12 | ng | -0.16 |
| 13) 2,4,6-Tribromophenol    | 16.26 | 330 | 234  | 0.43 | ng | -0.02 |
| 14) 2-Fluorobiphenyl        | 13.39 | 172 | 2290 | 0.37 | ng | -0.01 |
| 26) Terphenyl-d14           | 20.12 | 244 | 4027 | 0.37 | ng | -0.02 |


|                                |       |     |       |      |    |        |
|--------------------------------|-------|-----|-------|------|----|--------|
| Target Compounds               |       |     |       |      |    | Qvalue |
| 2) 1,4-Dioxane                 | 3.33  | 88  | 568   | 0.52 | ng | # 81   |
| 3) n-Nitrosodimethylamine      | 3.71  | 42  | 453   | 0.50 | ng | # 91   |
| 6) bis(2-Chloroethyl)ether     | 7.52  | 93  | 788   | 0.37 | ng | # 69   |
| 9) Naphthalene                 | 10.93 | 128 | 3182  | 0.40 | ng | 98     |
| 10) Hexachlorobutadiene        | 11.22 | 225 | 472   | 0.41 | ng | 99     |
| 11) 2-Methylnaphthalene        | 12.58 | 142 | 1950  | 0.40 | ng | 97     |
| 15) Acenaphthylene             | 14.47 | 152 | 14615 | 0.38 | ng | 100    |
| 16) Acenaphthene               | 14.82 | 154 | 2124  | 0.38 | ng | 84     |
| 17) Fluorene                   | 15.81 | 166 | 3027  | 0.42 | ng | 98     |
| 19) Hexachlorobenzene          | 16.82 | 284 | 879   | 0.36 | ng | 93     |
| 20) Pentachlorophenol          | 17.21 | 266 | 202   | 0.53 | ng | # 89   |
| 21) Phenanthrene               | 17.56 | 178 | 5478  | 0.37 | ng | 96     |
| 22) Anthracene                 | 17.65 | 178 | 5480  | 0.40 | ng | 99     |
| 23) Fluoranthene               | 19.54 | 202 | 27376 | 0.43 | ng | 98     |
| 25) Pyrene                     | 19.92 | 202 | 29706 | 0.37 | ng | 100    |
| 27) Benzo(a)anthracene         | 21.66 | 228 | 63170 | 0.76 | ng | 94     |
| 28) Chrysene                   | 21.66 | 228 | 63417 | 0.65 | ng | # 84   |
| 29) Bis(2-ethylhexyl)phthalate | 21.59 | 149 | 2328  | 0.26 | ng | # 70   |
| 30) Indeno(1,2,3-cd)pyrene     | 26.46 | 276 | 31038 | 0.33 | ng | 98     |
| 32) Benzo(b)fluoranthene       | 23.29 | 252 | 7189  | 0.40 | ng | 99     |
| 33) Benzo(k)fluoranthene       | 23.34 | 252 | 8045  | 0.39 | ng | 96     |
| 34) Benzo(a)pyrene             | 23.90 | 252 | 6752  | 0.35 | ng | 99     |
| 35) Dibenzo(a,h)anthracene     | 26.50 | 278 | 6480  | 0.35 | ng | 97     |
| 36) Benzo(g,h,i)perylene       | 27.21 | 276 | 28006 | 0.36 | ng | 98     |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

**Instrument :**  
BNA\_E  
**ClientSampleId :**  
SSTDCCC0.4

Quant Time: Mar 01 08:05:28 2017  
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE021517.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Feb 15 17:05:20 2017  
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.10 min Scan# 573

Delta R.T. -0.02 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08

Instrument :

BNA\_E

ClientSampleId :

SSTDCCC0.4


Tot Ion:152 Resp: 8027

Ion Ratio Lower Upper

152 100

150 161.9 106.0 159.0#




115 64.1 31.2 46.8#



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 0.52 ng

RT: 3.33 min Scan# 64

Delta R.T. -0.01 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08


Tgt Ion: 88 Resp: 568

Ion Ratio Lower Upper

88 100

58 61.3 63.6 95.4#



43 25.4 28.0 42.0#




Abundance Ion 88.00 (87.70 to 88.70): BE0

Ion 58.00 (57.70 to 58.70): BE0

Ion 43.00 (42.70 to 43.70): BE0





#3

n-Nitrosodimethylamine

Concen: 0.50 ng

RT: 3.71 min Scan# 97

Delta R.T. -0.02 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08

Instrument :

BNA\_E

ClientSampleId :

SSTDCCC0.4


Tot Ion: 42 Resp: 453

Ion Ratio Lower Upper

42 100

74 115.9 86.0 129.0




44 0.0 5.1 7.7#



Abundance Ion 42.00 (41.70 to 42.70): BE0

Ion 74.00 (73.70 to 74.70): BE0

Ion 44.00 (43.70 to 44.70): BE0



#4

2-Fluorophenol

Concen: 0.40 ng

RT: 5.62 min Scan# 354

Delta R.T. -0.02 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08


Tgt Ion: 112 Resp: 625

Ion Ratio Lower Upper

112 100

64 68.2 53.5 80.3



63 33.9 27.9 41.9




Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BE0

Ion 63.00 (62.70 to 63.70): BE0





#5

Phenol-d6

Concen: 0.36 ng

RT: 7.31 min Scan# 539

Delta R.T. 0.00 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08

Instrument :

BNA\_E

ClientSampleId :

SSTDCCC0.4


Tot Ion: 99 Resp: 699

Ion Ratio Lower Upper

99 100

42 0.0 16.0 24.0#


71 0.0 30.6 45.8#





Abundance Ion 99.00 (98.70 to 99.70): BE0

Ion 42.00 (41.70 to 42.70): BE0

Ion 71.00 (70.70 to 71.70): BE0



Time-->



#6

bis(2-Chloroethyl)ether

Concen: 0.37 ng

RT: 7.52 min Scan# 548

Delta R.T. -0.02 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08


Tgt Ion: 93 Resp: 788

Ion Ratio Lower Upper

93 100

63 68.5 71.6 107.4#


95 0.0 24.0 36.0#




Abundance Ion 93.00 (92.70 to 93.70): BE0


Ion 63.00 (62.70 to 63.70): BE0

Ion 95.00 (94.70 to 95.70): BE0



Time-->





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.87 min Scan# 704

Delta R.T. -0.02 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08

Instrument :

BNA\_E

ClientSampleId :

SSTDCCC0.4

Tot Ion:136 Resp: 36298


Ion Ratio Lower Upper

136 100

137 10.6 8.2 12.4

54 10.5 9.6 14.4

68 7.5 6.7 10.1






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0



#8

Nitrobenzene-d5

Concen: 0.12 ng

RT: 9.16 min Scan# 619

Delta R.T. -0.16 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08


Tgt Ion: 82 Resp: 105

Ion Ratio Lower Upper

82 100

128 49.5 39.0 58.4



54 53.3 44.8 67.2




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0





#9

Naphthalene

Concen: 0.40 ng

RT: 10.93 min Scan# 707

Delta R.T. -0.00 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08

Instrument :

BNA\_E

ClientSampleId :

SSTDCCC0.4





Tot Ion:128 Resp: 3182

Ion Ratio Lower Upper

128 100

129 15.1 11.2 16.8

127 15.3 11.6 17.4



#10

Hexachlorobutadiene

Concen: 0.41 ng

RT: 11.22 min Scan# 722

Delta R.T. -0.00 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08




Tgt Ion:225 Resp: 472


Ion Ratio Lower Upper

225 100

223 62.9 50.6 76.0

227 62.7 51.4 77.0





#11

2-Methylnaphthalene

Concen: 0.40 ng

RT: 12.58 min Scan# 810

Delta R.T. -0.01 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08

Instrument :

BNA\_E

ClientSampleId :

SSTDCCC0.4


Tot Ion:142 Resp: 1950

Ion Ratio Lower Upper

142 100

141 88.4 73.7 110.5




115 41.5 34.0 51.0



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#12

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.75 min Scan# 976

Delta R.T. -0.02 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08


Tgt Ion:164 Resp: 25677

Ion Ratio Lower Upper

164 100

162 113.7 71.4 107.0#



160 55.5 27.4 41.2#




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E











#13  
2,4,6-Tribromophenol  
Concen: 0.43 ng  
RT: 16.26 min Scan# 1084  
Delta R.T. -0.02 min  
Lab File: BE092762.D  
Acq: 28 Feb 2017 14:08




Instrument :  
BNA\_E  
ClientSampled :  
SSTDCCC0.4


Tot Ion:330 Resp: 234  
Ion Ratio Lower Upper  
330 100  
332 92.7 75.4 113.0  
141 40.6 31.0 46.6



#14  
2-Fluorobiphenyl  
Concen: 0.37 ng  
RT: 13.39 min Scan# 880  
Delta R.T. -0.01 min  
Lab File: BE092762.D  
Acq: 28 Feb 2017 14:08

Tgt Ion:172 Resp: 2290  
Ion Ratio Lower Upper  
172 100  
171 35.3 30.1 45.1  
170 25.0 21.8 32.6





#15

Acenaphthylene

Concen: 0.38 ng

RT: 14.47 min Scan# 960

Delta R.T. 0.00 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08

Instrument :

BNA\_E

ClientSampleId :

SSTDCCC0.4





Tot Ion:152 Resp: 14615

Ion Ratio Lower Upper

152 100

151 4.9 3.9 5.9

153 13.1 10.4 15.6



#16

Acenaphthene

Concen: 0.38 ng

RT: 14.82 min Scan# 980

Delta R.T. -0.02 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08




Tgt Ion:154 Resp: 2124


Ion Ratio Lower Upper

154 100

153 477.2 350.8 526.2

152 243.5 171.1 256.7





#17

Fluorene

Concen: 0.42 ng

RT: 15.81 min Scan# 1045

Delta R.T. -0.01 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08

Instrument :

BNA\_E

ClientSampleId :

SSTDCCC0.4


Tot Ion:166 Resp: 3027

Ion Ratio Lower Upper

166 100

165 100.0 78.2 117.4




167 13.1 11.0 16.4



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#18

Phenanthrene-d10

Concen: 5.00 ng

RT: 17.51 min Scan# 1151

Delta R.T. -0.00 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08


Tgt Ion:188 Resp: 62780

Ion Ratio Lower Upper

188 100

94 5.4 3.2 4.8#



80 4.9 2.6 3.8#




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BE0

Ion 80.00 (79.70 to 80.70): BE0









#19  
Hexachlorobenzene  
Concen: 0.36 ng  
RT: 16.82 min Scan# 1121  
Delta R.T. -0.02 min  
Lab File: BE092762.D  
Acq: 28 Feb 2017 14:08

Instrument :  
BNA\_E  
ClientSampleId :  
SSTDCCC0.4

Tot Ion:284 Resp: 879  
Ion Ratio Lower Upper  
284 100  
142 42.3 29.1 43.7  
249 37.1 27.9 41.9




Abundance Ion 283.80 (283.50 to 284.50): E  
Ion 142.00 (141.70 to 142.70): E  
Ion 249.00 (248.70 to 249.70): E






#20  
Pentachlorophenol  
Concen: 0.53 ng  
RT: 17.21 min Scan# 1138  
Delta R.T. -0.09 min  
Lab File: BE092762.D  
Acq: 28 Feb 2017 14:08

Tgt Ion:266 Resp: 202  
Ion Ratio Lower Upper  
266 100  
268 82.1 62.5 93.7  
264 85.9 57.2 85.8#



Abundance Ion 265.70 (265.40 to 266.40): E  
Ion 268.00 (267.70 to 268.70): E  
Ion 264.00 (263.70 to 264.70): E





#21

Phenanthrene

Concen: 0.37 ng

RT: 17.56 min Scan# 1153

Delta R.T. -0.00 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08

Instrument :

BNA\_E

ClientSampleId :

SSTDCCC0.4


Tot Ion:178 Resp: 5478

Ion Ratio Lower Upper

178 100

176 19.9 15.8 23.8




179 16.4 16.0 24.0



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#22

Anthracene

Concen: 0.40 ng

RT: 17.65 min Scan# 1157

Delta R.T. -0.00 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08


Tgt Ion:178 Resp: 5480

Ion Ratio Lower Upper

178 100

176 18.3 15.0 22.4



179 15.7 12.2 18.2




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#23

Fluoranthene

Concen: 0.43 ng

RT: 19.54 min Scan# 1262

Delta R.T. -0.02 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08

Instrument :

BNA\_E

ClientSampleId :

SSTDCCC0.4


Tot Ion:202 Resp: 27376

Ion Ratio Lower Upper

202 100

101 3.0 2.2 3.4




203 17.8 13.7 20.5



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#24

Chrysene-d12

Concen: 5.00 ng

RT: 21.68 min Scan# 1372

Delta R.T. -0.02 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08


Tgt Ion:240 Resp: 85174

Ion Ratio Lower Upper

240 100

120 4.5 2.6 4.0#



236 30.0 24.6 37.0




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E









#25  
 Pvrene  
 Concen: 0.37 ng  
 RT: 19.92 min Scan# 1284  
 Delta R.T. -0.00 min  
 Lab File: BE092762.D  
 Acq: 28 Feb 2017 14:08




Instrument :  
 BNA\_E  
 ClientSampleId :  
 SSTDCCC0.4


Tot Ion:202 Resp: 29706  
 Ion Ratio Lower Upper  
 202 100  
 200 5.2 4.2 6.4  
 203 17.4 13.9 20.9



#26  
 Terphenyl-d14  
 Concen: 0.37 ng  
 RT: 20.12 min Scan# 1296  
 Delta R.T. -0.02 min  
 Lab File: BE092762.D  
 Acq: 28 Feb 2017 14:08

Tgt Ion:244 Resp: 4027  
 Ion Ratio Lower Upper  
 244 100  
 212 10.6 6.7 10.1#  
 122 9.7 3.6 5.4#









#27  
Benzo(a)anthracene  
Concen: 0.76 ng  
RT: 21.66 min Scan# 1371  
Delta R.T. -0.02 min  
Lab File: BE092762.D  
Acq: 28 Feb 2017 14:08




Instrument :  
BNA\_E  
ClientSampleId :  
SSTDCCC0.4

Tot Ion:228 Resp: 63170  
Ion Ratio Lower Upper  
228 100  
226 32.9 23.6 35.4  
229 14.7 13.3 19.9




#28  
Chrysene  
Concen: 0.65 ng  
RT: 21.66 min Scan# 1371  
Delta R.T. -0.07 min  
Lab File: BE092762.D  
Acq: 28 Feb 2017 14:08

Tgt Ion:228 Resp: 63417  
Ion Ratio Lower Upper  
228 100  
226 32.9 36.3 54.5#  
229 14.7 10.6 16.0











#29  
 Bis(2-ethylhexyl)phthalate  
 Concen: 0.26 ng  
 RT: 21.59 min Scan# 1368  
 Delta R.T. 0.00 min  
 Lab File: BE092762.D  
 Acq: 28 Feb 2017 14:08




Instrument :  
 BNA\_E  
 ClientSampleId :  
 SSTDCCC0.4


Tot Ion:149 Resp: 2328  
 Ion Ratio Lower Upper  
 149 100  
 167 30.3 16.0 24.0#  
 279 3.0 16.0 24.0#



#30  
 Indeno(1,2,3-cd)pyrene  
 Concen: 0.33 ng  
 RT: 26.46 min Scan# 1708  
 Delta R.T. -0.03 min  
 Lab File: BE092762.D  
 Acq: 28 Feb 2017 14:08

Tgt Ion:276 Resp: 31038  
 Ion Ratio Lower Upper  
 276 100  
 138 27.8 20.3 30.5  
 277 24.6 19.7 29.5









#31  
Pervlene-d12  
Concen: 5.00 ng  
RT: 24.01 min Scan# 1537  
Delta R.T. -0.02 min  
Lab File: BE092762.D  
Acq: 28 Feb 2017 14:08

Instrument :  
BNA\_E  
ClientSampleId :  
SSTDCCC0.4

Tot Ion:264 Resp: 74485  




| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 264 | 100   |       |       |
| 260 | 25.5  | 20.1  | 30.1  |
| 265 | 21.3  | 17.9  | 26.9  |




#32  
Benzo(b)fluoranthene  
Concen: 0.40 ng  
RT: 23.29 min Scan# 1475  
Delta R.T. -0.02 min  
Lab File: BE092762.D  
Acq: 28 Feb 2017 14:08

Tgt Ion:252 Resp: 7189  

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 252 | 100   |       |       |
| 253 | 23.0  | 18.7  | 28.1  |
| 125 | 12.5  | 10.5  | 15.7  |





#33

Benzo(k)fluoranthene

Concen: 0.39 ng

RT: 23.34 min Scan# 1479

Delta R.T. -0.02 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08

Instrument :

BNA\_E

ClientSampleId :

SSTDCCC0.4





Tot Ion:252 Resp: 8045

Ion Ratio Lower Upper

252 100

253 23.1 20.3 30.5

125 12.8 9.6 14.4



#34

Benzo(a)pyrene

Concen: 0.35 ng

RT: 23.90 min Scan# 1528

Delta R.T. -0.02 min

Lab File: BE092762.D

Acq: 28 Feb 2017 14:08




Tgt Ion:252 Resp: 6752


Ion Ratio Lower Upper

252 100

253 24.3 19.0 28.6

125 14.0 11.8 17.8









#35  
Dibenzo(a,h)anthracene  
Concen: 0.35 ng  
RT: 26.50 min Scan# 1710  
Delta R.T. -0.02 min  
Lab File: BE092762.D  
Acq: 28 Feb 2017 14:08

Instrument :  
BNA\_E  
ClientSampleId :  
SSTDCCC0.4

Tot Ion:278 Resp: 6480  

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 278 | 100   |       |       |
| 139 | 19.1  | 13.8  | 20.8  |
| 279 | 27.6  | 21.4  | 32.2  |



#36  
Benzo(g,h,i)perylene  
Concen: 0.36 ng  
RT: 27.21 min Scan# 1752  
Delta R.T. -0.03 min  
Lab File: BE092762.D  
Acq: 28 Feb 2017 14:08

Tgt Ion:276 Resp: 28006  

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 276 | 100   |       |       |
| 277 | 25.6  | 19.8  | 29.8  |
| 138 | 25.8  | 19.8  | 29.6  |

