

Data Path : Z:\HPCHEM1\BNA_E\Data\BE022817\
 Data File : BE092762.D
 Acq On : 28 Feb 2017 14:08
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 21 10:22:48 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE021517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 15 17:05:20 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	91	-0.02
2	1,4-Dioxane	0.400	0.525	-31.3#	115	-0.01
3	n-Nitrosodimethylamine	0.400	0.503	-25.7#	99	-0.02
4 S	2-Fluorophenol	0.400	0.400	0.0	95	-0.02
5 S	Phenol-d6	0.400	0.365	8.8	86	0.00
6	bis(2-Chloroethyl)ether	0.400	0.372	7.0	92	-0.02
7 I	Naphthalene-d8	5.000	5.000	0.0	86	-0.02
8 S	Nitrobenzene-d5	0.400	0.425	-6.2	95	-0.02
9	Naphthalene	0.400	0.405	-1.3	87	0.00
10	Hexachlorobutadiene	0.400	0.409	-2.2	85	0.00
11	2-Methylnaphthalene	0.400	0.395	1.3	88	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	100	-0.02
13 S	2,4,6-Tribromophenol	0.400	0.429	-7.2	109	-0.02
14 S	2-Fluorobiphenyl	0.400	0.366	8.5	90	-0.01
15	Acenaphthylene	0.400	0.377	5.8	95	0.00
16	Acenaphthene	0.400	0.377	5.8	93	-0.02
17	Fluorene	0.400	0.419	-4.7	106	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	117	0.00
19	Hexachlorobenzene	0.400	0.363	9.3	97	-0.02
20	Pentachlorophenol	0.400	0.470	-17.5	103	-0.09
21	Phenanthrene	0.400	0.367	8.3	102	0.00
22	Anthracene	0.400	0.399	0.3	119	0.00
23	Fluoranthene	0.400	0.428	-7.0	125	-0.02
24 I	Chrysene-d12	5.000	5.000	0.0	136	-0.02
25	Pyrene	0.400	0.366	8.5	128	0.00
26 S	Terphenyl-d14	0.400	0.370	7.5	130	-0.02
27	Benzo(a)anthracene	0.400	0.372	7.0	129	-0.02
28	Chrysene	0.400	0.328	18.0	111	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.261	34.8#	80	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.334	16.5	95	-0.03
31 I	Perylene-d12	5.000	5.000	0.0	117	-0.02
32	Benzo(b)fluoranthene	0.400	0.397	0.8	103	-0.02
33	Benzo(k)fluoranthene	0.400	0.390	2.5	116	-0.02
34 C	Benzo(a)pyrene	0.400	0.353	11.8	109	-0.02
35	Dibenzo(a,h)anthracene	0.400	0.351	12.3	95	-0.02
36	Benzo(g,h,i)perylene	0.400	0.361	9.8	97	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0