

Data Path : Z:\HPCHEM1\BNA_E\Data\BE022817\
 Data File : BE092762.D
 Acq On : 28 Feb 2017 14:08
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 21 10:22:48 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE021517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 15 17:05:20 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	-0.02
2	1,4-Dioxane	0.710	0.885	-24.6	115	-0.01
3	n-Nitrosodimethylamine	0.740	0.705	4.7	99	-0.02
4 S	2-Fluorophenol	0.973	0.973	0.0	95	-0.02
5 S	Phenol-d6	1.194	1.089	8.8	86	0.00
6	bis(2-Chloroethyl)ether	1.321	1.227	7.1	92	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	86	-0.02
8 S	Nitrobenzene-d5	0.279	0.298	-6.8	95	-0.02
9	Naphthalene	1.082	1.096	-1.3	87	0.00
10	Hexachlorobutadiene	0.159	0.163	-2.5	85	0.00
11	2-Methylnaphthalene	0.680	0.672	1.2	88	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	100	-0.02
13 S	2,4,6-Tribromophenol	0.106	0.114	-7.5	109	-0.02
14 S	2-Fluorobiphenyl	1.219	1.115	8.5	90	-0.01
15	Acenaphthylene	7.544	7.115	5.7	95	0.00
16	Acenaphthene	1.096	1.034	5.7	93	-0.02
17	Fluorene	1.407	1.474	-4.8	106	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.00
19	Hexachlorobenzene	0.193	0.175	9.3	97	-0.02
20	Pentachlorophenol	0.042	0.031	26.2#	103	-0.09
21	Phenanthrene	1.188	1.091	8.2	102	0.00
22	Anthracene	1.094	1.091	0.3	119	0.00
23	Fluoranthene	5.096	5.451	-7.0	125	-0.02
24 I	Chrysene-d12	1.000	1.000	0.0	136	-0.02
25	Pyrene	4.761	4.360	8.4	128	0.00
26 S	Terphenyl-d14	0.638	0.591	7.4	130	-0.02
27	Benzo(a)anthracene	4.849	4.513	6.9	129	-0.02
28	Chrysene	5.752	4.719	18.0	111	0.00
29	Bis(2-ethylhexyl)phthalate	0.674	0.342	49.3#	80	0.00
30	Indeno(1,2,3-cd)pyrene	6.486	4.555	29.8#	95	-0.03
31 I	Perylene-d12	1.000	1.000	0.0	117	-0.02
32	Benzo(b)fluoranthene	1.414	1.206	14.7	103	-0.02
33	Benzo(k)fluoranthene	1.495	1.350	9.7	116	-0.02
34 C	Benzo(a)pyrene	1.285	1.133	11.8	109	-0.02
35	Dibenzo(a,h)anthracene	1.355	1.087	19.8	95	-0.02
36	Benzo(g,h,i)perylene	5.728	4.700	17.9	97	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0