

Data Path : Z:\HPCHEM1\BNA E\DATA\BE022817\
 Data File : BE092772.D
 Acq On : 28 Feb 2017 20:48
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Mar 01 08:21:09 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE021517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 15 17:05:20 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	-0.02
2	1,4-Dioxane	0.710	0.776	-9.3	109	0.00
3	n-Nitrosodimethylamine	0.740	0.693	6.4	105	-0.02
4 S	2-Fluorophenol	0.973	0.904	7.1	95	-0.02
5 S	Phenol-d6	1.194	1.106	7.4	95	0.00
6	bis(2-Chloroethyl)ether	1.321	1.190	9.9	97	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	94	-0.02
8 S	Nitrobenzene-d5	0.279	0.290	-3.9	101	-0.02
9	Naphthalene	1.082	1.086	-0.4	94	0.00
10	Hexachlorobutadiene	0.159	0.158	0.6	90	0.00
11	2-Methylnaphthalene	0.680	0.678	0.3	96	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	108	-0.02
13 S	2,4,6-Tribromophenol	0.106	0.120	-13.2	124	-0.02
14 S	2-Fluorobiphenyl	1.219	1.129	7.4	98	-0.01
15	Acenaphthylene	7.544	7.304	3.2	105	-0.02
16	Acenaphthene	1.096	1.060	3.3	103	-0.02
17	Fluorene	1.407	1.449	-3.0	112	-0.01
18 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
19	Hexachlorobenzene	0.193	0.182	5.7	103	-0.02
20	Pentachlorophenol	0.042	0.037	11.9	124	-0.09
21	Phenanthrene	1.188	1.090	8.2	105	0.00
22	Anthracene	1.094	1.128	-3.1	127	0.00
23	Fluoranthene	5.096	5.481	-7.6	129	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	138	-0.02
25	Pyrene	4.761	4.261	10.5	127	0.00
26 S	Terphenyl-d14	0.638	0.590	7.5	132	-0.02
27	Benzo(a)anthracene	4.849	4.549	6.2	132	-0.02
28	Chrysene	5.752	4.747	17.5	114	0.00
29	Bis(2-ethylhexyl)phthalate	0.674	0.380	43.6#	90	0.00
30	Indeno(1,2,3-cd)pyrene	6.486	4.536	30.1#	96	-0.03
31 I	Perylene-d12	1.000	1.000	0.0	117	-0.02
32	Benzo(b)fluoranthene	1.414	1.213	14.2	104	-0.02
33	Benzo(k)fluoranthene	1.495	1.365	8.7	118	-0.02
34 C	Benzo(a)pyrene	1.285	1.155	10.1	112	-0.02
35	Dibenzo(a,h)anthracene	1.355	1.077	20.5	95	-0.03
36	Benzo(g,h,i)perylene	5.728	4.834	15.6	100	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0