

Data Path : Z:\HPCHEM1\BNA E\DATA\BE030117\
 Data File : BE092776.D
 Acq On : 1 Mar 2017 10:08
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 01 13:11:34 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE021517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 15 17:05:20 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	-0.02
2	1,4-Dioxane	0.710	0.774	-9.0	113	-0.01
3	n-Nitrosodimethylamine	0.740	0.630	14.9	99	-0.02
4 S	2-Fluorophenol	0.973	0.941	3.3	103	-0.02
5 S	Phenol-d6	1.194	1.077	9.8	95	0.00
6	bis(2-Chloroethyl)ether	1.321	1.213	8.2	102	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	99	-0.02
8 S	Nitrobenzene-d5	0.279	0.267	4.3	98	-0.02
9	Naphthalene	1.082	1.087	-0.5	98	0.00
10	Hexachlorobutadiene	0.159	0.155	2.5	92	0.00
11	2-Methylnaphthalene	0.680	0.669	1.6	100	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	110	-0.02
13 S	2,4,6-Tribromophenol	0.106	0.111	-4.7	117	-0.02
14 S	2-Fluorobiphenyl	1.219	1.139	6.6	102	-0.01
15	Acenaphthylene	7.544	7.344	2.7	108	0.00
16	Acenaphthene	1.096	1.064	2.9	106	-0.02
17	Fluorene	1.407	1.467	-4.3	116	-0.01
18 I	Phenanthrene-d10	1.000	1.000	0.0	127	0.00
19	Hexachlorobenzene	0.193	0.177	8.3	106	-0.02
20	Pentachlorophenol	0.042	0.038	9.5	137	-0.09
21	Phenanthrene	1.188	1.085	8.7	110	0.00
22	Anthracene	1.094	1.087	0.6	129	0.00
23	Fluoranthene	5.096	5.444	-6.8	136	-0.02
24 I	Chrysene-d12	1.000	1.000	0.0	137	-0.02
25	Pyrene	4.761	4.669	1.9	138	0.00
26 S	Terphenyl-d14	0.638	0.614	3.8	136	-0.02
27	Benzo(a)anthracene	4.849	4.763	1.8	137	-0.02
28	Chrysene	5.752	5.103	11.3	121	0.00
29	Bis(2-ethylhexyl)phthalate	0.674	0.481	28.6#	113	0.00
30	Indeno(1,2,3-cd)pyrene	6.486	4.937	23.9	103	-0.02
31 I	Perylene-d12	1.000	1.000	0.0	126	-0.01
32	Benzo(b)fluoranthene	1.414	1.344	5.0	124	-0.01
33	Benzo(k)fluoranthene	1.495	1.237	17.3	115	-0.01
34 C	Benzo(a)pyrene	1.285	1.168	9.1	122	-0.03
35	Dibenzo(a,h)anthracene	1.355	1.095	19.2	104	-0.02
36	Benzo(g,h,i)perylene	5.728	4.715	17.7	105	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0