

Data Path : Z:\HPCHEM1\BNA E\DATA\BE030117\
 Data File : BE092776.D
 Acq On : 1 Mar 2017 10:08
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 01 13:11:34 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE021517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 15 17:05:20 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	102	-0.02
2	1,4-Dioxane	0.400	0.453	-13.2	113	-0.01
3	n-Nitrosodimethylamine	0.400	0.471	-17.7	99	-0.02
4 S	2-Fluorophenol	0.400	0.387	3.3	103	-0.02
5 S	Phenol-d6	0.400	0.361	9.8	95	0.00
6	bis(2-Chloroethyl)ether	0.400	0.368	8.0	102	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	99	-0.02
8 S	Nitrobenzene-d5	0.400	0.389	2.8	98	-0.02
9	Naphthalene	0.400	0.402	-0.5	98	0.00
10	Hexachlorobutadiene	0.400	0.389	2.8	92	0.00
11	2-Methylnaphthalene	0.400	0.394	1.5	100	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	110	-0.02
13 S	2,4,6-Tribromophenol	0.400	0.417	-4.2	117	-0.02
14 S	2-Fluorobiphenyl	0.400	0.374	6.5	102	-0.01
15	Acenaphthylene	0.400	0.389	2.8	108	0.00
16	Acenaphthene	0.400	0.388	3.0	106	-0.02
17	Fluorene	0.400	0.417	-4.2	116	-0.01
18 I	Phenanthrene-d10	5.000	5.000	0.0	127	0.00
19	Hexachlorobenzene	0.400	0.366	8.5	106	-0.02
20	Pentachlorophenol	0.400	0.517	-29.3#	137	-0.09
21	Phenanthrene	0.400	0.365	8.8	110	0.00
22	Anthracene	0.400	0.398	0.5	129	0.00
23	Fluoranthene	0.400	0.427	-6.7	136	-0.02
24 I	Chrysene-d12	5.000	5.000	0.0	137	-0.02
25	Pyrene	0.400	0.392	2.0	138	0.00
26 S	Terphenyl-d14	0.400	0.385	3.8	136	-0.02
27	Benzo(a)anthracene	0.400	0.393	1.8	137	-0.02
28	Chrysene	0.400	0.355	11.3	121	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.343	14.2	113	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.359	10.3	103	-0.02
31 I	Perylene-d12	5.000	5.000	0.0	126	-0.01
32	Benzo(b)fluoranthene	0.400	0.439	-9.7	124	-0.01
33	Benzo(k)fluoranthene	0.400	0.357	10.8	115	-0.01
34 C	Benzo(a)pyrene	0.400	0.364	9.0	122	-0.03
35	Dibenzo(a,h)anthracene	0.400	0.354	11.5	104	-0.02
36	Benzo(g,h,i)perylene	0.400	0.363	9.3	105	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0