

Data Path : Z:\HPCHEM1\BNA E\DATA\BE030117\
 Data File : BE092777.D
 Acq On : 1 Mar 2017 10:47
 Operator : SJ/MA
 Sample : PB97227BS
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB97227BS

Manual Integrations
 APPROVED

mohammad
 3/1/2017 3:08:02 PM

Quant Time: Mar 01 12:58:30 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE021517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 15 17:05:20 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.09	152	13165	5.00	ng	-0.02
7) Naphthalene-d8	10.88	136	54446	5.00	ng	-0.02
12) Acenaphthene-d10	14.75	164	28191	5.00	ng	-0.02
18) Phenanthrene-d10	17.51	188	59420	5.00	ng	0.00
24) Chrysene-d12	21.68	240	87525	5.00	ng	-0.02
31) Perylene-d12	24.01	264	89235	5.00	ng	-0.02

System Monitoring Compounds						
4) 2-Fluorophenol	5.60	112	21634	8.45	ng	-0.03
5) Phenol-d6	7.27	99	29256	9.31	ng	-0.05
8) Nitrobenzene-d5	9.27	82	12861	3.52	ng	-0.05
13) 2,4,6-Tribromophenol	16.25	330	6682	11.17	ng	-0.03
14) 2-Fluorobiphenyl	13.36	172	27833	4.05	ng	-0.03
26) Terphenyl-d14	20.12	244	37008	3.31	ng	-0.02

Target Compounds						Qvalue
2) 1,4-Dioxane	3.33	88	4842	2.94	ng	# 72
3) n-Nitrosodimethylamine	3.68	42	7784	3.28	ng	# 99
6) bis(2-Chloroethyl)ether	7.52	93	12681	3.65	ng	91
9) Naphthalene	10.91	128	38848	3.30	ng	# 95
10) Hexachlorobutadiene	11.20	225	5898	3.40	ng	99
11) 2-Methylnaphthalene	12.56	142	25325	3.42	ng	# 87
15) Acenaphthylene	14.46	152	163968	3.85	ng	100
16) Acenaphthene	14.82	154	23838	3.86	ng	90
17) Fluorene	15.80	166	31836	4.01	ng	99
19) Hexachlorobenzene	16.82	284	9280	4.04	ng	# 66
20) Pentachlorophenol	17.16	266	6986	10.28	ng	# 82
21) Phenanthrene	17.53	178	52582	3.72	ng	# 94
22) Anthracene	17.63	178	52866	4.07	ng	98
23) Fluoranthene	19.54	202	250271	4.13	ng	99
25) Pyrene	19.90	202	267675	3.21	ng	100
27) Benzo(a)anthracene	21.66	228	315075m	3.71	ng	
28) Chrysene	21.70	228	261427m	2.60	ng	
29) Bis(2-ethylhexyl)phthalate	21.59	149	27523	2.38	ng	# 67
30) Indeno(1,2,3-cd)pyrene	26.45	276	380874	3.65	ng	97
32) Benzo(b)fluoranthene	23.29	252	80987m	3.46	ng	
33) Benzo(k)fluoranthene	23.34	252	79431	3.19	ng	94
34) Benzo(a)pyrene	23.90	252	80366	3.50	ng	# 95
35) Dibenzo(a,h)anthracene	26.48	278	78758	3.49	ng	95
36) Benzo(g,h,i)perylene	27.21	276	328702	3.52	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE030117\
Data File : BE092777.D
Acq On : 1 Mar 2017 10:47
Operator : SJ/MA
Sample : PB97227BS
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
PB97227BS

Manual Integrations
APPROVED

mohammad
3/1/2017 3:08:02 PM

Quant Time: Mar 01 12:58:30 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE021517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 15 17:05:20 2017
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.09 min Scan# 573

Delta R.T. -0.02 min

Lab File: BE092777.D

Acq: 1 Mar 2017 10:47

Instrument :

BNA_E





ClientSampleId :

PB97227BS

Tot Ion	Ratio	Lower	Upper
152	100		
150	156.8	106.0	159.0
115	60.0	31.2	46.8

Manual Integrations
APPROVED

mohammad
3/1/2017 3:08:02 PM



#2

1,4-Dioxane

Concen: 2.94 ng




RT: 3.33 min Scan# 64


Delta R.T. -0.01 min

Lab File: BE092777.D

Acq: 1 Mar 2017 10:47

Tgt Ion	Ratio	Lower	Upper
88	100		
58	94.5	63.6	95.4
43	64.8	28.0	42.0





#3





n-Nitrosodimethylamine
Concen: 3.28 ng
RT: 3.68 min Scan# 94
Delta R.T. -0.06 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Instrument :
BNA_E
ClientSampleId :
PB97227BS

Tot Ion	Ratio	Lower	Upper
42	100		
74	107.6	86.0	129.0
44	9.6	5.1	7.7#

Manual Integrations
APPROVED




mohammad
3/1/2017 3:08:02 PM






#4

2-Fluorophenol
Concen: 8.45 ng
RT: 5.60 min Scan# 351
Delta R.T. -0.03 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Tgt Ion	Ratio	Lower	Upper
112	100		
64	68.1	53.5	80.3
63	36.2	27.9	41.9









#5
Phenol-d6
Concen: 9.31 ng
RT: 7.27 min Scan# 537
Delta R.T. -0.05 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Tot Ion	Ratio	Lower	Upper
99	100		
42	0.0	16.0	24.0#
71	31.5	30.6	45.8

Instrument :
BNA_E
ClientSampleId :
PB97227BS


Manual Integrations
APPROVED


mohammad
3/1/2017 3:08:02 PM



#6
bis(2-Chloroethyl)ether
Concen: 3.65 ng
RT: 7.52 min Scan# 548
Delta R.T. -0.02 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Tgt Ion	Ratio	Lower	Upper
93	100		
63	79.7	71.6	107.4
95	32.7	24.0	36.0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.88 min Scan# 704

Delta R.T. -0.02 min

Lab File: BE092777.D

Acq: 1 Mar 2017 10:47

Instrument :

BNA_E





ClientSampleId :

PB97227BS

Tot Ion:	136	Resp:	54446
Ion Ratio	Lower	Upper	
136	100		
137	11.2	8.2	12.4
54	6.7	9.6	14.4#
68	5.1	6.7	10.1#

Manual Integrations
APPROVED

mohammad
3/1/2017 3:08:02 PM



#8

Nitrobenzene-d5

Concen: 3.52 ng




RT: 9.27 min Scan# 624


Delta R.T. -0.05 min

Lab File: BE092777.D

Acq: 1 Mar 2017 10:47

Tgt Ion:	82	Resp:	12861
Ion Ratio	Lower	Upper	
82	100		
128	49.0	39.0	58.4
54	40.5	44.8	67.2#





#9

Naphthalene

Concen: 3.30 ng

RT: 10.91 min Scan# 706

Delta R.T. -0.02 min

Lab File: BE092777.D

Acq: 1 Mar 2017 10:47

Instrument :

BNA_E





ClientSampleId :

PB97227BS

Tot Ion	Ratio	Lower	Upper
128	100		
129	10.4	11.2	16.8#
127	14.3	11.6	17.4

Manual Integrations
APPROVED

mohammad
3/1/2017 3:08:02 PM



#10

Hexachlorobutadiene

Concen: 3.40 ng




RT: 11.20 min Scan# 721


Delta R.T. -0.02 min

Lab File: BE092777.D

Acq: 1 Mar 2017 10:47

Tgt Ion	Ratio	Lower	Upper
225	100		
223	63.4	50.6	76.0
227	62.8	51.4	77.0









#11
2-Methylnaphthalene
Concen: 3.42 ng
RT: 12.56 min Scan# 808
Delta R.T. -0.03 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Instrument :
BNA_E
ClientSampleId :
PB97227BS

Tot Ion	Ratio	Lower	Upper
142	100		
141	81.8	73.7	110.5
115	30.9	34.0	51.0#




Manual Integrations
APPROVED


mohammad
3/1/2017 3:08:02 PM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.75 min Scan# 976
Delta R.T. -0.02 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Tgt Ion	Ratio	Lower	Upper
164	100		
162	105.8	71.4	107.0
160	48.3	27.4	41.2#









#13
 2,4,6-Tribromophenol
 Concen: 11.17 ng
 RT: 16.25 min Scan# 1083
 Delta R.T. -0.03 min
 Lab File: BE092777.D
 Acq: 1 Mar 2017 10:47

Instrument :
 BNA_E
 ClientSampleId :
 PB97227BS

Tot Ion:330 Resp: 6682
 Ion Ratio Lower Upper
 330 100
 332 97.2 75.4 113.0
 141 43.3 31.0 46.6




Manual Integrations
 APPROVED


mohammad
 3/1/2017 3:08:02 PM



#14
 2-Fluorobiphenyl
 Concen: 4.05 ng
 RT: 13.36 min Scan# 878
 Delta R.T. -0.03 min
 Lab File: BE092777.D
 Acq: 1 Mar 2017 10:47

Tgt Ion:172 Resp: 27833
 Ion Ratio Lower Upper
 172 100
 171 35.7 30.1 45.1
 170 24.3 21.8 32.6





#15

Acenaphthylene

Concen: 3.85 ng

RT: 14.46 min Scan# 959

Delta R.T. -0.02 min

Lab File: BE092777.D

Acq: 1 Mar 2017 10:47

Instrument :

BNA_E





ClientSampleId :

PB97227BS

Tot Ion:	152	Resp:	163968
Ion Ratio	Lower	Upper	
152	100		
151	4.9	3.9	5.9
153	13.0	10.4	15.6

Manual Integrations
APPROVED

mohammad
3/1/2017 3:08:02 PM



#16

Acenaphthene

Concen: 3.86 ng




RT: 14.82 min Scan# 980


Delta R.T. -0.02 min

Lab File: BE092777.D

Acq: 1 Mar 2017 10:47

Tgt Ion:	154	Resp:	23838
Ion Ratio	Lower	Upper	
154	100		
153	460.2	350.8	526.2
152	234.0	171.1	256.7





#17

Fluorene

Concen: 4.01 ng

RT: 15.80 min Scan# 1044

Delta R.T. -0.02 min

Lab File: BE092777.D

Acq: 1 Mar 2017 10:47

Instrument :

BNA_E





ClientSampleId :

PB97227BS

Tot Ion	Ratio	Lower	Upper
166	100		
165	98.6	78.2	117.4
167	13.5	11.0	16.4

Manual Integrations
APPROVED

mohammad
3/1/2017 3:08:02 PM



#18

Phenanthrene-d10

Concen: 5.00 ng




RT: 17.51 min Scan# 1151


Delta R.T. -0.00 min

Lab File: BE092777.D

Acq: 1 Mar 2017 10:47

Tgt Ion	Ratio	Lower	Upper
188	100		
94	3.2	3.2	4.8
80	2.6	2.6	3.8









#19
Hexachlorobenzene
Concen: 4.04 ng
RT: 16.82 min Scan# 1121
Delta R.T. -0.02 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Instrument :
BNA_E
ClientSampleId :
PB97227BS

Tot Ion	Ratio	Lower	Upper
284	100		
142	0.0	29.1	43.7#
249	32.4	27.9	41.9




Manual Integrations
APPROVED


mohammad
3/1/2017 3:08:02 PM



#20
Pentachlorophenol
Concen: 10.28 ng
RT: 17.16 min Scan# 1136
Delta R.T. -0.14 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Tgt Ion	Ratio	Lower	Upper
266	100		
268	51.8	62.5	93.7#
264	75.8	57.2	85.8









#21
Phenanthrene
Concen: 3.72 ng
RT: 17.53 min Scan# 1152
Delta R.T. -0.02 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Instrument :
BNA_E
ClientSampleId :
PB97227BS

Tot Ion	Ratio	Lower	Upper
178	100		
176	20.5	15.8	23.8
179	14.8	16.0	24.0




Manual Integrations
APPROVED


mohammad
3/1/2017 3:08:02 PM



#22
Anthracene
Concen: 4.07 ng
RT: 17.63 min Scan# 1156
Delta R.T. -0.02 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.7	15.0	22.4
179	14.8	12.2	18.2





#23

Fluoranthene

Concen: 4.13 ng

RT: 19.54 min Scan# 1262

Delta R.T. -0.02 min

Lab File: BE092777.D

Acq: 1 Mar 2017 10:47

Instrument :

BNA_E





ClientSampleId :

PB97227BS

Tot Ion:	202	Resp:	250271
Ion Ratio	Lower	Upper	
202	100		
101	2.9	2.2	3.4
203	17.7	13.7	20.5

Manual Integrations
APPROVED

mohammad
3/1/2017 3:08:02 PM



#24

Chrysene-d12

Concen: 5.00 ng




RT: 21.68 min Scan# 1372


Delta R.T. -0.02 min

Lab File: BE092777.D

Acq: 1 Mar 2017 10:47

Tgt Ion:	240	Resp:	87525
Ion Ratio	Lower	Upper	
240	100		
120	4.2	2.6	4.0#
236	29.8	24.6	37.0









#25
Pvrene
Concen: 3.21 ng
RT: 19.90 min Scan# 1283
Delta R.T. -0.02 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Instrument :
BNA_E
ClientSampleId :
PB97227BS

Tot Ion	Ratio	Resp Lower	Upper
202	100		
200	5.4	4.2	6.4
203	17.3	13.9	20.9




Manual Integrations
APPROVED


mohammad
3/1/2017 3:08:02 PM



#26
Terphenyl-d14
Concen: 3.31 ng
RT: 20.12 min Scan# 1296
Delta R.T. -0.02 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Tgt Ion	Ratio	Resp Lower	Upper
244	100		
212	7.0	6.7	10.1
122	2.0	3.6	5.4









#27
Benzo(a)anthracene
Concen: 3.71 ng m
RT: 21.66 min Scan# 1371
Delta R.T. -0.02 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Instrument :
BNA_E
ClientSampleId :
PB97227BS

Tot Ion	Ratio	Lower	Upper
228	100		
226	30.0	23.6	35.4
229	14.5	13.3	19.9




Manual Integrations
APPROVED


mohammad
3/1/2017 3:08:02 PM



#28
Chrysene
Concen: 2.60 ng m
RT: 21.70 min Scan# 1373
Delta R.T. -0.02 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Tgt Ion	Ratio	Lower	Upper
228	100		
226	50.6	36.3	54.5
229	10.5	10.6	16.0









#29
 Bis(2-ethylhexyl)phthalate
 Concen: 2.38 ng
 RT: 21.59 min Scan# 1368
 Delta R.T. 0.00 min
 Lab File: BE092777.D
 Acq: 1 Mar 2017 10:47

Instrument :
 BNA_E
 ClientSampleId :
 PB97227BS

Tot Ion	Ratio	Lower	Upper
149	100		
167	33.1	16.0	24.0#
279	2.4	16.0	24.0#




Manual Integrations
 APPROVED


mohammad
 3/1/2017 3:08:02 PM



#30
 Indeno(1,2,3-cd)pyrene
 Concen: 3.65 ng
 RT: 26.45 min Scan# 1707
 Delta R.T. -0.05 min
 Lab File: BE092777.D
 Acq: 1 Mar 2017 10:47

Tgt Ion	Ratio	Lower	Upper
276	100		
138	28.1	20.3	30.5
277	24.8	19.7	29.5









#31
Pervlene-d12
Concen: 5.00 ng
RT: 24.01 min Scan# 1537
Delta R.T. -0.02 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Instrument :
BNA_E
ClientSampleId :
PB97227BS

Tot Ion	Ratio	Lower	Upper
264	100		
260	26.1	20.1	30.1
265	20.5	17.9	26.9




Manual Integrations
APPROVED


mohammad
3/1/2017 3:08:02 PM



#32
Benzo(b)fluoranthene
Concen: 3.46 ng m
RT: 23.29 min Scan# 1475
Delta R.T. -0.02 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.3	18.7	28.1
125	10.2	10.5	15.7









#33
Benzo(k)fluoranthene
Concen: 3.19 ng
RT: 23.34 min Scan# 1479
Delta R.T. -0.02 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Instrument :
BNA_E
ClientSampleId :
PB97227BS

Tot Ion	Ratio	Lower	Upper
252	100		
253	21.9	20.3	30.5
125	10.5	9.6	14.4




Manual Integrations
APPROVED


mohammad
3/1/2017 3:08:02 PM



#34
Benzo(a)pyrene
Concen: 3.50 ng
RT: 23.90 min Scan# 1528
Delta R.T. -0.02 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.1	19.0	28.6
125	11.4	11.8	17.8









#35
Dibenzo(a,h)anthracene
Concen: 3.49 ng
RT: 26.48 min Scan# 1709
Delta R.T. -0.03 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Instrument :
BNA_E
ClientSampleId :
PB97227BS

Tot Ion	Ratio	Lower	Upper
278	100		
139	14.4	13.8	20.8
279	24.5	21.4	32.2

Manual Integrations
APPROVED

mohammad
3/1/2017 3:08:02 PM



#36
Benzo(g,h,i)perylene
Concen: 3.52 ng
RT: 27.21 min Scan# 1752
Delta R.T. -0.03 min
Lab File: BE092777.D
Acq: 1 Mar 2017 10:47

Tgt Ion	Ratio	Lower	Upper
276	100		
277	24.5	19.8	29.8
138	22.4	19.8	29.6

