

Data Path : Z:\HPCHEM1\BNA\_E\DATA\BE030316\  
 Data File : BE091433.D  
 Acq On : 3 Mar 2016 21:34  
 Operator : UM/IZ  
 Sample : MDL-02-S  
 Misc : 0.08PPM  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 MDL-02-S

Quant Time: Mar 04 07:51:07 2016  
 Quant Method : Z:\HPCHEM1\BNA\_E\METHODS\SIM-BE030316.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Mar 04 06:35:39 2016  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|---------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 6.85  | 152  | 34320    | 5.00 | ng    | -0.02    |
| 4) Naphthalene-d8         | 9.53  | 136  | 147565   | 5.00 | ng    | 0.00     |
| 8) Acenaphthene-d10       | 13.34 | 164  | 93265    | 5.00 | ng    | 0.00     |
| 13) Phenanthrene-d10      | 16.10 | 188  | 157123   | 5.00 | ng    | 0.00     |
| 19) Chrysene-d12          | 20.21 | 240  | 249898   | 5.00 | ng    | 0.00     |
| 25) Perylene-d12          | 22.09 | 264  | 281243   | 5.00 | ng    | 0.00     |

|                             |       |     |        |      |    |       |
|-----------------------------|-------|-----|--------|------|----|-------|
| System Monitoring Compounds |       |     |        |      |    |       |
| 5) Nitrobenzene-d5          | 8.15  | 82  | 37174  | 4.07 | ng | -0.03 |
| 9) 2-Fluorobiphenyl         | 12.00 | 172 | 107225 | 4.74 | ng | 0.00  |
| 21) Terphenyl-d14           | 18.59 | 244 | 184410 | 4.85 | ng | -0.02 |


|                            |       |     |       |           |    |       |
|----------------------------|-------|-----|-------|-----------|----|-------|
| Target Compounds           |       |     |       | Qvalue    |    |       |
| 2) 1,4-Dioxane             | 3.00  | 88  | 367   | 0.12      | ng | 98    |
| 3) n-Nitrosodimethylamine  | 3.63  | 42  | 129   | 0.09      | ng | # 1   |
| 6) Naphthalene             | 9.57  | 128 | 4760  | 0.14      | ng | 94    |
| 7) 2-Methylnaphthalene     | 11.15 | 142 | 2812  | 0.14      | ng | # 86  |
| 10) Acenaphthylene         | 13.09 | 152 | 18643 | 0.12      | ng | 99    |
| 11) Acenaphthene           | 13.41 | 154 | 3912  | 0.14      | ng | # 100 |
| 12) Fluorene               | 14.41 | 166 | 4414  | 0.12      | ng | 97    |
| 14) Hexachlorobenzene      | 15.38 | 284 | 1456  | 0.16      | ng | # 100 |
| 15) Pentachlorophenol      | 15.93 | 266 | 103   | Below Cal |    | 94    |
| 16) Phenanthrene           | 16.13 | 178 | 7861  | 0.14      | ng | 99    |
| 17) Anthracene             | 16.24 | 178 | 4301  | 0.22      | ng | # 82  |
| 18) Fluoranthene           | 18.09 | 202 | 8207  | 0.21      | ng | 98    |
| 20) Pyrene                 | 18.48 | 202 | 8614  | 0.12      | ng | 99    |
| 22) Benzo(a)anthracene     | 20.17 | 228 | 9862  | 0.11      | ng | 99    |
| 23) Chrysene               | 20.24 | 228 | 12025 | 0.13      | ng | 98    |
| 24) Indeno(1,2,3-cd)pyrene | 23.70 | 276 | 9782  | 0.13      | ng | # 89  |
| 26) Benzo(b)fluoranthene   | 21.57 | 252 | 18828 | 0.24      | ng | 95    |
| 27) Benzo(k)fluoranthene   | 21.57 | 252 | 19007 | 0.25      | ng | 99    |
| 28) Benzo(a)pyrene         | 21.99 | 252 | 8057  | 0.12      | ng | 95    |
| 29) Dibenzo(a,h)anthracene | 23.65 | 278 | 7546  | 0.12      | ng | 96    |
| 30) Benzo(g,h,i)perylene   | 24.24 | 276 | 8744  | 0.12      | ng | 95    |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_E\DATA\BE030316\  
Data File : BE091433.D  
Acq On : 3 Mar 2016 21:34  
Operator : UM/IZ  
Sample : MDL-02-S  
Misc : 0.08PPM  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
BNA\_E  
ClientSampleId :  
MDL-02-S

Quant Time: Mar 04 07:51:07 2016  
Quant Method : Z:\HPCHEM1\BNA\_E\METHODS\SIM-BE030316.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Mar 04 06:35:39 2016  
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 6.85 min Scan# 245

Delta R.T. -0.02 min

Lab File: BE091433.D

Acq: 3 Mar 2016 21:34

Instrument :

BNA\_E

ClientSampleId :

MDL-02-S





Tgt Ion:152 Resp: 34320

Ion Ratio Lower Upper

152 100

150 168.0 115.0 172.4

115 70.8 42.2 63.4#



#2

1,4-Dioxane

Concen: 0.12 ng

RT: 3.00 min Scan# 24

Delta R.T. 0.03 min

Lab File: BE091433.D

Acq: 3 Mar 2016 21:34




Tgt Ion: 88 Resp: 367


Ion Ratio Lower Upper

88 100

58 124.8 97.9 146.9

43 38.1 29.0 43.4





#3

n-Nitrosodimethylamine

Concen: 0.09 ng

RT: 3.63 min Scan# 60

Delta R.T. 0.09 min

Lab File: BE091433.D

Acq: 3 Mar 2016 21:34

Instrument :

BNA\_E

ClientSampleId :

MDL-02-S


Tgt Ion: 42 Resp: 129

Ion Ratio Lower Upper

42 100

74 0.0 93.8 140.6#




44 31.0 0.0 0.0#



Abundance Ion 42.00 (41.70 to 42.70): BE0

Ion 74.00 (73.70 to 74.70): BE0

Ion 44.00 (43.70 to 44.70): BE0



#4

Naphthalene-d8

Concen: 5.00 ng

RT: 9.53 min Scan# 450

Delta R.T. 0.00 min

Lab File: BE091433.D

Acq: 3 Mar 2016 21:34


Tgt Ion: 136 Resp: 147565

Ion Ratio Lower Upper

136 100

137 10.8 8.7 13.1



54 10.4 6.9 10.3#




Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0





#5

Nitrobenzene-d5

Concen: 4.07 ng

RT: 8.15 min Scan# 336

Delta R.T. -0.03 min

Lab File: BE091433.D

Acq: 3 Mar 2016 21:34

Instrument :

BNA\_E

ClientSampleId :

MDL-02-S


Tgt Ion: 82 Resp: 37174

Ion Ratio Lower Upper

82 100

128 28.5 30.0 45.0#




54 60.3 41.5 62.3



Abundance Ion 82.00 (81.70 to 82.70): BE091425.D (-333) (-)

Ion 128.00 (127.70 to 128.70): BE091425.D (-333) (-)

Ion 54.00 (53.70 to 54.70): BE091425.D (-333) (-)



#6

Naphthalene

Concen: 0.14 ng

RT: 9.57 min Scan# 452

Delta R.T. 0.00 min

Lab File: BE091433.D

Acq: 3 Mar 2016 21:34


Tgt Ion: 128 Resp: 4760

Ion Ratio Lower Upper

128 100

129 13.8 9.5 14.3



127 16.7 11.4 17.0




Abundance Ion 128.00 (127.70 to 128.70): BE091425.D (-444) (-)

Ion 129.00 (128.70 to 129.70): BE091425.D (-444) (-)

Ion 127.00 (126.70 to 127.70): BE091425.D (-444) (-)





#7

2-Methylnaphthalene

Concen: 0.14 ng

RT: 11.15 min Scan# 554

Delta R.T. 0.01 min

Lab File: BE091433.D

Acq: 3 Mar 2016 21:34

Instrument :

BNA\_E

ClientSampleId :

MDL-02-S


Tgt Ion:142 Resp: 2812

Ion Ratio Lower Upper

142 100

141 97.1 68.3 102.5




115 45.7 28.9 43.3#



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#8

Acenaphthene-d10

Concen: 5.00 ng

RT: 13.34 min Scan# 720

Delta R.T. 0.00 min

Lab File: BE091433.D

Acq: 3 Mar 2016 21:34


Tgt Ion:164 Resp: 93265

Ion Ratio Lower Upper

164 100

162 101.4 79.1 118.7



160 43.2 32.4 48.6




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E





#9

2-Fluorobiphenyl

Concen: 4.74 ng

RT: 12.00 min Scan# 628

Delta R.T. -0.00 min

Lab File: BE091433.D

Acq: 3 Mar 2016 21:34

Instrument :

BNA\_E

ClientSampleId :

MDL-02-S


Tgt Ion:172 Resp: 107225

Ion Ratio Lower Upper

172 100

171 31.7 30.2 45.4




170 19.4 20.9 31.3#



Abundance Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E



#10

Acenaphthylene

Concen: 0.12 ng

RT: 13.09 min Scan# 709

Delta R.T. 0.00 min

Lab File: BE091433.D

Acq: 3 Mar 2016 21:34


Tgt Ion:152 Resp: 18643

Ion Ratio Lower Upper

152 100

151 4.6 3.9 5.9



153 12.7 10.5 15.7




Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E





#11

Acenaphthene

Concen: 0.14 ng

RT: 13.41 min Scan# 723

Delta R.T. 0.00 min

Lab File: BE091433.D

Acq: 3 Mar 2016 21:34

Instrument :

BNA\_E

ClientSampleId :

MDL-02-S


Tgt Ion:154 Resp: 3912

Ion Ratio Lower Upper




154 100

153 408.9 0.0 0.0#

152 0.0 0.0 0.0



Abundance Ion 154.00 (153.70 to 154.70): E  
Ion 153.00 (152.70 to 153.70): E  
Ion 152.00 (151.70 to 152.70): E



#12

Fluorene

Concen: 0.12 ng

RT: 14.41 min Scan# 767

Delta R.T. 0.00 min

Lab File: BE091433.D

Acq: 3 Mar 2016 21:34


Tgt Ion:166 Resp: 4414

Ion Ratio Lower Upper



166 100


165 96.8 79.4 119.2

167 14.6 10.7 16.1



Abundance Ion 166.00 (165.70 to 166.70): E  
Ion 165.00 (164.70 to 165.70): E  
Ion 167.00 (166.70 to 167.70): E





#13

Phenanthrene-d10

Concen: 5.00 ng

RT: 16.10 min Scan# 818

Delta R.T. 0.00 min

Lab File: BE091433.D

Acq: 3 Mar 2016 21:34

Instrument :

BNA\_E

ClientSampleId :

MDL-02-S


Tgt Ion:188 Resp: 157123

Ion Ratio Lower Upper

188 100

94 3.4 2.1 3.1#




80 3.2 1.9 2.9#



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BE0

Ion 80.00 (79.70 to 80.70): BE0



#14

Hexachlorobenzene

Concen: 0.16 ng

RT: 15.38 min Scan# 797

Delta R.T. 0.00 min

Lab File: BE091433.D

Acq: 3 Mar 2016 21:34


Tgt Ion:284 Resp: 1456

Ion Ratio Lower Upper

284 100

142 0.0 0.0 0.0



249 0.0 0.0 0.0




Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E









#15  
 Pentachlorophenol  
 Concen: Below Cal  
 RT: 15.93 min Scan# 813  
 Delta R.T. 0.00 min  
 Lab File: BE091433.D  
 Acq: 3 Mar 2016 21:34




Instrument :  
 BNA\_E  
 ClientSampleId :  
 MDL-02-S


Tgt Ion: 266 Resp: 103  
 Ion Ratio Lower Upper  
 266 100  
 268 93.7 73.0 109.6  
 264 92.1 66.1 99.1



#16  
 Phenanthrene  
 Concen: 0.14 ng  
 RT: 16.13 min Scan# 819  
 Delta R.T. 0.00 min  
 Lab File: BE091433.D  
 Acq: 3 Mar 2016 21:34

Tgt Ion: 178 Resp: 7861  
 Ion Ratio Lower Upper  
 178 100  
 176 17.7 13.8 20.8  
 179 17.4 14.2 21.4









#17  
 Anthracene  
 Concen: 0.22 ng  
 RT: 16.24 min Scan# 822  
 Delta R.T. 0.00 min  
 Lab File: BE091433.D  
 Acq: 3 Mar 2016 21:34




Instrument :  
 BNA\_E  
 ClientSampleId :  
 MDL-02-S


Tgt Ion:178 Resp: 4301  
 Ion Ratio Lower Upper  
 178 100  
 176 16.2 24.4 36.6#  
 179 16.8 14.8 22.2



#18  
 Fluoranthene  
 Concen: 0.21 ng  
 RT: 18.09 min Scan# 900  
 Delta R.T. 0.02 min  
 Lab File: BE091433.D  
 Acq: 3 Mar 2016 21:34

Tgt Ion:202 Resp: 8207  
 Ion Ratio Lower Upper  
 202 100  
 101 12.7 11.0 16.6  
 203 17.7 13.7 20.5





#19

Chrysene-d12

Concen: 5.00 ng

RT: 20.21 min Scan# 1000

Delta R.T. -0.00 min

Lab File: BE091433.D

Acq: 3 Mar 2016 21:34

Instrument :

BNA\_E

ClientSampleId :

MDL-02-S


Tgt Ion:240 Resp: 249898

Ion Ratio Lower Upper

240 100

120 2.1 1.4 2.0#




236 21.5 16.7 25.1



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#20

Pyrene

Concen: 0.12 ng

RT: 18.48 min Scan# 920

Delta R.T. 0.02 min

Lab File: BE091433.D

Acq: 3 Mar 2016 21:34


Tgt Ion:202 Resp: 8614

Ion Ratio Lower Upper

202 100

200 20.9 16.5 24.7



203 17.9 13.8 20.6




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E









#21  
 Terphenyl-d14  
 Concen: 4.85 ng  
 RT: 18.59 min Scan# 926  
 Delta R.T. -0.02 min  
 Lab File: BE091433.D  
 Acq: 3 Mar 2016 21:34




Instrument :  
 BNA\_E  
 ClientSampleId :  
 MDL-02-S


Tgt Ion: 244 Resp: 184410  
 Ion Ratio Lower Upper  
 244 100  
 212 9.7 7.0 10.6  
 122 15.3 6.8 10.2#



#22  
 Benzo(a)anthracene  
 Concen: 0.11 ng  
 RT: 20.17 min Scan# 998  
 Delta R.T. -0.00 min  
 Lab File: BE091433.D  
 Acq: 3 Mar 2016 21:34

Tgt Ion: 228 Resp: 9862  
 Ion Ratio Lower Upper  
 228 100  
 226 29.8 23.1 34.7  
 229 18.0 14.5 21.7









#23  
Chrysene  
Concen: 0.13 ng  
RT: 20.24 min Scan# 1001  
Delta R.T. -0.00 min  
Lab File: BE091433.D  
Acq: 3 Mar 2016 21:34




Instrument :  
BNA\_E  
ClientSampleId :  
MDL-02-S


Tgt Ion:228 Resp: 12025  
Ion Ratio Lower Upper  
228 100  
226 28.8 22.0 33.0  
229 20.3 16.2 24.2



#24  
Indeno(1,2,3-cd)pyrene  
Concen: 0.13 ng  
RT: 23.70 min Scan# 1263  
Delta R.T. 0.01 min  
Lab File: BE091433.D  
Acq: 3 Mar 2016 21:34

Tgt Ion:276 Resp: 9782  
Ion Ratio Lower Upper  
276 100  
138 26.8 16.1 24.1#  
277 24.4 22.9 34.3









#25  
 Perylene-d12  
 Concen: 5.00 ng  
 RT: 22.09 min Scan# 1126  
 Delta R.T. -0.00 min  
 Lab File: BE091433.D  
 Acq: 3 Mar 2016 21:34




Instrument :  
 BNA\_E  
 ClientSampleId :  
 MDL-02-S


Tgt Ion: 264 Resp: 281243  
 Ion Ratio Lower Upper  
 264 100  
 260 23.3 18.0 27.0  
 265 22.6 18.2 27.4



#26  
 Benzo(b)fluoranthene  
 Concen: 0.24 ng  
 RT: 21.57 min Scan# 1081  
 Delta R.T. 0.03 min  
 Lab File: BE091433.D  
 Acq: 3 Mar 2016 21:34

Tgt Ion: 252 Resp: 18828  
 Ion Ratio Lower Upper  
 252 100  
 253 24.3 17.0 25.4  
 125 10.5 8.3 12.5









#27  
Benzo(k)fluoranthene  
Concen: 0.25 ng  
RT: 21.57 min Scan# 1081  
Delta R.T. -0.00 min  
Lab File: BE091433.D  
Acq: 3 Mar 2016 21:34




Instrument :  
BNA\_E  
ClientSampleId :  
MDL-02-S


Tgt Ion: 252 Resp: 19007  
Ion Ratio Lower Upper  
252 100  
253 24.3 19.5 29.3  
125 10.5 7.3 10.9



#28  
Benzo(a)pyrene  
Concen: 0.12 ng  
RT: 21.99 min Scan# 1117  
Delta R.T. -0.00 min  
Lab File: BE091433.D  
Acq: 3 Mar 2016 21:34

Tgt Ion: 252 Resp: 8057  
Ion Ratio Lower Upper  
252 100  
253 22.8 16.2 24.4  
125 13.6 9.6 14.4





#29

Dibenzo(a,h)anthracene

Concen: 0.12 ng

RT: 23.65 min Scan# 1259

Delta R.T. 0.01 min

Lab File: BE091433.D

Acq: 3 Mar 2016 21:34

Instrument :

BNA\_E

ClientSampleId :

MDL-02-S


Tgt Ion:278 Resp: 7546

Ion Ratio Lower Upper




278 100

139 17.3 12.1 18.1

279 25.8 19.4 29.0



Abundance Ion 278.00 (277.70 to 278.70): E  
Ion 139.00 (138.70 to 139.70): E  
Ion 279.00 (278.70 to 279.70): E



#30

Benzo(g,h,i)perylene

Concen: 0.12 ng

RT: 24.24 min Scan# 1310

Delta R.T. 0.01 min

Lab File: BE091433.D

Acq: 3 Mar 2016 21:34


Tgt Ion:276 Resp: 8744

Ion Ratio Lower Upper

276 100

277 26.1 19.0 28.6

138 25.0 18.2 27.4



Abundance Ion 276.00 (275.70 to 276.70): E  
Ion 277.00 (276.70 to 277.70): E  
Ion 138.00 (137.70 to 138.70): E

