

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE030316\
 Data File : BE091426.D
 Acq On : 3 Mar 2016 17:17
 Operator : UM/IZ
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.8

Quant Time: Mar 04 03:24:16 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE030316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 04 03:11:46 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	30726	5.00	ng	0.00
4) Naphthalene-d8	9.53	136	144145	5.00	ng	0.00
8) Acenaphthene-d10	13.34	164	83460	5.00	ng	0.00
13) Phenanthrene-d10	16.10	188	112414	5.00	ng	0.00
19) Chrysene-d12	20.21	240	234201	5.00	ng	0.00
25) Perylene-d12	22.09	264	271138	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	8.16	82	5617	0.71	ng	-0.02
9) 2-Fluorobiphenyl	12.00	172	16130	0.75	ng	0.00
21) Terphenyl-d14	18.61	244	28147	0.80	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.95	88	2120	0.42	ng	89
3) n-Nitrosodimethylamine	3.45	42	2528	5.93	ng	# 90
6) Naphthalene	9.57	128	25341	0.82	ng	98
7) 2-Methylnaphthalene	11.11	142	15038m	0.81	ng	
10) Acenaphthylene	13.09	152	103866	0.93	ng	# 96
11) Acenaphthene	13.38	154	18486	0.90	ng	# 100
12) Fluorene	14.39	166	25867	0.95	ng	100
14) Hexachlorobenzene	15.38	284	4166	0.81	ng	# 100
15) Pentachlorophenol	15.89	266	433	0.87	ng	85
16) Phenanthrene	16.13	178	32170	1.12	ng	# 79
17) Anthracene	16.20	178	46538m	1.81	ng	
18) Fluoranthene	18.07	202	47075	1.65	ng	98
20) Pyrene	18.46	202	49522	0.92	ng	100
22) Benzo(a)anthracene	20.17	228	56341	1.17	ng	99
23) Chrysene	20.23	228	66842	1.02	ng	99
24) Indeno(1,2,3-cd)pyrene	23.67	276	53802m	1.15	ng	
26) Benzo(b)fluoranthene	21.53	252	53479m	0.96	ng	
27) Benzo(k)fluoranthene	21.56	252	61718m	0.93	ng	
28) Benzo(a)pyrene	21.99	252	47646	0.93	ng	99
29) Dibenzo(a,h)anthracene	23.63	278	47765	1.11	ng	97
30) Benzo(g,h,i)perylene	24.23	276	52016	0.94	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE030316\
Data File : BE091426.D
Acq On : 3 Mar 2016 17:17
Operator : UM/IZ
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.8

Quant Time: Mar 04 03:24:16 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE030316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 04 03:11:46 2016
Response via : Initial Calibration

