

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE030316\
 Data File : BE091429.D
 Acq On : 3 Mar 2016 19:08
 Operator : UM/IZ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE030316

Quant Time: Mar 04 07:56:44 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE030316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 04 06:35:39 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	29473	5.00	ng	0.00
4) Naphthalene-d8	9.53	136	138950	5.00	ng	0.00
8) Acenaphthene-d10	13.34	164	86238	5.00	ng	0.00
13) Phenanthrene-d10	16.10	188	139238	5.00	ng	0.00
19) Chrysene-d12	20.21	240	225123	5.00	ng	0.00
25) Perylene-d12	22.09	264	259119	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	8.18	82	2727	0.41	ng	0.00
9) 2-Fluorobiphenyl	12.01	172	8564	0.41	ng	0.01
21) Terphenyl-d14	18.61	244	14587	0.43	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.97	88	1025	0.39	ng	91
3) n-Nitrosodimethylamine	3.54	42	1022	0.45	ng	# 88
6) Naphthalene	9.57	128	13203	0.41	ng	99
7) 2-Methylnaphthalene	11.14	142	7383m	0.39	ng	
10) Acenaphthylene	13.09	152	55293	0.38	ng	99
11) Acenaphthene	13.41	154	10449	0.40	ng	# 100
12) Fluorene	14.41	166	12567	0.38	ng	99
14) Hexachlorobenzene	15.38	284	3376	0.41	ng	# 100
15) Pentachlorophenol	15.93	266	223m	0.50	ng	
16) Phenanthrene	16.13	178	22558	0.46	ng	99
17) Anthracene	16.23	178	20864m	0.37	ng	
18) Fluoranthene	18.09	202	24109	0.37	ng	99
20) Pyrene	18.46	202	25906	0.40	ng	99
22) Benzo(a)anthracene	20.17	228	29047	0.38	ng	100
23) Chrysene	20.24	228	35403	0.42	ng	98
24) Indeno(1,2,3-cd)pyrene	23.69	276	28613m	0.42	ng	
26) Benzo(b)fluoranthene	21.54	252	27282m	0.38	ng	
27) Benzo(k)fluoranthene	21.56	252	33409m	0.48	ng	
28) Benzo(a)pyrene	21.99	252	24739	0.40	ng	99
29) Dibenzo(a,h)anthracene	23.64	278	24381	0.41	ng	99
30) Benzo(g,h,i)perylene	24.23	276	26907	0.42	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE030316\
Data File : BE091429.D
Acq On : 3 Mar 2016 19:08
Operator : UM/IZ
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE030316

Quant Time: Mar 04 07:56:44 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE030316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 04 06:35:39 2016
Response via : Initial Calibration

