

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE030316\
 Data File : BE091436.D
 Acq On : 3 Mar 2016 23:24
 Operator : UM/IZ
 Sample : MDL-05-S
 Misc : 0.008PPM
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 MDL-05-S

Quant Time: Mar 04 08:04:53 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE030316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 04 06:35:39 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	33412	5.00	ng	-0.02
4) Naphthalene-d8	9.53	136	143622	5.00	ng	0.00
8) Acenaphthene-d10	13.34	164	90730	5.00	ng	0.00
13) Phenanthrene-d10	16.10	188	171868	5.00	ng	0.00
19) Chrysene-d12	20.21	240	237238	5.00	ng	0.00
25) Perylene-d12	22.09	264	277482	5.00	ng	0.00

System Monitoring Compounds						
5) Nitrobenzene-d5	8.16	82	36797	4.14	ng	-0.02
9) 2-Fluorobiphenyl	12.00	172	106505	4.84	ng	0.00
21) Terphenyl-d14	18.61	244	177744	4.92	ng	0.00

Target Compounds						Qvalue
2) 1,4-Dioxane	3.00	88	373	0.13	ng	95
3) n-Nitrosodimethylamine	3.44	42	11	0.05	ng	# 84
6) Naphthalene	9.57	128	4464	0.13	ng	# 94
7) 2-Methylnaphthalene	11.16	142	2684	0.14	ng	98
10) Acenaphthylene	13.09	152	17083	0.11	ng	100
11) Acenaphthene	13.41	154	3660	0.13	ng	# 100
12) Fluorene	14.41	166	4243	0.12	ng	97
14) Hexachlorobenzene	15.38	284	1541	0.15	ng	# 100
15) Pentachlorophenol	15.96	266	68	Below Cal		94
16) Phenanthrene	16.13	178	8615	0.14	ng	99
17) Anthracene	16.23	178	7080m	0.24	ng	
18) Fluoranthene	18.09	202	7582	0.19	ng	99
20) Pyrene	18.48	202	8049	0.12	ng	98
22) Benzo(a)anthracene	20.17	228	9082	0.11	ng	99
23) Chrysene	20.24	228	11110	0.13	ng	99
24) Indeno(1,2,3-cd)pyrene	23.70	276	7925m	0.11	ng	
26) Benzo(b)fluoranthene	21.54	252	10250m	0.13	ng	
27) Benzo(k)fluoranthene	21.57	252	8311m	0.11	ng	
28) Benzo(a)pyrene	21.99	252	7294	0.11	ng	94
29) Dibenzo(a,h)anthracene	23.64	278	6866	0.11	ng	# 94
30) Benzo(g,h,i)perylene	24.24	276	8128	0.12	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE030316\
Data File : BE091436.D
Acq On : 3 Mar 2016 23:24
Operator : UM/IZ
Sample : MDL-05-S
Misc : 0.008PPM
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
MDL-05-S

Quant Time: Mar 04 08:04:53 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE030316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 04 06:35:39 2016
Response via : Initial Calibration

