

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE030316\
 Data File : BE091429.D
 Acq On : 3 Mar 2016 19:08
 Operator : UM/IZ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE030316

Quant Time: Mar 04 07:56:44 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE030316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 04 06:35:39 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	98	0.00
2	1,4-Dioxane	0.400	0.393	1.8	109	0.00
3	n-Nitrosodimethylamine	0.400	0.447	-11.7	99	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	98	0.00
5 S	Nitrobenzene-d5	0.400	0.413	-3.2	104	0.00
6	Naphthalene	0.400	0.408	-2.0	95	0.00
7	2-Methylnaphthalene	0.400	0.392	2.0	98	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	98	0.00
9 S	2-Fluorobiphenyl	0.400	0.409	-2.2	103	0.01
10	Acenaphthylene	0.400	0.382	4.5	105	0.00
11	Acenaphthene	0.400	0.402	-0.5	108	0.00
12	Fluorene	0.400	0.376	6.0	103	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	95	0.00
14	Hexachlorobenzene	0.400	0.409	-2.2	107	0.00
15	Pentachlorophenol	0.400	0.500	-25.0	162	0.00
16	Phenanthrene	0.400	0.461	-15.3	114	0.00
17	Anthracene	0.400	0.365	8.8	109	0.00
18	Fluoranthene	0.400	0.368	8.0	104	0.02
19 I	Chrysene-d12	5.000	5.000	0.0	99	0.00
20	Pyrene	0.400	0.404	-1.0	102	0.00
21 S	Terphenyl-d14	0.400	0.426	-6.5	108	0.00
22	Benzo(a)anthracene	0.400	0.375	6.3	103	0.00
23	Chrysene	0.400	0.424	-6.0	103	0.00
24	Indeno(1,2,3-cd)pyrene	0.400	0.421	-5.2	110	0.00
25 I	Perylene-d12	5.000	5.000	0.0	97	0.00
26	Benzo(b)fluoranthene	0.400	0.377	5.8	105	0.00
27	Benzo(k)fluoranthene	0.400	0.477	-19.2	106	-0.01
28 C	Benzo(a)pyrene	0.400	0.397	0.8	102	0.00
29	Dibenzo(a,h)anthracene	0.400	0.408	-2.0	104	0.00
30	Benzo(g,h,i)perylene	0.400	0.416	-4.0	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0