

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE030316\
 Data File : BE091429.D
 Acq On : 3 Mar 2016 19:08
 Operator : UM/IZ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE030316

Quant Time: Mar 04 07:56:44 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE030316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 04 06:35:39 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.443	0.435	1.8	109	0.00
3	n-Nitrosodimethylamine	0.464	0.433	6.7	99	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
5 S	Nitrobenzene-d5	0.241	0.245	-1.7	104	0.00
6	Naphthalene	1.164	1.188	-2.1	95	0.00
7	2-Methylnaphthalene	0.677	0.664	1.9	98	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
9 S	2-Fluorobiphenyl	1.214	1.241	-2.2	103	0.01
10	Acenaphthylene	8.388	8.015	4.4	105	0.00
11	Acenaphthene	1.508	1.515	-0.5	108	0.00
12	Fluorene	1.936	1.822	5.9	103	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
14	Hexachlorobenzene	0.296	0.303	-2.4	107	0.00
15	Pentachlorophenol	0.052	0.020	61.5#	162#	0.00
16	Phenanthrene	1.759	2.025	-15.1	114	0.00
17	Anthracene	2.513	1.873	25.5#	109	0.00
18	Fluoranthene	2.559	2.164	15.4	104	0.02
19 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
20	Pyrene	1.425	1.438	-0.9	102	0.00
21 S	Terphenyl-d14	0.761	0.810	-6.4	108	0.00
22	Benzo(a)anthracene	1.720	1.613	6.2	103	0.00
23	Chrysene	1.853	1.966	-6.1	103	0.00
24	Indeno(1,2,3-cd)pyrene	1.509	1.589	-5.3	110	0.00
25 I	Perylene-d12	1.000	1.000	0.0	97	0.00
26	Benzo(b)fluoranthene	1.395	1.316	5.7	105	0.00
27	Benzo(k)fluoranthene	1.351	1.612	-19.3	106	-0.01
28 C	Benzo(a)pyrene	1.204	1.193	0.9	102	0.00
29	Dibenzo(a,h)anthracene	1.152	1.176	-2.1	104	0.00
30	Benzo(g,h,i)perylene	1.247	1.298	-4.1	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0