

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE030316\
 Data File : BE091440.D
 Acq On : 4 Mar 2016 1:50
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001

Quant Time: Mar 04 08:09:06 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE030316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 04 06:35:39 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	104	0.00
2	1,4-Dioxane	0.400	0.388	3.0	114	0.00
3	n-Nitrosodimethylamine	0.400	0.409	-2.2	94	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	102	0.00
5 S	Nitrobenzene-d5	0.400	0.408	-2.0	106	0.00
6	Naphthalene	0.400	0.431	-7.7	104	0.00
7	2-Methylnaphthalene	0.400	0.375	6.3	97	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	101	0.00
9 S	2-Fluorobiphenyl	0.400	0.414	-3.5	108	0.01
10	Acenaphthylene	0.400	0.361	9.8	102	0.00
11	Acenaphthene	0.400	0.381	4.8	106	0.00
12	Fluorene	0.400	0.349	12.8	99	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	100	0.00
14	Hexachlorobenzene	0.400	0.388	3.0	106	0.00
15	Pentachlorophenol	0.400	-1.000	350.0#	82	-0.03
16	Phenanthrene	0.400	0.415	-3.7	108	0.00
17	Anthracene	0.400	0.338	15.5	97	0.00
18	Fluoranthene	0.400	0.360	10.0	105	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	103	0.00
20	Pyrene	0.400	0.381	4.8	100	0.00
21 S	Terphenyl-d14	0.400	0.445	-11.2	117	0.00
22	Benzo(a)anthracene	0.400	0.363	9.3	104	0.00
23	Chrysene	0.400	0.400	0.0	101	0.00
24	Indeno(1,2,3-cd)pyrene	0.400	0.376	6.0	102	0.00
25 I	Perylene-d12	5.000	5.000	0.0	104	0.00
26	Benzo(b)fluoranthene	0.400	0.335	16.3	100	0.00
27	Benzo(k)fluoranthene	0.400	0.415	-3.7	99	-0.01
28 C	Benzo(a)pyrene	0.400	0.361	9.8	100	0.00
29	Dibenzo(a,h)anthracene	0.400	0.358	10.5	98	-0.01
30	Benzo(g,h,i)perylene	0.400	0.371	7.3	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0