

Data Path : Z:\HPCHEM1\BNA_E\Data\BE030515\
 Data File : BE089741.D
 Acq On : 5 Mar 2015 12:44
 Operator : TP/JJ
 Sample : SSTD0.408
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Quant Time: Mar 06 09:38:35 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM01.2-EPA-SIM-BE030315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 04 14:17:55 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	149	0.00
2 SURR	1,4-Dioxane-d8	0.526	0.460	12.5	129	0.00
3	1,4-Dioxane	0.609	0.534	12.3	136	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	156#	0.00
5	Naphthalene	0.927	0.850	8.3	155#	0.00
6 SURR	2-Methylnaphthalene-d10	0.428	0.414	3.3	168#	0.00
7	2-Methylnaphthalene	0.539	0.517	4.1	167#	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	174#	0.00
9	Acenaphthylene	5.722	5.964	-4.2	183#	0.00
10 C	Acenaphthene	4.229	4.246	-0.4	173#	0.00
11	Fluorene	1.124	1.159	-3.1	176#	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	169#	0.00
13	Pentachlorophenol	0.016	0.017	-6.3	185#	-0.01
14	Phenanthrene	4.027	3.849	4.4	174#	0.00
15	Anthracene	3.342	3.308	1.0	181#	0.00
16 SURR	Fluoranthene-d10	0.664	0.654	1.5	174#	0.00
17 C	Fluoranthene	3.748	3.633	3.1	170#	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	158#	0.00
19	Pyrene	5.239	5.261	-0.4	171#	0.00
20	Benzo(a)anthracene	0.555	0.569	-2.5	190#	0.00
21	Chrysene	1.167	1.125	3.6	157#	0.00
22 I	Perylene-d12	1.000	1.000	0.0	172#	0.00
23	Benzo(b)fluoranthene	0.407	0.328	19.4	168#	0.00
24	Benzo(k)fluoranthene	1.303	1.292	0.8	173#	0.00
25 C	Benzo(a)pyrene	0.538	0.507	5.8	189#	0.00
26	Indeno(1,2,3-cd)pyrene	1.943	2.034	-4.7	260#	0.00
27	Dibenzo(a,h)anthracene	0.350	0.281	19.7	172#	0.00
28	Benzo(g,h,i)perylene	2.660	2.518	5.3	187#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0