



1,4-Dioxane

Response Ratio



Resp Ratio = $5.77 \times 10^{-1} \times \text{Amt} + 1.37 \times 10^{-2}$
Coef of Det (r^2) = 1.000 Curve Fit: Linear


Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE030617.M
Calibration Table Last Updated: Mon Mar 06 15:46:01 2017



Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE030617.M
Calibration Table Last Updated: Mon Mar 06 15:46:01 2017

Benzo(a)anthracene

Response Ratio




Resp Ratio = 5.27e+000 * Amt - 7.08e-002
Coef of Det (r^2) = 0.997 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE030617.M
Calibration Table Last Updated: Mon Mar 06 15:46:01 2017

Bis(2-ethylhexyl)phthalate

Response Ratio



$R = 1.19e-001 A^2 + 3.57e-001 A + 3.23e-003$
Coef of Det (r^2) = 0.996 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE030617.M
Calibration Table Last Updated: Mon Mar 06 15:46:01 2017