




Data Path : Z:\HPCHEM1\BNA_E\Data\BE030617\
Data File : BE092789.D
Acq On : 6 Mar 2017 16:23
Operator : SJ/MA
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 07 06:31:24 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 17 17:31:55 2017
Response via : Initial Calibration



TIC: BE092789.D

(69) Pentachlorophenol (C)




16.647min (+5.462) 1162953.20ng m

response 122604

Ion	Exp%	Act%
265.70	100	100
268.00	66.60	64.87
264.00	62.90	65.80
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE030617\
 Data File : BE092789.D
 Acq On : 6 Mar 2017 16:23
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 07 06:31:24 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 17 17:31:55 2017
 Response via : Initial Calibration



TIC: BE092789.D

(76) Benzidine

12.705min (-0.012) 80.46ng

response 24

Ion	Exp%	Act%
184.00	100	100
185.00	16.70	80.49#
183.00	11.50	29.27#
0.00	0.00	0.00