

Data Path : Z:\HPCHEM1\BNA E\DATA\BE030617\
 Data File : BE092788.D
 Acq On : 6 Mar 2017 15:15
 Operator : SJ/MA
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE030617

Quant Time: Mar 06 15:50:47 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE030617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 06 15:46:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	-0.02
2	1,4-Dioxane	0.743	0.762	-2.6	125	-0.01
3	n-Nitrosodimethylamine	0.839	0.834	0.6	121	-0.03
4 S	2-Fluorophenol	1.058	0.962	9.1	116	-0.02
5 S	Phenol-d6	1.260	1.101	12.6	113	-0.02
6	bis(2-Chloroethyl)ether	1.374	1.332	3.1	117	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	109	-0.02
8 S	Nitrobenzene-d5	0.306	0.293	4.2	127	-0.02
9	Naphthalene	1.070	1.104	-3.2	123	0.00
10	Hexachlorobutadiene	0.154	0.166	-7.8	124	-0.02
11	2-Methylnaphthalene	0.692	0.694	-0.3	123	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	127	-0.02
13 S	2,4,6-Tribromophenol	0.132	0.120	9.1	117	-0.02
14 S	2-Fluorobiphenyl	1.179	1.156	2.0	124	-0.02
15	Acenaphthylene	7.684	7.299	5.0	125	-0.02
16	Acenaphthene	1.076	1.069	0.7	125	-0.02
17	Fluorene	1.463	1.453	0.7	126	-0.01
18 I	Phenanthrene-d10	1.000	1.000	0.0	113	0.00
19	Hexachlorobenzene	0.205	0.201	2.0	129	-0.02
20	Pentachlorophenol	0.069	0.051	26.1#	106	-0.12
21	Phenanthrene	1.199	1.160	3.3	124	0.00
22	Anthracene	1.209	1.143	5.5	125	0.00
23	Fluoranthene	5.839	5.704	2.3	126	-0.02
24 I	Chrysene-d12	1.000	1.000	0.0	116	-0.02
25	Pyrene	4.529	4.366	3.6	124	0.00
26 S	Terphenyl-d14	0.603	0.581	3.6	122	-0.02
27	Benzo(a)anthracene	5.028	4.714	6.2	134	-0.02
28	Chrysene	4.585	5.001	-9.1	122	-0.02
29	Bis(2-ethylhexyl)phthalate	0.549	0.338	38.4#	102	0.00
30	Indeno(1,2,3-cd)pyrene	4.965	4.550	8.4	125	-0.03
31 I	Perylene-d12	1.000	1.000	0.0	110	-0.02
32	Benzo(b)fluoranthene	1.332	1.312	1.5	129	-0.02
33	Benzo(k)fluoranthene	1.340	1.394	-4.0	118	-0.02
34 C	Benzo(a)pyrene	1.211	1.192	1.6	124	-0.02
35	Dibenzo(a,h)anthracene	1.157	1.125	2.8	126	-0.02
36	Benzo(g,h,i)perylene	4.951	4.904	0.9	122	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0