

Data Path : Z:\HPCHEM1\BNA E\DATA\BE030617\
 Data File : BE092788.D
 Acq On : 6 Mar 2017 15:15
 Operator : SJ/MA
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE030617

Quant Time: Mar 06 15:50:47 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE030617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 06 15:46:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	106	-0.02
2	1,4-Dioxane	0.400	0.409	-2.2	125	-0.01
3	n-Nitrosodimethylamine	0.400	0.398	0.5	121	-0.03
4 S	2-Fluorophenol	0.400	0.364	9.0	116	-0.02
5 S	Phenol-d6	0.400	0.350	12.5	113	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.388	3.0	117	-0.02
7 I	Naphthalene-d8	5.000	5.000	0.0	109	-0.02
8 S	Nitrobenzene-d5	0.400	0.384	4.0	127	-0.02
9	Naphthalene	0.400	0.413	-3.2	123	0.00
10	Hexachlorobutadiene	0.400	0.432	-8.0	124	-0.02
11	2-Methylnaphthalene	0.400	0.401	-0.3	123	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	127	-0.02
13 S	2,4,6-Tribromophenol	0.400	0.365	8.8	117	-0.02
14 S	2-Fluorobiphenyl	0.400	0.392	2.0	124	-0.02
15	Acenaphthylene	0.400	0.380	5.0	125	-0.02
16	Acenaphthene	0.400	0.397	0.8	125	-0.02
17	Fluorene	0.400	0.397	0.8	126	-0.01
18 I	Phenanthrene-d10	5.000	5.000	0.0	113	0.00
19	Hexachlorobenzene	0.400	0.392	2.0	129	-0.02
20	Pentachlorophenol	0.400	0.357	10.8	106	-0.12
21	Phenanthrene	0.400	0.387	3.3	124	0.00
22	Anthracene	0.400	0.378	5.5	125	0.00
23	Fluoranthene	0.400	0.391	2.3	126	-0.02
24 I	Chrysene-d12	5.000	5.000	0.0	116	-0.02
25	Pyrene	0.400	0.386	3.5	124	0.00
26 S	Terphenyl-d14	0.400	0.385	3.8	122	-0.02
27	Benzo(a)anthracene	0.400	0.425	-6.2	134	-0.02
28	Chrysene	0.400	0.436	-9.0	122	-0.02
29	Bis(2-ethylhexyl)phthalate	0.400	0.326	18.5	102	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.367	8.3	125	-0.03
31 I	Perylene-d12	5.000	5.000	0.0	110	-0.02
32	Benzo(b)fluoranthene	0.400	0.394	1.5	129	-0.02
33	Benzo(k)fluoranthene	0.400	0.416	-4.0	118	-0.02
34 C	Benzo(a)pyrene	0.400	0.394	1.5	124	-0.02
35	Dibenzo(a,h)anthracene	0.400	0.389	2.8	126	-0.02
36	Benzo(g,h,i)perylene	0.400	0.396	1.0	122	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0