

Data Path : Z:\HPCHEM1\BNA E\DATA\BE030617\
 Data File : BE092790.D
 Acq On : 6 Mar 2017 17:02
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 07 02:29:10 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE030617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 06 15:46:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	-0.02
2	1,4-Dioxane	0.743	0.827	-11.3	132	-0.01
3	n-Nitrosodimethylamine	0.839	0.782	6.8	111	-0.03
4 S	2-Fluorophenol	1.058	0.985	6.9	116	-0.02
5 S	Phenol-d6	1.260	1.073	14.8	107	-0.02
6	bis(2-Chloroethyl)ether	1.374	1.300	5.4	111	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	104	-0.02
8 S	Nitrobenzene-d5	0.306	0.285	6.9	119	-0.02
9	Naphthalene	1.070	1.119	-4.6	120	0.00
10	Hexachlorobutadiene	0.154	0.165	-7.1	117	-0.02
11	2-Methylnaphthalene	0.692	0.688	0.6	117	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	116	-0.02
13 S	2,4,6-Tribromophenol	0.132	0.116	12.1	104	-0.02
14 S	2-Fluorobiphenyl	1.179	1.206	-2.3	119	-0.01
15	Acenaphthylene	7.684	7.466	2.8	117	-0.02
16	Acenaphthene	1.076	1.096	-1.9	118	-0.02
17	Fluorene	1.463	1.480	-1.2	118	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	108	0.00
19	Hexachlorobenzene	0.205	0.193	5.9	119	-0.02
20	Pentachlorophenol	0.069	0.054	21.7	106	-0.12
21	Phenanthrene	1.199	1.152	3.9	117	0.00
22	Anthracene	1.209	1.130	6.5	118	0.00
23	Fluoranthene	5.839	5.644	3.3	119	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	105	-0.02
25	Pyrene	4.529	4.569	-0.9	117	0.00
26 S	Terphenyl-d14	0.603	0.612	-1.5	116	-0.02
27	Benzo(a)anthracene	5.028	4.735	5.8	122	-0.02
28	Chrysene	4.585	5.211	-13.7	114	0.00
29	Bis(2-ethylhexyl)phthalate	0.549	0.378	31.1#	103	0.00
30	Indeno(1,2,3-cd)pyrene	4.965	4.752	4.3	118	-0.03
31 I	Perylene-d12	1.000	1.000	0.0	106	-0.02
32	Benzo(b)fluoranthene	1.332	1.252	6.0	119	-0.02
33	Benzo(k)fluoranthene	1.340	1.419	-5.9	116	-0.02
34 C	Benzo(a)pyrene	1.211	1.193	1.5	120	-0.02
35	Dibenzo(a,h)anthracene	1.157	1.097	5.2	118	-0.02
36	Benzo(g,h,i)perylene	4.951	4.843	2.2	117	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0