

Data Path : Z:\HPCHEM1\BNA E\DATA\BE030617\
 Data File : BE092803.D
 Acq On : 7 Mar 2017 1:09
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 07 02:49:32 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE030617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 06 15:46:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	-0.02
2	1,4-Dioxane	0.743	0.731	1.6	118	-0.01
3	n-Nitrosodimethylamine	0.839	0.764	8.9	110	-0.02
4 S	2-Fluorophenol	1.058	0.981	7.3	117	-0.02
5 S	Phenol-d6	1.260	1.120	11.1	113	0.00
6	bis(2-Chloroethyl)ether	1.374	1.279	6.9	111	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	111	-0.02
8 S	Nitrobenzene-d5	0.306	0.281	8.2	124	-0.02
9	Naphthalene	1.070	1.109	-3.6	126	0.00
10	Hexachlorobutadiene	0.154	0.162	-5.2	123	-0.02
11	2-Methylnaphthalene	0.692	0.695	-0.4	125	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	128	-0.02
13 S	2,4,6-Tribromophenol	0.132	0.119	9.8	118	-0.02
14 S	2-Fluorobiphenyl	1.179	1.181	-0.2	128	-0.02
15	Acenaphthylene	7.684	7.436	3.2	129	-0.02
16	Acenaphthene	1.076	1.101	-2.3	130	-0.02
17	Fluorene	1.463	1.489	-1.8	130	-0.01
18 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
19	Hexachlorobenzene	0.205	0.190	7.3	131	-0.02
20	Pentachlorophenol	0.069	0.045	34.8#	99	-0.12
21	Phenanthrene	1.199	1.169	2.5	133	0.00
22	Anthracene	1.209	1.127	6.8	131	0.00
23	Fluoranthene	5.839	5.598	4.1	131	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	107	-0.02
25	Pyrene	4.529	4.929	-8.8	130	0.00
26 S	Terphenyl-d14	0.603	0.668	-10.8	129	-0.02
27	Benzo(a)anthracene	5.028	4.732	5.9	125	-0.02
28	Chrysene	4.585	5.472	-19.3	123	0.00
29	Bis(2-ethylhexyl)phthalate	0.549	0.491	10.6	137	0.00
30	Indeno(1,2,3-cd)pyrene	4.965	5.179	-4.3	132	-0.03
31 I	Perylene-d12	1.000	1.000	0.0	116	-0.02
32	Benzo(b)fluoranthene	1.332	1.265	5.0	132	-0.02
33	Benzo(k)fluoranthene	1.340	1.438	-7.3	129	-0.02
34 C	Benzo(a)pyrene	1.211	1.239	-2.3	137	-0.02
35	Dibenzo(a,h)anthracene	1.157	1.103	4.7	130	-0.02
36	Benzo(g,h,i)perylene	4.951	4.858	1.9	128	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0