

Data Path : Z:\HPCHEM1\BNA E\DATA\BE030617\
 Data File : BE092803.D
 Acq On : 7 Mar 2017 1:09
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 07 02:49:32 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE030617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 06 15:46:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	105	-0.02
2	1,4-Dioxane	0.400	0.388	3.0	118	-0.01
3	n-Nitrosodimethylamine	0.400	0.364	9.0	110	-0.02
4 S	2-Fluorophenol	0.400	0.371	7.3	117	-0.02
5 S	Phenol-d6	0.400	0.355	11.3	113	0.00
6	bis(2-Chloroethyl)ether	0.400	0.372	7.0	111	-0.02
7 I	Naphthalene-d8	5.000	5.000	0.0	111	-0.02
8 S	Nitrobenzene-d5	0.400	0.367	8.3	124	-0.02
9	Naphthalene	0.400	0.415	-3.7	126	0.00
10	Hexachlorobutadiene	0.400	0.422	-5.5	123	-0.02
11	2-Methylnaphthalene	0.400	0.402	-0.5	125	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	128	-0.02
13 S	2,4,6-Tribromophenol	0.400	0.363	9.3	118	-0.02
14 S	2-Fluorobiphenyl	0.400	0.401	-0.3	128	-0.02
15	Acenaphthylene	0.400	0.387	3.3	129	-0.02
16	Acenaphthene	0.400	0.410	-2.5	130	-0.02
17	Fluorene	0.400	0.407	-1.7	130	-0.01
18 I	Phenanthrene-d10	5.000	5.000	0.0	120	0.00
19	Hexachlorobenzene	0.400	0.372	7.0	131	-0.02
20	Pentachlorophenol	0.400	0.324	19.0	99	-0.12
21	Phenanthrene	0.400	0.390	2.5	133	0.00
22	Anthracene	0.400	0.373	6.8	131	0.00
23	Fluoranthene	0.400	0.383	4.3	131	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	107	-0.02
25	Pyrene	0.400	0.435	-8.7	130	0.00
26 S	Terphenyl-d14	0.400	0.443	-10.7	129	-0.02
27	Benzo(a)anthracene	0.400	0.426	-6.5	125	-0.02
28	Chrysene	0.400	0.477	-19.2	123	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.489	-22.2	137	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.417	-4.2	132	-0.03
31 I	Perylene-d12	5.000	5.000	0.0	116	-0.02
32	Benzo(b)fluoranthene	0.400	0.380	5.0	132	-0.02
33	Benzo(k)fluoranthene	0.400	0.429	-7.2	129	-0.02
34 C	Benzo(a)pyrene	0.400	0.409	-2.2	137	-0.02
35	Dibenzo(a,h)anthracene	0.400	0.381	4.8	130	-0.02
36	Benzo(g,h,i)perylene	0.400	0.392	2.0	128	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0