

Data Path : Z:\HPCHEM1\BNA E\DATA\BE030617\
 Data File : BE092780.D
 Acq On : 6 Mar 2017 10:16
 Operator : SJ/MA
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Manual Integrations
 APPROVED

mohammad
 3/7/2017 4:22:32 PM

Quant Time: Mar 06 14:41:38 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE030617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 06 14:29:09 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.10	152	9854	5.00	ng	0.00
7) Naphthalene-d8	10.88	136	43796	5.00	ng	0.00
12) Acenaphthene-d10	14.75	164	29223	5.00	ng	0.00
18) Phenanthrene-d10	17.51	188	64761	5.00	ng	0.00
24) Chrysene-d12	21.68	240	88520	5.00	ng	0.00
31) Perylene-d12	24.01	264	81176	5.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.61	112	215	0.11	ng	0.00
5) Phenol-d6	7.31	99	227	0.10	ng	0.02
8) Nitrobenzene-d5	9.29	82	238	0.10	ng	0.00
13) 2,4,6-Tribromophenol	16.27	330	69	0.11	ng	0.01
14) 2-Fluorobiphenyl	13.39	172	681	0.10	ng	0.00
26) Terphenyl-d14	20.14	244	1105	0.10	ng	0.02

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.33	88	259	0.19	ng	# 66
3) n-Nitrosodimethylamine	3.70	42	158	0.11	ng	# 93
6) bis(2-Chloroethyl)ether	7.52	93	265	0.10	ng	# 67
9) Naphthalene	10.93	128	1019	0.11	ng	# 78
10) Hexachlorobutadiene	11.20	225	137	0.10	ng	90
11) 2-Methylnaphthalene	12.59	142	587	0.10	ng	# 94
15) Acenaphthylene	14.47	152	4846	0.11	ng	99
16) Acenaphthene	14.82	154	547	0.09	ng	# 65
17) Fluorene	15.82	166	834	0.10	ng	96
19) Hexachlorobenzene	16.84	284	242	0.10	ng	# 39
20) Pentachlorophenol	17.21	266	122	0.22	ng	87
21) Phenanthrene	17.56	178	1496	0.10	ng	98
22) Anthracene	17.65	178	1472	0.10	ng	98
23) Fluoranthene	19.56	202	7529	0.11	ng	97
25) Pyrene	19.92	202	8068	0.10	ng	100
27) Benzo(a)anthracene	21.66	228	11873m	0.14	ng	
28) Chrysene	21.72	228	7349m	0.07	ng	
29) Bis(2-ethylhexyl)phthalate	21.59	149	2240	0.19	ng	# 35
30) Indeno(1,2,3-cd)pyrene	26.50	276	9361	0.08	ng	99
32) Benzo(b)fluoranthene	23.30	252	2314	0.10	ng	# 78
33) Benzo(k)fluoranthene	23.35	252	2339	0.10	ng	# 82
34) Benzo(a)pyrene	23.90	252	1958	0.09	ng	# 72
35) Dibenzo(a,h)anthracene	26.51	278	1902	0.09	ng	# 75
36) Benzo(g,h,i)perylene	27.25	276	8630	0.09	ng	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE030617\
Data File : BE092780.D
Acq On : 6 Mar 2017 10:16
Operator : SJ/MA
Sample : SSTDIC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:32 PM

Quant Time: Mar 06 14:41:38 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE030617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 06 14:29:09 2017
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.10 min Scan# 573

Delta R.T. -0.00 min

Lab File: BE092780.D

Acq: 6 Mar 2017 10:16

Instrument :

BNA_E


ClientSampleId :

SSTDIC0.1

Tot Ion	Ratio	Lower	Upper
152	100		
150	156.7	106.0	159.0
115	59.3	31.2	46.8

Manual Integrations
APPROVED




mohammad
3/7/2017 4:22:32 PM



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 0.19 ng


RT: 3.33 min Scan# 64

Delta R.T. 0.00 min

Lab File: BE092780.D

Acq: 6 Mar 2017 10:16



Tgt Ion	Ratio	Lower	Upper
88	100		
58	43.2	63.6	95.4
43	24.3	28.0	42.0






Abundance Ion 88.00 (87.70 to 88.70): BE0

Ion 58.00 (57.70 to 58.70): BE0

Ion 43.00 (42.70 to 43.70): BE0





#3

n-Nitrosodimethylamine

Concen: 0.11 ng

RT: 3.70 min Scan# 96

Delta R.T. 0.00 min

Lab File: BE092780.D

Acq: 6 Mar 2017 10:16

Tot Ion:	42	Resp:	158
Ion	Ratio	Lower	Upper
42	100		
74	113.9	86.0	129.0
44	0.0	5.1	7.7#

Instrument :





BNA_E

ClientSampleId :

SSTDIC0.1

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:32 PM



#4

2-Fluorophenol

Concen: 0.11 ng


RT: 5.61 min Scan# 352




Delta R.T. -0.00 min

Lab File: BE092780.D

Acq: 6 Mar 2017 10:16

Tgt Ion:	112	Resp:	215
Ion	Ratio	Lower	Upper
112	100		
64	65.6	53.5	80.3
63	34.9	27.9	41.9









#5
Phenol-d6
Concen: 0.10 ng
RT: 7.31 min Scan# 539
Delta R.T. 0.02 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Tot Ion	Ratio	Lower	Upper
99	100		
42	0.0	16.0	24.0#
71	49.8	30.6	45.8#

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1


Manual Integrations
APPROVED


mohammad
3/7/2017 4:22:32 PM



#6
bis(2-Chloroethyl)ether
Concen: 0.10 ng
RT: 7.52 min Scan# 548
Delta R.T. -0.00 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Tgt Ion	Ratio	Lower	Upper
93	100		
63	65.7	71.6	107.4#
95	0.0	24.0	36.0#









#7
Naphthalene-d8
Concen: 5.00 ng
RT: 10.88 min Scan# 704
Delta R.T. -0.00 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tot Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.2	12.4
54	8.8	9.6	14.4#
68	6.3	6.7	10.1#




Manual Integrations
APPROVED


mohammad
3/7/2017 4:22:32 PM



#8
Nitrobenzene-d5
Concen: 0.10 ng
RT: 9.29 min Scan# 625
Delta R.T. -0.00 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Tgt Ion	Ratio	Lower	Upper
82	100		
128	66.7	39.0	58.4#
54	96.9	44.8	67.2#





#9

Naphthalene

Concen: 0.11 ng

RT: 10.93 min Scan# 707

Delta R.T. -0.00 min

Lab File: BE092780.D

Acq: 6 Mar 2017 10:16

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.1

Tot Ion	Ratio	Lower	Upper
128	100		
129	23.6	11.2	16.8#
127	22.8	11.6	17.4#

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:32 PM



#10

Hexachlorobutadiene

Concen: 0.10 ng




RT: 11.20 min Scan# 721


Delta R.T. -0.00 min

Lab File: BE092780.D

Acq: 6 Mar 2017 10:16

Tgt Ion	Ratio	Lower	Upper
225	100		
223	70.1	50.6	76.0
227	72.3	51.4	77.0









#11
2-Methylnaphthalene
Concen: 0.10 ng
RT: 12.59 min Scan# 811
Delta R.T. 0.01 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.1

Tot Ion	Ratio	Lower	Upper
142	100		
141	91.4	73.7	110.5
115	52.9	34.0	51.0#




Manual Integrations
APPROVED


mohammad
3/7/2017 4:22:32 PM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.75 min Scan# 976
Delta R.T. 0.00 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Tgt Ion	Ratio	Lower	Upper
164	100		
162	109.6	71.4	107.0#
160	51.7	27.4	41.2#









#13
 2,4,6-Tribromophenol
 Concen: 0.11 ng
 RT: 16.27 min Scan# 1085
 Delta R.T. 0.01 min
 Lab File: BE092780.D
 Acq: 6 Mar 2017 10:16

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Tot Ion	Ratio	Lower	Upper
330	100		
332	117.4	75.4	113.0#
141	56.5	31.0	46.6#




Manual Integrations
 APPROVED


mohammad
 3/7/2017 4:22:32 PM



#14
 2-Fluorobiphenyl
 Concen: 0.10 ng
 RT: 13.39 min Scan# 880
 Delta R.T. 0.00 min
 Lab File: BE092780.D
 Acq: 6 Mar 2017 10:16

Tgt Ion	Ratio	Lower	Upper
172	100		
171	45.9	30.1	45.1#
170	38.5	21.8	32.6#









#15
Acenaphthylene
Concen: 0.11 ng
RT: 14.47 min Scan# 960
Delta R.T. 0.02 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.1

Tot Ion	Ratio	Lower	Upper
152	100		
151	4.5	3.9	5.9
153	12.4	10.4	15.6




Manual Integrations
APPROVED


mohammad
3/7/2017 4:22:32 PM



#16
Acenaphthene
Concen: 0.09 ng
RT: 14.82 min Scan# 980
Delta R.T. 0.00 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Tgt Ion	Ratio	Lower	Upper
154	100		
153	531.4	350.8	526.2#
152	263.4	171.1	256.7#





#17

Fluorene

Concen: 0.10 ng

RT: 15.82 min Scan# 1046

Delta R.T. 0.01 min

Lab File: BE092780.D

Acq: 6 Mar 2017 10:16

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.1

Tot Ion	Ion	Ratio	Lower	Upper
166	100			
165	102.0	78.2	117.4	
167	13.2	11.0	16.4	

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:32 PM



#18

Phenanthrene-d10

Concen: 5.00 ng




RT: 17.51 min Scan# 1151


Delta R.T. -0.00 min

Lab File: BE092780.D

Acq: 6 Mar 2017 10:16

Tgt Ion	Ion	Ratio	Lower	Upper
188	100			
94	6.2	3.2	4.8#	
80	5.9	2.6	3.8#	






#19
Hexachlorobenzene
Concen: 0.10 ng
RT: 16.84 min Scan# 1122
Delta R.T. 0.02 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1




Tot Ion	Ratio	Lower	Upper
284	100		
142	0.0	29.1	43.7#
249	0.0	27.9	41.9#

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:32 PM






Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E




#20
Pentachlorophenol
Concen: 0.22 ng
RT: 17.21 min Scan# 1138
Delta R.T. 0.02 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Tgt Ion	Ratio	Lower	Upper
266	100		
268	89.5	62.5	93.7
264	82.9	57.2	85.8

Abundance Ion 265.70 (265.40 to 266.40): E
Ion 268.00 (267.70 to 268.70): E
Ion 264.00 (263.70 to 264.70): E









#21
Phenanthrene
Concen: 0.10 ng
RT: 17.56 min Scan# 1153
Delta R.T. -0.00 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tot Ion	Ratio	Lower	Upper
178	100		
176	19.2	15.8	23.8
179	18.9	16.0	24.0




Manual Integrations
APPROVED


mohammad
3/7/2017 4:22:32 PM



#22
Anthracene
Concen: 0.10 ng
RT: 17.65 min Scan# 1157
Delta R.T. -0.00 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.0	22.4
179	16.4	12.2	18.2









#23
 Fluoranthene
 Concen: 0.11 ng
 RT: 19.56 min Scan# 1263
 Delta R.T. -0.00 min
 Lab File: BE092780.D
 Acq: 6 Mar 2017 10:16

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Tot Ion	Ratio	Lower	Upper
202	100		
101	2.8	2.2	3.4
203	18.5	13.7	20.5




Manual Integrations
 APPROVED


mohammad
 3/7/2017 4:22:32 PM



#24
 Chrysene-d12
 Concen: 5.00 ng
 RT: 21.68 min Scan# 1372
 Delta R.T. -0.00 min
 Lab File: BE092780.D
 Acq: 6 Mar 2017 10:16

Tgt Ion	Ratio	Lower	Upper
240	100		
120	7.8	2.6	4.0#
236	31.1	24.6	37.0









#25
Pvrene
Concen: 0.10 ng
RT: 19.92 min Scan# 1284
Delta R.T. -0.00 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.1

Tot Ion	Ratio	Lower	Upper
202	100		
200	5.3	4.2	6.4
203	17.5	13.9	20.9




Manual Integrations
APPROVED


mohammad
3/7/2017 4:22:32 PM



#26
Terphenyl-d14
Concen: 0.10 ng
RT: 20.14 min Scan# 1297
Delta R.T. 0.02 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Tgt Ion	Ratio	Lower	Upper
244	100		
212	16.3	6.7	10.1#
122	14.4	3.6	5.4#









#27
Benzo(a)anthracene
Concen: 0.14 ng m
RT: 21.66 min Scan# 1371
Delta R.T. -0.00 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.1

Tot Ion	Ratio	Lower	Upper
228	100		
226	42.1	23.6	35.4#
229	17.6	13.3	19.9




Manual Integrations
APPROVED


mohammad
3/7/2017 4:22:32 PM



#28
Chrysene
Concen: 0.07 ng m
RT: 21.72 min Scan# 1374
Delta R.T. -0.00 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Tgt Ion	Ratio	Lower	Upper
228	100		
226	25.4	36.3	54.5#
229	28.2	10.6	16.0#






#29
Bis(2-ethylhexyl)phthalate
Concen: 0.19 ng
RT: 21.59 min Scan# 1368
Delta R.T. -0.00 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tot Ion	Ratio	Resp	Lower	Upper
149	100	2240		
167	59.9	16.0	24.0#	
279	0.0	16.0	24.0#	

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:32 PM






Abundance

Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#30
Indeno(1,2,3-cd)pyrene
Concen: 0.08 ng
RT: 26.50 min Scan# 1710
Delta R.T. 0.02 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16




Tgt Ion	Ratio	Resp	Lower	Upper
276	100	9361		
138	25.9	20.3	30.5	
277	24.3	19.7	29.5	


Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#31
Pervlene-d12
Concen: 5.00 ng
RT: 24.01 min Scan# 1537
Delta R.T. 0.00 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tot Ion:264 Resp: 81176
Ion Ratio Lower Upper
264 100
260 25.7 20.1 30.1
265 21.4 17.9 26.9




Manual Integrations
APPROVED


mohammad
3/7/2017 4:22:32 PM



#32
Benzo(b)fluoranthene
Concen: 0.10 ng
RT: 23.30 min Scan# 1476
Delta R.T. 0.01 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Tgt Ion:252 Resp: 2314
Ion Ratio Lower Upper
252 100
253 34.0 18.7 28.1#
125 22.1 10.5 15.7#









#33
Benzo(k)fluoranthene
Concen: 0.10 ng
RT: 23.35 min Scan# 1480
Delta R.T. 0.01 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tot Ion	Ratio	Lower	Upper
252	100		
253	33.0	20.3	30.5#
125	21.9	9.6	14.4#




Manual Integrations
APPROVED


mohammad
3/7/2017 4:22:32 PM



#34
Benzo(a)pyrene
Concen: 0.09 ng
RT: 23.90 min Scan# 1528
Delta R.T. 0.00 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Tgt Ion	Ratio	Lower	Upper
252	100		
253	35.8	19.0	28.6#
125	28.7	11.8	17.8#









#35
Dibenzo(a,h)anthracene
Concen: 0.09 ng
RT: 26.51 min Scan# 1711
Delta R.T. 0.02 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tot Ion	Ratio	Lower	Upper
278	100		
139	32.1	13.8	20.8#
279	36.3	21.4	32.2#

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:32 PM



#36
Benzo(g,h,i)perylene
Concen: 0.09 ng
RT: 27.25 min Scan# 1754
Delta R.T. 0.02 min
Lab File: BE092780.D
Acq: 6 Mar 2017 10:16

Tgt Ion	Ratio	Lower	Upper
276	100		
277	31.8	19.8	29.8#
138	34.3	19.8	29.6#

