

Data Path : Z:\HPCHEM1\BNA E\DATA\BE030617\
 Data File : BE092782.D
 Acq On : 6 Mar 2017 11:31
 Operator : SJ/MA
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICCC0.4

Manual Integrations
 APPROVED

mohammad
 3/7/2017 4:22:37 PM

Quant Time: Mar 06 14:33:19 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE030617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 06 14:29:53 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.10	152	10730	5.00	ng	0.00
7) Naphthalene-d8	10.88	136	46457	5.00	ng	0.00
12) Acenaphthene-d10	14.75	164	27216	5.00	ng	0.00
18) Phenanthrene-d10	17.51	188	65914	5.00	ng	0.00
24) Chrysene-d12	21.68	240	87801	5.00	ng	0.00
31) Perylene-d12	24.01	264	77180	5.00	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.61	112	753	0.36	ng	0.00
5) Phenol-d6	7.29	99	887	0.35	ng	0.00
8) Nitrobenzene-d5	9.29	82	930	0.36	ng	0.00
13) 2,4,6-Tribromophenol	16.26	330	283	0.49	ng	0.00
14) 2-Fluorobiphenyl	13.39	172	2578	0.39	ng	0.00
26) Terphenyl-d14	20.12	244	3892	0.35	ng	0.00

Target Compounds						Qvalue
2) 1,4-Dioxane	3.33	88	555	0.36	ng	93
3) n-Nitrosodimethylamine	3.70	42	626	0.39	ng	100
6) bis(2-Chloroethyl)ether	7.52	93	1038	0.37	ng	# 85
9) Naphthalene	10.93	128	3619	0.36	ng	98
10) Hexachlorobutadiene	11.20	225	544	0.37	ng	100
11) 2-Methylnaphthalene	12.58	142	2276	0.36	ng	94
15) Acenaphthylene	14.46	152	16129	0.39	ng	100
16) Acenaphthene	14.81	154	2356	0.39	ng	87
17) Fluorene	15.81	166	3191	0.42	ng	98
19) Hexachlorobenzene	16.82	284	923	0.36	ng	96
20) Pentachlorophenol	17.19	266	289	0.52	ng	94
21) Phenanthrene	17.56	178	5581	0.36	ng	# 74
22) Anthracene	17.65	178	5439	0.38	ng	100
23) Fluoranthene	19.56	202	26971	0.40	ng	99
25) Pyrene	19.92	202	28681	0.34	ng	100
27) Benzo(a)anthracene	21.66	228	28667m	0.34	ng	
28) Chrysene	21.72	228	33502m	0.33	ng	
29) Bis(2-ethylhexyl)phthalate	21.59	149	2706	0.23	ng	# 70
30) Indeno(1,2,3-cd)pyrene	26.48	276	29707	0.26	ng	98
32) Benzo(b)fluoranthene	23.29	252	6864	0.31	ng	99
33) Benzo(k)fluoranthene	23.34	252	7999	0.35	ng	96
34) Benzo(a)pyrene	23.90	252	6485	0.33	ng	98
35) Dibenzo(a,h)anthracene	26.50	278	6063	0.29	ng	# 95
36) Benzo(g,h,i)perylene	27.23	276	27154	0.31	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE030617\
 Data File : BE092782.D
 Acq On : 6 Mar 2017 11:31
 Operator : SJ/MA
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_E
 ClientSampleId :
 SSTDICCC0.4

Manual Integrations
 APPROVED

mohammad
 3/7/2017 4:22:37 PM

Quant Time: Mar 06 14:33:19 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE030617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 06 14:29:53 2017
 Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 5.00 ng
RT: 8.10 min Scan# 573
Delta R.T. 0.00 min
Lab File: BE092782.D
Acq: 6 Mar 2017 11:31

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
152	100		
150	157.1	106.0	159.0
115	57.8	31.2	46.8

Manual Integrations
APPROVED




mohammad
3/7/2017 4:22:37 PM




#2

1,4-Dioxane
Concen: 0.36 ng
RT: 3.33 min Scan# 64
Delta R.T. 0.00 min
Lab File: BE092782.D
Acq: 6 Mar 2017 11:31

Tgt Ion	Ratio	Lower	Upper
88	100		
58	71.9	63.6	95.4
43	32.3	28.0	42.0





#3





n-Nitrosodimethylamine
Concen: 0.39 ng
RT: 3.70 min Scan# 96
Delta R.T. 0.00 min
Lab File: BE092782.D
Acq: 6 Mar 2017 11:31

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
42	100		
74	107.7	86.0	129.0
44	7.5	5.1	7.7

Manual Integrations
APPROVED




mohammad
3/7/2017 4:22:37 PM




#4

2-Fluorophenol
Concen: 0.36 ng
RT: 5.61 min Scan# 352
Delta R.T. 0.00 min
Lab File: BE092782.D
Acq: 6 Mar 2017 11:31

Tgt Ion	Ratio	Lower	Upper
112	100		
64	72.5	53.5	80.3
63	38.0	27.9	41.9





#5

Phenol-d6

Concen: 0.35 ng

RT: 7.29 min Scan# 538

Delta R.T. 0.00 min

Lab File: BE092782.D

Acq: 6 Mar 2017 11:31

Instrument :

BNA_E





ClientSampleId :

SSTDICCC0.4

Tot Ion:	99	Resp:	887
Ion	Ratio	Lower	Upper
99	100		
42	26.5	16.0	24.0#
71	35.5	30.6	45.8

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:37 PM



#6

bis(2-Chloroethyl)ether

Concen: 0.37 ng




RT: 7.52 min Scan# 548


Delta R.T. 0.00 min

Lab File: BE092782.D

Acq: 6 Mar 2017 11:31

Tgt Ion:	93	Resp:	1038
Ion	Ratio	Lower	Upper
93	100		
63	71.4	71.6	107.4#
95	31.6	24.0	36.0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.88 min Scan# 704

Delta R.T. 0.00 min

Lab File: BE092782.D

Acq: 6 Mar 2017 11:31

Instrument :

BNA_E





ClientSampleId :

SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.2	12.4
54	8.2	9.6	14.4#
68	6.0	6.7	10.1#

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:37 PM



#8

Nitrobenzene-d5

Concen: 0.36 ng




RT: 9.29 min Scan# 625


Delta R.T. 0.00 min

Lab File: BE092782.D

Acq: 6 Mar 2017 11:31

Tgt Ion	Ratio	Lower	Upper
82	100		
128	50.2	39.0	58.4
54	62.6	44.8	67.2





#9

Naphthalene

Concen: 0.36 ng

RT: 10.93 min Scan# 707

Delta R.T. 0.00 min

Lab File: BE092782.D

Acq: 6 Mar 2017 11:31

Instrument :

BNA_E





ClientSampleId :

SSTDICCC0.4

Tot Ion:	128	Resp:	3619
Ion Ratio	Lower	Upper	
128	100		
129	15.2	11.2	16.8
127	15.2	11.6	17.4

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:37 PM



#10

Hexachlorobutadiene

Concen: 0.37 ng




RT: 11.20 min Scan# 721


Delta R.T. 0.00 min

Lab File: BE092782.D

Acq: 6 Mar 2017 11:31

Tgt Ion:	225	Resp:	544
Ion Ratio	Lower	Upper	
225	100		
223	63.4	50.6	76.0
227	63.8	51.4	77.0






#11
2-Methylnaphthalene
Concen: 0.36 ng
RT: 12.58 min Scan# 810
Delta R.T. 0.00 min
Lab File: BE092782.D
Acq: 6 Mar 2017 11:31

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
142	100		
141	86.4	73.7	110.5
115	38.9	34.0	51.0

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:37 PM






Abundance

Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.75 min Scan# 976
Delta R.T. 0.00 min
Lab File: BE092782.D
Acq: 6 Mar 2017 11:31




Tgt Ion	Ratio	Lower	Upper
164	100		
162	111.4	71.4	107.0#
160	53.4	27.4	41.2#


Abundance

Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#13
 2,4,6-Tribromophenol
 Concen: 0.49 ng
 RT: 16.26 min Scan# 1084
 Delta R.T. 0.00 min
 Lab File: BE092782.D
 Acq: 6 Mar 2017 11:31

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.4

Tot Ion:330 Resp: 283
 Ion Ratio Lower Upper
 330 100
 332 96.8 75.4 113.0
 141 37.8 31.0 46.6




Manual Integrations
 APPROVED




mohammad
 3/7/2017 4:22:37 PM



#14
 2-Fluorobiphenyl
 Concen: 0.39 ng
 RT: 13.39 min Scan# 880
 Delta R.T. 0.00 min
 Lab File: BE092782.D
 Acq: 6 Mar 2017 11:31

Tgt Ion:172 Resp: 2578
 Ion Ratio Lower Upper
 172 100
 171 34.9 30.1 45.1
 170 23.2 21.8 32.6





#15

Acenaphthylene

Concen: 0.39 ng

RT: 14.46 min Scan# 959

Delta R.T. 0.00 min

Lab File: BE092782.D

Acq: 6 Mar 2017 11:31

Tot Ion:	152	Resp:	16129
Ion Ratio	Lower	Upper	
152	100		
151	5.0	3.9	5.9
153	13.0	10.4	15.6

Instrument :





BNA_E

ClientSampleId :

SSTDICCC0.4

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:37 PM



#16

Acenaphthene

Concen: 0.39 ng


RT: 14.81 min Scan# 980


Delta R.T. 0.00 min

Lab File: BE092782.D

Acq: 6 Mar 2017 11:31

Tgt Ion:	154	Resp:	2356
Ion Ratio	Lower	Upper	
154	100		
153	470.6	350.8	526.2
152	234.3	171.1	256.7





#17

Fluorene

Concen: 0.42 ng

RT: 15.81 min Scan# 1045

Delta R.T. 0.00 min

Lab File: BE092782.D

Acq: 6 Mar 2017 11:31

Instrument :

BNA_E





ClientSampleId :

SSTDICCC0.4

Tot Ion:	166	Resp:	3191
Ion	Ratio	Lower	Upper
166	100		
165	99.8	78.2	117.4
167	13.2	11.0	16.4

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:37 PM



#18

Phenanthrene-d10

Concen: 5.00 ng




RT: 17.51 min Scan# 1151


Delta R.T. 0.00 min

Lab File: BE092782.D

Acq: 6 Mar 2017 11:31

Tgt Ion:	188	Resp:	65914
Ion	Ratio	Lower	Upper
188	100		
94	5.0	3.2	4.8#
80	4.6	2.6	3.8#









#19
Hexachlorobenzene
Concen: 0.36 ng
RT: 16.82 min Scan# 1121
Delta R.T. 0.00 min
Lab File: BE092782.D
Acq: 6 Mar 2017 11:31

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
284	100		
142	39.5	29.1	43.7
249	36.2	27.9	41.9




Manual Integrations
APPROVED


mohammad
3/7/2017 4:22:37 PM



#20
Pentachlorophenol
Concen: 0.52 ng
RT: 17.19 min Scan# 1137
Delta R.T. 0.00 min
Lab File: BE092782.D
Acq: 6 Mar 2017 11:31

Tgt Ion	Ratio	Lower	Upper
266	100		
268	73.0	62.5	93.7
264	75.9	57.2	85.8





#21

Phenanthrene

Concen: 0.36 ng

RT: 17.56 min Scan# 1153

Delta R.T. 0.00 min

Lab File: BE092782.D

Acq: 6 Mar 2017 11:31

Instrument :

BNA_E





ClientSampleId :

SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
178	100		
176	0.0	15.8	23.8#
179	15.9	16.0	24.0#

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:37 PM



#22

Anthracene

Concen: 0.38 ng




RT: 17.65 min Scan# 1157


Delta R.T. 0.00 min

Lab File: BE092782.D

Acq: 6 Mar 2017 11:31

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.0	22.4
179	15.5	12.2	18.2





#23

Fluoranthene

Concen: 0.40 ng

RT: 19.56 min Scan# 1263

Delta R.T. 0.00 min

Lab File: BE092782.D

Acq: 6 Mar 2017 11:31

Instrument :

BNA_E


ClientSampleId :

SSTDICCC0.4


Tot Ion:	202	Resp:	26971
Ion Ratio	Lower	Upper	
202	100		
101	3.0	2.2	3.4
203	17.8	13.7	20.5

Manual Integrations
APPROVED



mohammad
3/7/2017 4:22:37 PM



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E



Time--> 19.20 19.40 19.60 19.80



#24

Chrysene-d12

Concen: 5.00 ng

RT: 21.68 min Scan# 1372


Delta R.T. 0.00 min

Lab File: BE092782.D



Acq: 6 Mar 2017 11:31


Tgt Ion:	240	Resp:	87801
Ion Ratio	Lower	Upper	
240	100		
120	6.6	2.6	4.0#
236	30.6	24.6	37.0

Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E



Time--> 21.50 21.60 21.70 21.80 21.90 22.00









#25
Pvrene
Concen: 0.34 ng
RT: 19.92 min Scan# 1284
Delta R.T. 0.00 min
Lab File: BE092782.D
Acq: 6 Mar 2017 11:31

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
202	100		
200	5.3	4.2	6.4
203	17.3	13.9	20.9




Manual Integrations
APPROVED


mohammad
3/7/2017 4:22:37 PM



#26
Terphenyl-d14
Concen: 0.35 ng
RT: 20.12 min Scan# 1296
Delta R.T. 0.00 min
Lab File: BE092782.D
Acq: 6 Mar 2017 11:31

Tgt Ion	Ratio	Lower	Upper
244	100		
212	10.5	6.7	10.1#
122	10.3	3.6	5.4#






#27
Benzo(a)anthracene
Concen: 0.34 ng m
RT: 21.66 min Scan# 1371
Delta R.T. 0.00 min
Lab File: BE092782.D
Acq: 6 Mar 2017 11:31

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4




Tot Ion	Ratio	Lower	Upper
228	100		
226	36.3	23.6	35.4#
229	13.6	13.3	19.9

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:37 PM






Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E




#28
Chrysene
Concen: 0.33 ng m
RT: 21.72 min Scan# 1374
Delta R.T. 0.00 min
Lab File: BE092782.D
Acq: 6 Mar 2017 11:31

Tgt Ion	Ratio	Lower	Upper
228	100		
226	22.1	36.3	54.5#
229	27.1	10.6	16.0#

Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E





#29

Bis(2-ethylhexyl)phthalate

Concen: 0.23 ng

RT: 21.59 min Scan# 1368

Delta R.T. 0.00 min

Lab File: BE092782.D

Acq: 6 Mar 2017 11:31

Instrument :

BNA_E





ClientSampleId :

SSTDICCC0.4

Tot Ion:	149	Resp:	2706
Ion Ratio	Lower	Upper	
149	100		
167	30.0	16.0	24.0#
279	2.8	16.0	24.0#

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:37 PM



#30

Indeno(1,2,3-cd)pyrene

Concen: 0.26 ng




RT: 26.48 min Scan# 1709


Delta R.T. 0.00 min

Lab File: BE092782.D

Acq: 6 Mar 2017 11:31

Tgt Ion:	276	Resp:	29707
Ion Ratio	Lower	Upper	
276	100		
138	27.2	20.3	30.5
277	24.9	19.7	29.5









#31
Pervlene-d12
Concen: 5.00 ng
RT: 24.01 min Scan# 1537
Delta R.T. 0.00 min
Lab File: BE092782.D
Acq: 6 Mar 2017 11:31

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion:264 Resp: 77180
Ion Ratio Lower Upper
264 100
260 25.7 20.1 30.1
265 21.2 17.9 26.9




Manual Integrations
APPROVED


mohammad
3/7/2017 4:22:37 PM



#32
Benzo(b)fluoranthene
Concen: 0.31 ng
RT: 23.29 min Scan# 1475
Delta R.T. 0.00 min
Lab File: BE092782.D
Acq: 6 Mar 2017 11:31

Tgt Ion:252 Resp: 6864
Ion Ratio Lower Upper
252 100
253 22.8 18.7 28.1
125 13.0 10.5 15.7









#33
Benzo(k)fluoranthene
Concen: 0.35 ng
RT: 23.34 min Scan# 1479
Delta R.T. 0.00 min
Lab File: BE092782.D
Acq: 6 Mar 2017 11:31

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.5	20.3	30.5
125	13.2	9.6	14.4




Manual Integrations
APPROVED


mohammad
3/7/2017 4:22:37 PM



#34
Benzo(a)pyrene
Concen: 0.33 ng
RT: 23.90 min Scan# 1528
Delta R.T. 0.00 min
Lab File: BE092782.D
Acq: 6 Mar 2017 11:31

Tgt Ion	Ratio	Lower	Upper
252	100		
253	25.3	19.0	28.6
125	14.6	11.8	17.8









#35
Dibenzo(a,h)anthracene
Concen: 0.29 ng
RT: 26.50 min Scan# 1710
Delta R.T. 0.00 min
Lab File: BE092782.D
Acq: 6 Mar 2017 11:31

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
278	100		
139	22.6	13.8	20.8#
279	27.3	21.4	32.2

Manual Integrations
APPROVED

mohammad
3/7/2017 4:22:37 PM



#36
Benzo(g,h,i)perylene
Concen: 0.31 ng
RT: 27.23 min Scan# 1753
Delta R.T. 0.00 min
Lab File: BE092782.D
Acq: 6 Mar 2017 11:31

Tgt Ion	Ratio	Lower	Upper
276	100		
277	26.6	19.8	29.8
138	25.5	19.8	29.6

